

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD  
Time of Inspection: 09:55

Date of Inspection: 06/25/2021

Serial Number: 80-001073  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201910D<br>Exp: 10/22/2021 | 0.08g/210L Test (g/210L)<br>Lot#:201908E<br>Exp: 08/07/2021 | 0.20g/210L Test (g/210L)<br>Lot#:202005A<br>Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:24119080A1<br>Exp: 11/05/2021 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.079   | 0.199   | 0.079   |
| 0.000                      | 0.048   | 0.079   | 0.199   | 0.080   |
| 0.000                      | 0.048   | 0.079   | 0.200   | 0.080   |

Number of Simulators Used: 4


**Remarks:**

BACK IN SERVICE

*Expired Mouth Alcohol Used*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_ DEAN BALDWIN  
Signature and Printed Name

06/25/2021  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD  
Time of Inspection: 17:18

Date of Inspection: 04/26/2021

Serial Number: 80-001073  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201910D<br>Exp: 10/22/2021 | 0.08g/210L Test (g/210L)<br>Lot#:201908B<br>Exp: 08/07/2021 | 0.20g/210L Test (g/210L)<br>Lot#:202005A<br>Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:24119080A1<br>Exp: 11/05/2021 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.079   | 0.199   | 0.080   |
| 0.000                      | 0.049   | 0.079   | 0.199   | 0.081   |
| 0.000                      | 0.049   | 0.079   | 0.200   | 0.080   |

Number of Simulators Used: 4

Remarks:

OUT FOR DEPARTMENT INSPECTION

*Expired Mouth Alcohol Used*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DEAN BALDWIN  
Signature and Printed Name

04/26/2021  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD  
Time of Inspection: 16:51

Date of Inspection: 03/24/2021

Serial Number: 80-001073  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201910D<br>Exp: 10/22/2021 | 0.08g/210L Test (g/210L)<br>Lot#:201908B<br>Exp: 08/07/2021 | 0.20g/210L Test (g/210L)<br>Lot#:202005A<br>Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:24119080A1<br>Exp: 11/05/2021 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.079   | 0.199   | 0.080   |
| 0.000                      | 0.049   | 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.049   | 0.079   | 0.200   | 0.080   |


Number of Simulators Used: 4

Remarks:

*Expired Mouth Alcohol Used*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_  
Signature and Printed Name DEAN BALDWIN

03/24/2021  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD  
Time of Inspection: 17:50

Date of Inspection: 02/24/2021

Serial Number: 80-001073  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 201910D<br>Exp: 10/22/2021 | 0.08g/210L Test (g/210L)<br>Lot#: 201908B<br>Exp: 08/07/2021 | 0.20g/210L Test (g/210L)<br>Lot#: 202005A<br>Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: 24119080A1<br>Exp: 11/05/2021 |
|----------------------------|--|--|--|--|
| 0.000                      | 0.049  | 0.080  | 0.198  | 0.079  |
| 0.000                      | 0.049  | 0.079  | 0.199  | 0.080  |
| 0.000                      | 0.049  | 0.079  | 0.199  | 0.080  |


Number of Simulators Used: 4

Remarks:

*Expired Mouth Alcohol Used*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_  
Signature and Printed Name DEAN BALDWIN

02/24/2021  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD  
Time of Inspection: 16:46

Date of Inspection: 01/19/2021

Serial Number: 80-001073  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 201910D<br>Exp: 10/22/2021 | 0.08g/210L Test (g/210L)<br>Lot#: 201908B<br>Exp: 08/07/2021 | 0.20g/210L Test (g/210L)<br>Lot#: 202005A<br>Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: 24119080A1<br>Exp: 11/05/2021 |
|----------------------------|--|--|--|--|
| 0.000                      | 0.049  | 0.079  | 0.199  | 0.081  |
| 0.000                      | 0.049  | 0.079  | 0.199  | 0.080  |
| 0.000                      | 0.049  | 0.079  | 0.200  | 0.080  |


Number of Simulators Used: 5

Remarks:

*Expired Mouth Alcohol Used*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name DEAN BALDWIN

01/19/2021  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD  
Time of Inspection: 17:03

Date of Inspection: 09/20/2021

Serial Number: 80-001073  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202007A<br>Exp: 07/21/2022 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:04012006A1<br>Exp: 11/05/2021 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.080   | 0.197   | 0.079   |
| 0.000                      | 0.049   | 0.079   | 0.198   | 0.079   |
| 0.000                      | 0.049   | 0.080   | 0.199   | 0.080   |


Number of Simulators Used: 4

Remarks:

*Approved 11/12/2021*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
\_\_\_\_\_  
Signature and Printed Name

DEAN BALDWIN

09/20/2021  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD  
Time of Inspection: 17:07

Date of Inspection: 10/12/2021

Serial Number: 80-001073  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#: 202007A<br>Exp: 07/21/2022 | 0.20g/210L Test (g/210L)<br>Lot#: 202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: <del>202100001</del> 2462419080A1<br>Exp: 11/05/2021 |
|----------------------------|--|--|--|---|
| 0.000                      | 0.049  | 0.079  | 0.198  | 0.079   |
| 0.000                      | 0.049  | 0.079  | 0.199  | 0.079   |
| 0.000                      | 0.049  | 0.078  | 0.199  | 0.079   |

Number of Simulators Used: 4

Remarks:

*APPROVED 11/12/2021*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

\_\_\_\_\_  
Signature and Printed Name DEAN BALDWIN

10/12/2021  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD  
Time of Inspection: 16:46

Date of Inspection: 11/03/2021

Serial Number: 80-001073  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#: 202007A<br>Exp: 07/21/2022 | 0.20g/210L Test (g/210L)<br>Lot#: 202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: 04012020<br>Exp: 11/05/2021 |
|----------------------------|--|--|--|--|
| 0.000                      | 0.049  | 0.080  | 0.199  | 0.080  |
| 0.000                      | 0.049  | 0.080  | 0.199  | 0.079  |
| 0.000                      | 0.049  | 0.080  | 0.198  | 0.079  |

Number of Simulators Used: 4


Remarks:

OUT OF SERVICE DUE TO EXPIRING .08 GAS

*Approved 11/12/2021*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DEAN BALDWIN

Signature and Printed Name

11/03/2021  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD  
Time of Inspection: 12:48

Date of Inspection: 08/22/2021

Serial Number: 80-001073  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202007A<br>Exp: 07/21/2022 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L) <i>24/1080A1 NB</i><br>Lot#:04012080A1<br>Exp: 11/05/2021 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.079   | 0.198   | 0.000 / 0.078   |
| 0.000                      | 0.049   | 0.079   | 0.198   | 0.080 / 0.079   |
| 0.000                      | 0.049   | 0.078   | 0.199   | 0.080 / 0.079   |

Number of Simulators Used: 4


Remarks:

08: Control Outside, .08 GAS NOT CONNECTED

*Amended 11/12/2021*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name DEAN BALDWIN

08/22/2021  
 Date