

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Hendry County S.O.

Serial Number: 80-001207

Time of Inspection:

Date of Inspection:

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks: Instrument did not receive an Agency Inspection in Oct 2021. After a successful Agency Inspection in late Sept, the instrument was sent to FDLE for its Department Inspection in early Oct with a breath test record on it. After returning from FDLE, instrument underwent a successful Agency Inspection and was put back in service in Nov 2021

The above instrument complies (\_\_\_\_) does not comply (\_\_\_\_) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
James R. Hester

---

Signature and Printed Name

1/26/22  
 Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY S.O.  
Time of Inspection: 18:12

Date of Inspection: 02/25/2021

Serial Number: 80-001207  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: <del>201910A</del> 201910D Exp: 10/22/2021 JK	0.08g/210L Test (g/210L) Lot#: 202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1012059 Exp: 09/05/2021
0.000	0.049	0.079	0.197	0.081
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.080	0.198	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAMES R HESTER  
Signature and Printed Name

02/25/2021  
Date