

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 09:10

Date of Inspection: 10/28/2021

Serial Number: 80-001253
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24620080A2 Exp: 11/05/2022
0.000	0.049	0.080	0.198	0.081
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.200	0.080

Number of Simulators Used: 5

Remarks:

Int Det: .

Observed Interferent Detect test error. Checked all connections of simulator with no obstructions or deficiencies observed. Ran test again with no errors observed.

[Signature]
Gaston A Robin, #3877

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] - Gaston A Robin, #3877
Signature and Printed Name

10/28/2021
Date