

(Amended)

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCOA P.D.

Time of Inspection: 20:21

Date of Inspection: 07/01/2021

Serial Number: 80-001260

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000				
0.000				
0.000				

Number of Simulators Used: 5

Remarks:

LOOSE CONNECTION, . NOT SHOWING INTERFERENT.

Made a new solution from different bottle of acetone stock solution, and it tested correctly. (Sec Inspection at 20:21 hours.)

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAMES H FRAZIER

Signature and Printed Name

Frazier #161

07/01/2021
Date

Amended
**Florida Department of Law Enforcement
 Alcohol Testing Program**

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCOA P.D.
 Time of Inspection: 22:05

Date of Inspection: 03/18/2021

Serial Number: 80-001260
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A4 Exp: 01/05/2022
0.000	/ 0.050	0.081	0.192	0.078
0.000	/ 0.051	0.081	0.203	0.079
0.000	/ 0.050	0.080	0.202	0.079

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded. 05: Ambient Fail.

see attached note.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Christopher E Bradshaw #75 Bradshaw

CHRISTOPHER E BRADSHAW

Signature and Printed Name

03/18/2021
Date