

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD
Time of Inspection: 00:37

Date of Inspection: 02/04/2021

Serial Number: 80-001306
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17919080A1 Exp: 08/05/2021
0.000	0.047	0.082	0.198	0.081
0.000	0.047	0.082	0.201	0.081
0.000	0.048	0.082	0.201	0.081

Number of Simulators Used: 4

Remarks:

Int Det: .

Checked simulator seal and all connections. Re-run test. - JLR

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Thomas C. Roberson

THOMAS C ROBERSON

Signature and Printed Name

02/04/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD
Time of Inspection: 14:03

Date of Inspection: 01/04/2021

Serial Number: 80-001306
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1298240 Exp: 03/19/2023
0.000	0.046	0.081	0.196	0.085 / 0.082
0.000	0.047	0.082	0.198	0.086 / 0.082
0.000	0.048	0.082	0.199	0.086 / 0.082

Number of Simulators Used: 4

Lot # 17919080A1
Exp 8/5/2021

Remarks:

08: Control Outside Tolerance.

Supply cylinder expended during inspection. Replaced cylinder and reran test.
- JLR

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Thomas C. Roberson

THOMAS C ROBERSON

Signature and Printed Name

01/04/2021
Date