

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PATRICK AFB SFS
Time of Inspection: 23:59

Date of Inspection: 10/31/2021

Serial Number: 80-001471
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:25721080A2 Exp: 11/05/2023
0.000	0.048	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.050	0.080	0.200	0.077

Number of Simulators Used: 5

Remarks:

DATE OF INSPECTION: 10/31/2021 AND THE START TIME WAS 23:20 HOURS.

[Signature] Dep. Robin, G. #3877

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] GASTON A ROBIN
Signature and Printed Name

10/31/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PATRICK AFB SFS
Time of Inspection: 00:46

Date of Inspection: 11/01/2021

Serial Number: 80-001471
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#: 202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 25721080A2 Exp: 11/05/2023
0.000	0.048	0.079	0.198	0.078
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.078

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded.

The alcohol free / mouth alcohol test was run with range exceeded notification. Test was redone without any issues.

[Signature] Sgt. Robin, G. #3877

* DATE OF INSPECTION: 11/01/2021 AND THE START TIME WAS 0006 HOURS.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] Sgt. Robin, G. #3877

GASTON A ROBIN

Signature and Printed Name

11/01/2021
Date