

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TYNDALL AFB

Time of Inspection: 10:58

Date of Inspection: 02/25/2021

Serial Number: 80-001646

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Sed Test (g/210L) Lot#:270670 Exp: 06/07/2021
0.000	0.046	0.078	0.197	0.080
0.000	0.047	0.078	0.198	0.081
0.000	0.047	0.078	0.197	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, F.A.C.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector's Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, F.A.C.



LANNY A LORD  
Signature and Printed Name

02/25/2021  
Date