

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.  
Time of Inspection: 17:22

Date of Inspection: 12/27/2021

Serial Number: 80-001721  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:24620080A2 Exp: 11/05/2022	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048	0.000 / 0.000		
0.000	0.048	0.000 / 0.000		
0.000	0.048	0.000 / 0.000		

Number of Simulators Used: 5

Remarks:


08 Control Outside Tolerance, Control Outside Tolerance. Non-compliance:

**OPERATOR ERROR: MISTAKENLY BELIEVED INSTRUMENT WAS ON .08 DRY GAS, WHICH WAS ATTACHED.**

**REPEATING INSPECTION.**

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



EDWIN J LUTZ

Signature and Printed Name

12/27/2021  
Date

**CARLY ROSAS  
TPD**