

*Amended*

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD  
Time of Inspection: 13:55

Date of Inspection: 04/01/2021

Serial Number: 80-002097  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A3 Exp: 03/05/2021
0.000	0.046	0.077	0.192	0.081
0.000	0.046	0.078	0.194	0.082
0.000	0.046	0.079	0.194	0.082

Number of Simulators Used: 5

Remarks:

*Agency inspection was NOT compliant due to expired Dry Gas, used.*

The above instrument complies ( ~~✓~~ ) does not comply ( *XND* ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]*  
Signature and Printed Name

LUIS H JUSINO

04/01/2021  
Date

*Amended*

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD  
Time of Inspection: 21:04

Date of Inspection: 01/24/2021

Serial Number: 80-002097  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:19390 Exp: 12/30/2021	0.08g/210L Test (g/210L) Lot#:20330 Exp: 07/10/2022	0.20g/210L Test (g/210L) Lot#:20160 Exp: 03/18/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A3 Exp: 03/05/2021
0.000	0.046	0.077	0.136 / 0.194	0.082
0.000	0.046	0.077	0.189 / 0.196	0.082
0.000	0.047	0.078	0.194 / 0.195	0.082

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance.

*Inspection was non-compliant due to the use of non-approved solutions.*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC. *7 May 21*

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Andrew E. Wright* ANDREW E WRIGHT  
Signature and Printed Name

01/24/2021  
Date