

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 18:48

Date of Inspection: 01/19/2021

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22419080A3 Exp: 10/05/2021
0.000 / 0.000	0.048	0.078	0.195	0.081
0.000 / 0.000	0.048	0.079	0.196	0.080
/ 0.000	0.048	0.078	0.196	0.081

Number of Simulators Used: 4

Remarks:

00: RFI Detect-RADIO IN ROOM.

Another officer came into room w/o turning off equipment
Remove all equipment and report
JW 7572
1/19/21

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BARRY NIXON

Signature and Printed Name

01/19/2021
Date