

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 13:33

Date of Inspection: 07/09/2021

Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:11821080A2 Exp: 06/05/2023
0.000	0.049	0.082	0.201	0.078
0.000	0.049	0.081	0.205	0.078
0.000	0.047	0.080	0.206	0.079

Number of Simulators Used: 5

Remarks: **.05g*

HOSE AIR TIGHT SEAL CHECKED AND CONFIRMED, RECONNECTED, MET STANDARD. 9-14-21 CA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



CAMILO J ALMEIDA

Signature and Printed Name

07/09/2021
Date