

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: HENDRY COUNTY SO
Time of Inspection: 09:58

Date of Inspection: 06/28/2023

Serial Number: 80-006386
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No *
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1474747 Exp: 04/12/2024
0.000	0.047	0.076	0.198	0.080
0.000	0.047	0.077	0.199	0.080
0.000	0.047	0.077	0.198	0.080

Number of Simulators Used: 5


Remarks:

Non-compliance:

* I was doing two instruments at one time and started the second one (this one) too soon. On the Alcohol Free Test before all of the mouth alcohol from the previous instrument was gone from my mouth. The instrument registered a slight amount of alcohol. Spoke with the department inspector and completed another inspection.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

06/28/2023
Date