

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Serial Number: 80-006472

Time of Inspection: 13:26

Date of Inspection: 11/17/2021

Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | Yes | |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 202101C Exp: 01/12/2023 | 0.08g/210L Test (g/210L) Lot#: 202007A Exp: 07/21/2022 | 0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 1355293 Exp: 07/29/2023 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.078 | 0.196 | 0.078 |
| 0.000 | 0.049 | 0.078 | 0.195 | 0.078 |
| 0.000 | 0.049 | 0.079 | 0.195 | 0.079 |

Number of Simulators Used: 5

Remarks:

Time-Date changed. A F / M A: RFI Detect.

Repeat done due to RFI *Marlon Boggan*
12/6/21

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MARLON M BOGGAN

Signature and Printed Name

11/17/2021

Date

