

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: HENDRY COUNTY SO
Time of Inspection: 11:31

Date of Inspection: 12/15/2021

Serial Number: 80-006234
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1474747 Exp: 04/12/2024
/ 0.000	0.047	0.076	0.199	0.083
/ 0.000	0.047	0.078	0.200	0.082
/ 0.000	0.048	0.079	0.201	0.082

Number of Simulators Used: 5

Remarks:

00: Ambient Fail.

** I walked by the instrument as the test began. I spoke with FDLE Department Inspector, and believe the issue was my Cologne too close to Instrument. The Room was Aired out and Test restarted without issue.
12/15/2021 JH⁵¹²*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JH⁵¹²

Signature and Printed Name

JAMES R HESTER

12/15/2021
Date