

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 10:04

Date of Inspection: 10/18/2021

Serial Number: 80-~~000000~~ ⁰⁹⁶⁷⁵⁸
Software: 8100.27

TL
10/18/21

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 111:579 Exp: 03/19/2022
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.049	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks: Instrument S/N, NOT displaying on Form 40. Per Dept. Inspector Taylor Gutschow, Will upload & send to FOLE Inspection.
(TL) 10/18/21

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

TPR. *[Signature]* #3892

TAYLOR M LEDFORD

Signature and Printed Name

10/18/2021
Date