

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 01:44

Date of Inspection: 11/13/2021

Serial Number: 80-007439
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09820080A3 Exp: 06/05/2022
0.000	0.048	0.077	0.198	0.079
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks:

Alcohol Testing Program

NOV 23 2021

RECEIVED
FDLE

RECEIVED

NOV 15 2021

Records Department
Baker County Sheriffs Office

RECEIVED
FDLE
NOV 23 2021
Alcohol Testing Program

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

LTC Coll 909

CALEB A COLLINS

Signature and Printed Name

11/13/2021
Date