

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 10:44

Date of Inspection: 10/29/2021

Serial Number: 80-007505  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22620080A3 Exp: 10/05/2022
0.000	0.046	0.077	0.196	0.080
0.000	0.047	0.078	0.198	0.080
0.000	0.047	0.078	0.198	0.080

Number of Simulators Used: 5

Remarks:

**A TRUE COPY**  
**PAUL BLACKMAN, SHERIFF**  
**HIGHLANDS COUNTY, FLORIDA**  
BY Shon W. Abeln 2184 D.S.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184 SHON W ABELN  
Signature and Printed Name

10/29/2021  
Date