



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Lakeland PD

Instrument Serial Number: 80-003945

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Camilo Almeida
Date of Inspection: 5/07/2021 and 5/20/2021
Time of Inspection: 09:09:19 (5/7/21) 09:55:26 (5/7/21) 15:19:43 (5/20/21)
Agency Inspection Discrepancy: [] Incomplete [] Untimely/Not Received [] Erroneous Information [] Procedural [x] Other (Missing Required Information)
[] Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
[] Lot Number [] Expiration Date for ___g/ 210L [] Alcohol Reference Solution [] Dry Gas Standard is [] Incorrect [] Expired.
[x] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [x] REASON for repeating the following test(s); OR the [x] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[] Alcohol Free Subject Test [] Mouth Alcohol Test [] Alcohol Free Test [] Interferent Detect Test
[x] 0.05 g/210L Test [x] 0.08 g/210L Test [] 0.20 g/210L Test [] 0.08 g/210L Dry Gas Standard Test
[] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[] Other: _____

OTHER ELECTRONIC DATA REVIEW
[] Login Records Date: _____
[] Cylinder Change Records Date: _____
[] Control Test Records Date: _____
[] Diagnostic Check Records Date: _____
Comments:
On all of the listed Agency Inspections, the 0.05 and/or 0.08 g/210 L Tests were repeated. The reason for repeating the tests, as well as any corrective action taken prior to repeating them, must be included.

CORRECTIVE ACTION
[x] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 9/24/2021.
[] Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
[] Upload the Agency Inspection(s).
[] Remove the instrument from evidentiary use until otherwise directed by the Department.
[] No action required
[] Other: _____

Signature of Alcohol Testing Program Staff Member

8/24/2021 Date