



INSTRUMENT PROCESSING SHEET

Agency FDLES/N 80-000224Florida Department of
Law EnforcementDate In n/aDI Completion Date 5/4/2021☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>TDG</u>	Quality Checks	By <u>TDG</u> Date <u>5/4/2021</u>	Flow Calibration	By _____ Date _____																																																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Inspection conducted prior to loaning to Polk CSO</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>227</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5092</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705 09/23/2022</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A 10/05/2022	0.080	MP5093	202010B 10/05/2022	0.200	MP5094	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ DI Temp. Checks By <u>MH</u> ☒ Lab Temp °C <u>22.13</u> External Digital Therm. ID#: <u>300504</u> ☒ 34°C +/- .2 Serial #: <u>MP5092</u> ☒ 34°C +/- .2 Serial #: <u>MP5093</u> ☒ 34°C +/- .2 Serial #: <u>MP5094</u>																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP5092	202010A 10/05/2022																																																															
0.080	MP5093	202010B 10/05/2022																																																															
0.200	MP5094	202010D 10/06/2022																																																															
0.080 DGS	N/A	AG026705 09/23/2022																																																															
Calibration Adjustment By _____ Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>TDG</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1016</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>SD1017</td></tr> <tr><td>Interferent</td><td>SD3966</td></tr> <tr><td>0.050</td><td>MP5092</td></tr> <tr><td>0.080</td><td>MP5093</td></tr> <tr><td>0.200</td><td>MP5094</td></tr> </tbody> </table>				Simulator	Serial Number	0.000	SD1017	Interferent	SD3966	0.050	MP5092	0.080	MP5093	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	SD1017																																																																
Interferent	SD3966																																																																
0.050	MP5092																																																																
0.080	MP5093																																																																
0.200	MP5094																																																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2021.05.04 <i>Michael D. Haughey</i> 12:03:37 -04'00' 2021.05.04 <i>SR</i> 4 13:33:38 04'00' Tech Review / Date Admin Review / Date																																																															

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 09:39

Date of Inspection: 05/04/2021

Serial Number: 80-000224

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

AI BYPASSED. COMPLIANCE NOT DETERMINED.

Not determined *TMG*
5/4/2021

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

05/04/2021
Date

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
05/04/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.049	09:44
Air Blank	0.000	09:45
Control Test	0.049	09:46
Air Blank	0.000	09:46
Control Test	0.049	09:47
Air Blank	0.000	09:48
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 11:32

Date of Inspection: 05/04/2021

Serial Number: 80-000224

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#: 202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#: 202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG026705 Exp: 09/23/2022
0.000	0.050	0.080	0.203	0.079
0.000	0.050	0.081	0.203	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.050	0.081	0.202	0.078
0.000	0.050	0.080	0.203	0.078
0.000	0.051	0.081	0.203	0.078
0.000	0.051	0.081	0.202	0.079
0.000	0.051	0.081	0.203	0.079
0.000	0.051	0.081	0.202	0.079
0.000	0.051	0.081	0.202	0.078


Standard Deviations	0.0005	0.0004	0.0006	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

05/04/2021
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000224, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000224</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>05/04/2021</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:32</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/04/2021

Date

Taylor D Gutschow

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality