



Lafayette County 1-27-2021

INSTRUMENT PROCESSING SHEET

Agency Lafayette County Sheriff's Office

S/N 80-000773

Florida Department of
Law Enforcement

Date In 12/29/2020 DI Completion Date 01-22-2021

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>RAW</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		Quality Checks Performed By <u>RAW</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>191</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.230</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2021.01.28 10:45:35 -05'00'		Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____		Temperature Checks Performed By <u>IS</u> ☒ Lab Temp °C <u>21.03</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C +/- .2 Serial #: <u>MP5088</u> ☒ 34°C +/- .2 Serial #: <u>MP5086</u> ☒ 34°C +/- .2 Serial #: <u>MP5090</u>																																									
Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____		Department Inspection Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1013</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u>																																											
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G11621</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	G11621	Interferent	MP5087	0.050	MP5088	0.080	MP5086	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	G11621																																												
Interferent	MP5087																																												
0.050	MP5088																																												
0.080	MP5086																																												
0.200	MP5090																																												
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																											
Notes/Suggested Service: <u>Tech Review: corrected agency name</u> <u>1-27-2021</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2021.01.																																											
David Eliezer Reyes-Rivera Digitally signed by David Eliezer Reyes-Rivera Date: 2021.01.27 15:15:33 -05'00'		28 Tech Review / Date <u>10:45:12</u> Admin Review / Date <u>05'00'</u>																																											

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAFAYETTE CO SO
Time of Inspection: 10:11

Date of Inspection: 01/22/2021

Serial Number: 80-000773
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.200	0.079

Standard Deviations	0.0003	0.0000	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 ISRAEL SOTO
Signature and Printed Name

01/22/2021
Date

DERR Digitally signed
by DERR
Date: 2021.01.27
15:14:52 -05'00'

 2021.01.
28
10:44:56
-05'00'

Stability Checks

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000773
01/21/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:08
Control Test	0.048	15:08
Air Blank	0.000	15:09
Control Test	0.049	15:09
Air Blank	0.000	15:10
Control Test	0.049	15:11
Air Blank	0.000	15:11
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

0.05 g/210L

PAW
Operator's Signature

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000773
01/21/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:13
Control Test	0.079	15:14
Air Blank	0.000	15:14
Control Test	0.078	15:15
Air Blank	0.000	15:16
Control Test	0.079	15:16
Air Blank	0.000	15:17
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

0.08 g/210L
wet

PAW
Operator's Signature

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000773
01/21/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:19
Control Test	0.198	15:20
Air Blank	0.000	15:21
Control Test	0.198	15:21
Air Blank	0.000	15:22
Control Test	0.198	15:23
Air Blank	0.000	15:23
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.20 g/210L

PAW
Operator's Signature

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000773
01/21/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:25
Control Test	0.079	15:25
Air Blank	0.000	15:25
Control Test	0.080	15:26
Air Blank	0.000	15:26
Control Test	0.079	15:27
Air Blank	0.000	15:27
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

0.08 g/210L
DGS

PAW
Operator's Signature

Digitally signed
by DERR
Date:
2021.01.27
15:14:17 -05'00'

DERR

2021.01
.28
10:44:0
7 -05'00'



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000773, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000773	UNCERTAINTY* $\pm$	
Owning Agency:	LAFAYETTE CO SO	0.050 g/ 210 L	0.005
Calibration Date:	01/22/2021	0.080 g/ 210 L	0.004
Calibration Time:	10:11	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/22/2021

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Digitally signed
by DERR
Date: 2021.01.27
15:13:29 -05'00'

DERR
2021.01.
28
10:38:44
-05'00'