



INSTRUMENT PROCESSING SHEET

Agency NAS Jacksonville

S/N 80-000778

Florida Department of
Law Enforcement

Date In 01/12/2021

DI Completion Date 01-22-2021

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>RAW</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____		Quality Checks Performed By <u>IS</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>201</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.167</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.523</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2021.01.28 11:21:07 -05'00'		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A 10-05-2022</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> <td>202010B 10-05-2022</td> </tr> <tr> <td>0.200</td> <td>G2078</td> <td>202010D 10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102 04-20-2022</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	202010A 10-05-2022	0.080	MP5086	202010B 10-05-2022	0.200	G2078	202010D 10-06-2022	0.080 DGS	N/A	AG011102 04-20-2022	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>IS</u> ☒ Lab Temp °C <u>21.33</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C +/- .2 Serial #: <u>MP5088</u> ☒ 34°C +/- .2 Serial #: <u>MP5086</u> ☒ 34°C +/- .2 Serial #: <u>MP5090</u>																								
Simulator	Serial #	Lot #/Exp																																									
0.050	MP5088	202010A 10-05-2022																																									
0.080	MP5086	202010B 10-05-2022																																									
0.200	G2078	202010D 10-06-2022																																									
0.080 DGS	N/A	AG011102 04-20-2022																																									
Calibration Adjustment Performed By <u>IS</u> Barometric Pressure Gauge <u>1026</u> ID # <u>28421</u>		Department Inspection Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1014</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>20060</td> <td>02-10-2022</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20190</td> <td>04-06-2022</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>03-18-2022</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>01-21-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>24620080A3</td> <td>11-05-2022</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	02-10-2022	0.100	MP5083	20190	04-06-2022	0.200	MP5084	20160	03-18-2022	0.300	MP5085	20030	01-21-2022	0.080 DGS	N/A	24620080A3	11-05-2022	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G11621</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	G11621	Interferent	MP5087	0.050	MP5088	0.080	MP5086	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	MP5091	N/A	N/A																																								
0.040	MP5082	20060	02-10-2022																																								
0.100	MP5083	20190	04-06-2022																																								
0.200	MP5084	20160	03-18-2022																																								
0.300	MP5085	20030	01-21-2022																																								
0.080 DGS	N/A	24620080A3	11-05-2022																																								
Simulator	Serial Number																																										
0.000	G11621																																										
Interferent	MP5087																																										
0.050	MP5088																																										
0.080	MP5086																																										
0.200	MP5090																																										
☒ Post Calibration Adjustment Stability Checks		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A</td> <td>10-05-2022</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> <td>202010B</td> <td>10-05-2022</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D</td> <td>10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102</td> <td>04-20-2022</td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.050	MP5088	202010A	10-05-2022	0.080	MP5086	202010B	10-05-2022	0.200	MP5090	202010D	10-06-2022	0.080 DGS	N/A	AG011102	04-20-2022																		
Simulator	Serial Number	Lot Number	Expiration																																								
0.050	MP5088	202010A	10-05-2022																																								
0.080	MP5086	202010B	10-05-2022																																								
0.200	MP5090	202010D	10-06-2022																																								
0.080 DGS	N/A	AG011102	04-20-2022																																								
Notes/Suggested Service: <u>Dry gas stability values were not nominal, performed optical bench calibration adjust.</u>		Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☒ Calibration Adjustment ☐ Other _____																																									
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																									
		2021.01.26 08:47:30 -05'00' Michael D. Haughey Tech Review / Date																																									
		2021.01.28 11:20:43 Admin. Review / Date																																									

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD
Time of Inspection: 13:01

Date of Inspection: 01/22/2021

Serial Number: 80-000778
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.198	0.080
Standard Deviations	0.0004	0.0004	0.0005	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 ISRAEL SOTO
Signature and Printed Name

01/22/2021
Date


 2021.01.
28
11:20:26
-05'00'

Stability Checks

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
01/13/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:00
Control Test	0.049	12:01
Air Blank	0.000	12:01
Control Test	0.050	12:02
Air Blank	0.000	12:02
Control Test	0.051	12:03
Air Blank	0.000	12:03
Control Test Stats		
Average	0.0500	
Std Dev	0.0010	
Rel Std Dev(%)	2.0000	

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
01/13/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:05
Control Test	0.079	12:06
Air Blank	0.000	12:06
Control Test	0.080	12:07
Air Blank	0.000	12:07
Control Test	0.079	12:08
Air Blank	0.000	12:09
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

wet

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
01/13/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:10
Control Test	0.200	12:11
Air Blank	0.000	12:11
Control Test	0.201	12:12
Air Blank	0.000	12:12
Control Test	0.200	12:13
Air Blank	0.000	12:13
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
01/13/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:14
Control Test	0.085	12:15
Air Blank	0.000	12:15
Control Test	0.084	12:15
Air Blank	0.000	12:16
Control Test	0.084	12:16
Air Blank	0.000	12:17
Control Test Stats		
Average	0.0843	
Std Dev	0.0006	
Rel Std Dev(%)	0.6846	

Dry

Operator's Signature

2021.01.
28
11:20:12
-05'00"



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000778, manufactured by CML, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000778</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>NAS JACKSONVILLE PD</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>01/22/2021</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:01</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/22/2021

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 January 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

2021.01.
28
11:19:58
-05'00'

Optical Bench Cal. Adj.

SS

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
01/19/2021 14:27:15

Auto Calibration
Max Power Res Value = 22
Auto Range Res Value = 10

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum lo = 12719, Sum lo = 13135

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1440 (-0.0110)
Sample #2 = 0.1550 (-0.0200)
Sample #3 = 0.1520 (0.0280)
Sample #4 = 0.1260 (0.0360)
Avg % Abs = 0.1443 (0.0147)
STD DEV = 0.0159 (0.0303)
REL STD DEV = 11.049 (206.506)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1250 (-0.0030)
Sample #2 = 0.1400 (-0.0140)
Sample #3 = 0.1410 (0.0030)
Sample #4 = 0.1180 (0.0130)
Avg % Abs = 0.1330 (0.0007)
STD DEV = 0.0130 (0.0137)
REL STD DEV = 9.774 (2047.559)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum lo = 12705, Sum lo = 13125

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.9600 (-0.0430)
Sample #2 = 0.9270 (-0.0220)
Sample #3 = 0.9270 (-0.0140)
Sample #4 = 0.8940 (0.0260)
Avg % Abs = 0.9160 (-0.0033)
STD DEV = 0.0191 (0.0257)
REL STD DEV = 2.080 (771.492)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5930 (-0.0270)
Sample #2 = 1.5940 (-0.0360)
Sample #3 = 1.5600 (-0.0210)
Sample #4 = 1.5880 (-0.0220)
Avg % Abs = 1.5807 (-0.0263)
STD DEV = 0.0181 (0.0084)
REL STD DEV = 1.148 (31.847)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum lo = 12699, Sum lo = 13125

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 2.0880 (-0.0280)
Sample #2 = 2.0410 (-0.0020)
Sample #3 = 2.0580 (0.0130)
Sample #4 = 2.0370 (0.0490)
Avg % Abs = 2.0453 (0.0200)
STD DEV = 0.0112 (0.0262)
REL STD DEV = 0.545 (131.053)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6650 (-0.0060)
Sample #2 = 3.6310 (0.0070)
Sample #3 = 3.6400 (0.0090)
Sample #4 = 3.6370 (0.0240)
Avg % Abs = 3.6360 (0.0133)
STD DEV = 0.0046 (0.0093)
REL STD DEV = 0.126 (69.687)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum lo = 12687, Sum lo = 13117

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.8770 (-0.0230)
Sample #2 = 3.9310 (-0.0300)
Sample #3 = 3.9050 (-0.0260)
Sample #4 = 3.8800 (0.0070)
Avg % Abs = 3.9053 (-0.0163)
STD DEV = 0.0255 (0.0203)
REL STD DEV = 0.653 (124.322)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.9480 (-0.0110)
Sample #2 = 6.9630 (-0.0240)
Sample #3 = 6.9540 (-0.0040)
Sample #4 = 6.9650 (0.0030)
Avg % Abs = 6.9640 (-0.0083)
STD DEV = 0.0010 (0.0140)
REL STD DEV = 0.014 (168.143)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum lo = 12683, Sum lo = 13115

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.6400 (-0.0330)
Sample #2 = 5.6000 (-0.0260)
Sample #3 = 5.6150 (-0.0070)
Sample #4 = 5.6230 (0.0060)
Avg % Abs = 5.6127 (-0.0090)
STD DEV = 0.0117 (0.0161)
REL STD DEV = 0.208 (178.816)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 10.0450 (-0.0370)
Sample #2 = 9.9440 (-0.0150)
Sample #3 = 9.9720 (0.0030)
Sample #4 = 9.9950 (0.0030)
Avg % Abs = 9.9703 (-0.0030)
STD DEV = 0.0255 (0.0104)
REL STD DEV = 0.256 (345.410)

***** AUTO CAL DATA *****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.144
Std Dev = 0.02 Rel Std Dev = 11.05
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.916
Std Dev = 0.02 Rel Std Dev = 2.08
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 2.045
Std Dev = 0.01 Rel Std Dev = 0.55
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.905
Std Dev = 0.03 Rel Std Dev = 0.65
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.613
Std Dev = 0.01 Rel Std Dev = 0.21
Zero Order Coef = -324.19
First Order Coef = 2387.78
Second Order Coef = 37.73
Standard Deviation = 37.642040

<<<< CHANNEL 2 >>>>
Sol Val = 0.0300 mg/l or 0.000 g/210L
% Abs = 0.133
Std Dev = 0.01 Rel Std Dev = 9.77
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.581
Std Dev = 0.02 Rel Std Dev = 1.15
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.636
Std Dev = 0.00 Rel Std Dev = 0.13
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.964
Std Dev = 0.00 Rel Std Dev = 0.01
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.970
Std Dev = 0.03 Rel Std Dev = 0.26
Zero Order Coef = -158.67
First Order Coef = 1280.63
Second Order Coef = 16.69
Standard Deviation = 33.094650

2021.01.28
11:19:37 -05'00'

MX

Optical Bench Cal Adj. *SD*

Post Stability Checks

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0004
0.040	0.040	0.0002
0.100	0.099	0.0009
0.200	0.201	-0.0011
0.300	0.300	0.0004

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	-0.0001
0.100	0.099	0.0009
0.200	0.201	-0.0010
0.300	0.300	0.0004

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 2805.00
 Sample #2 = 2802.00
 Sample #3 = 2746.00
 Sample #4 = 2812.00
 Average Result = 2786.6667
 STD DEV = 35.5715
 REL STD DEV = 1.276

 ***** CHANNEL 2
 Sample #1 = 3261.00
 Sample #2 = 3200.00
 Sample #3 = 3203.00
 Sample #4 = 3207.00
 Average Result = 3203.3333
 STD DEV = 3.5119
 REL STD DEV = 0.110

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1026
 3 in H2O Adjust (mg/l*10,000) = 1023
 9 in H2O Adjust (mg/l*10,000) = 606
 **** AUTO CAL PASS

NAS JACKSONVILLE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000778
 01/19/2021
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:05
Control Test	0.048	16:06
Air Blank	0.000	16:07
Control Test	0.048	16:07
Air Blank	0.000	16:08
Control Test	0.048	16:09
Air Blank	0.000	16:09
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SD
 Operator's Signature

NAS JACKSONVILLE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000778
 01/19/2021
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:15
Control Test	0.198	16:16
Air Blank	0.000	16:17
Control Test	0.198	16:17
Air Blank	0.000	16:18
Control Test	0.198	16:18
Air Blank	0.000	16:19
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SD
 Operator's Signature

NAS JACKSONVILLE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000778
 01/19/2021
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:10
Control Test	0.078	16:11
Air Blank	0.000	16:11
Control Test	0.079	16:12
Air Blank	0.000	16:13
Control Test	0.078	16:13
Air Blank	0.000	16:14
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

wet

SD
 Operator's Signature

NAS JACKSONVILLE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000778
 01/19/2021
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:21
Control Test	0.080	16:21
Air Blank	0.000	16:21
Control Test	0.079	16:22
Air Blank	0.000	16:22
Control Test	0.079	16:23
Air Blank	0.000	16:23
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Dry

SD
 Operator's Signature

2021.01.
 28
 11:19:24
 -05'00"

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD
Time of Inspection: 14:13

Date of Inspection: 01/12/2021

Serial Number: 80-000778
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

N/A compliance not determined

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

01/12/2021
Date

MX
SR 2021.01.
28
11:19:11
-05'00"