



## INSTRUMENT PROCESSING SHEET

Agency FHP Sneads Police DepartmentS/N 80-000781Florida Department of  
Law EnforcementDate In 01/19/2021 DI Completion Date 01-22-2021☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

<b>Intake</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____		<b>Quality Checks</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>201</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.523</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks		<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
<b>Final Release Date</b> FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2021.01.29 11:46:05 -05'00'		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____		<b>Temperature Checks</b> Performed By <u>IS</u> <input checked="" type="checkbox"/> Lab Temp °C <u>20.69</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5086</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5090</u>																																							
<b>Calibration Adjustment</b> Performed By <u>IS</u> Barometric Pressure Gauge <u>1026</u> ID # <u>28421</u>		<b>Department Inspection</b> Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1012</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>20060</td> <td>02-10-2022</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20190</td> <td>04-06-2022</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>03-18-2022</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>01-21-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>24620080A3</td> <td>11-05-2022</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	02-10-2022	0.100	MP5083	20190	04-06-2022	0.200	MP5084	20160	03-18-2022	0.300	MP5085	20030	01-21-2022	0.080 DGS	N/A	24620080A3	11-05-2022	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G11621</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	G11621	Interferent	MP5087	0.050	MP5088	0.080	MP5086	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	MP5091	N/A	N/A																																								
0.040	MP5082	20060	02-10-2022																																								
0.100	MP5083	20190	04-06-2022																																								
0.200	MP5084	20160	03-18-2022																																								
0.300	MP5085	20030	01-21-2022																																								
0.080 DGS	N/A	24620080A3	11-05-2022																																								
Simulator	Serial Number																																										
0.000	G11621																																										
Interferent	MP5087																																										
0.050	MP5088																																										
0.080	MP5086																																										
0.200	MP5090																																										
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks		<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																									
Notes/Suggested Service: <u>Admin Review: Corrected agency name - 1-29-2021</u> <u>and added calibration certificate.</u>		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place Into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use 2021.01.26 08:43:28 -05'00' <u>Michael D. Hagley</u> Tech Review / Date 2021.01.29 11:45:17 <u>CR</u> Admin Review / Date																																									

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 14:12

Date of Inspection: 01/22/2021

Serial Number: 80-000781

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.050	0.079	0.201	0.078
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.078
0.000	0.050	0.079	0.202	0.078
0.000	0.050	0.079	0.202	0.078
0.000	0.050	0.079	0.201	0.078
0.000	0.050	0.079	0.201	0.079
0.000	0.051	0.079	0.201	0.078
0.000	0.050	0.079	0.201	0.079

Standard Deviations	0.0006	0.0000	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

*MX*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Israel Soto*

ISRAEL SOTO

Signature and Printed Name

01/22/2021

Date

# Stability Checks

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000781  
01/19/2021  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:09
Control Test	0.051	11:10
Air Blank	0.000	11:10
Control Test	0.050	11:11
Air Blank	0.000	11:12
Control Test	0.051	11:12
Air Blank	0.000	11:13
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

0.05g/210L

*PAW*  
Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000781  
01/19/2021  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:14
Control Test	0.082	11:15
Air Blank	0.000	11:16
Control Test	0.082	11:16
Air Blank	0.000	11:17
Control Test	0.081	11:18
Air Blank	0.000	11:18
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

0.08g/210L  
wet

*PAW*  
Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000781  
01/19/2021  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:24
Control Test	0.207	11:25
Air Blank	0.000	11:25
Control Test	0.204	11:26
Air Blank	0.000	11:27
Control Test	0.203	11:27
Air Blank	0.000	11:28
Control Test Stats		
Average	0.2047	
Std Dev	0.0021	
Rel Std Dev(%)	1.0171	

0.20g/210L

*PAW*  
Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000781  
01/19/2021  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.003	11:30
Control Test	0.080	11:30
Air Blank	0.000	11:31
Control Test	0.081	11:31
Air Blank	0.000	11:32
Control Test	0.081	11:32
Air Blank	0.000	11:32
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

0.08g/210L  
DGS

*PAW*  
Operator's Signature



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000781, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000781</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>01/22/2021</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:12</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

*Israel Soto*

01/22/2021

Date

ISRAEL SOTO,  
Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

# Optical Bench Cal. Adjust.

FHP

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000781

01/19/2021

13:03:57

Auto Calibration

Max Power Res Value = 45

Auto Range Res Value = 29

Sol Value = 0.000 g/210L \*\*\*

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12530, Sum Io = 13082

CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1100	(-0.0130)
Sample #2 =	0.0930	(0.0180)
Sample #3 =	0.1290	(0.0140)
Sample #4 =	0.1480	(0.0110)
Avg % Abs =	0.1233	(0.0143)
STD DEV =	0.0279	(0.0035)
REL STD DEV =	22.650	(24.502)

CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1400	(-0.0180)
Sample #2 =	0.1170	(-0.0030)
Sample #3 =	0.1280	(-0.0060)
Sample #4 =	0.1250	(-0.0030)
Avg % Abs =	0.1233	(-0.0040)
STD DEV =	0.0057	(0.0017)
REL STD DEV =	4.610	(43.301)

Sol Value = 0.040 g/210L \*\*\*

Fit value = 0.1905 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12530, Sum Io = 13082

CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.9430	(-0.0180)
Sample #2 =	0.9100	(0.0310)
Sample #3 =	0.9100	(0.0390)
Sample #4 =	0.9000	(0.0250)
Avg % Abs =	0.9067	(0.0317)
STD DEV =	0.0058	(0.0070)
REL STD DEV =	0.637	(22.180)

CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.5870	(0.0020)
Sample #2 =	1.5790	(0.0230)
Sample #3 =	1.5640	(0.0330)
Sample #4 =	1.5590	(0.0300)
Avg % Abs =	1.5673	(0.0287)
STD DEV =	0.0104	(0.0051)
REL STD DEV =	0.664	(17.931)

Sol Value = 0.100 g/210L \*\*\*

Fit value = 0.4762 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12524, Sum Io = 13078

CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	2.0360	(-0.0160)
Sample #2 =	2.0150	(0.0180)
Sample #3 =	1.9910	(0.0530)
Sample #4 =	1.9810	(0.0440)
Avg % Abs =	1.9957	(0.0383)
STD DEV =	0.0175	(0.0162)
REL STD DEV =	0.875	(47.413)

CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.6760	(-0.0140)
Sample #2 =	3.6400	(0.0320)
Sample #3 =	3.5970	(0.0650)
Sample #4 =	3.5990	(0.0480)
Avg % Abs =	3.6120	(0.0483)
STD DEV =	0.0243	(0.0165)
REL STD DEV =	0.672	(34.143)

Sol Value = 0.200 g/210L \*\*\*

Fit value = 0.9524 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12525, Sum Io = 13080

CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.8530	(-0.0210)
Sample #2 =	3.7670	(0.0350)
Sample #3 =	3.7660	(0.0500)
Sample #4 =	3.7500	(0.0900)
Avg % Abs =	3.7510	(0.0583)
STD DEV =	0.0095	(0.0234)
REL STD DEV =	0.254	(48.739)

CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	7.0560	(-0.0030)
Sample #2 =	6.9340	(0.0990)
Sample #3 =	6.9060	(0.1270)
Sample #4 =	6.9010	(0.1440)
Avg % Abs =	6.9137	(0.1233)
STD DEV =	0.0178	(0.0227)
REL STD DEV =	0.257	(18.424)

Sol Value = 0.300 g/210L \*\*\*

Fit value = 1.4286 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12517, Sum Io = 13673

CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	5.4910	(-0.0110)
Sample #2 =	5.4570	(0.0470)
Sample #3 =	5.4660	(0.0540)
Sample #4 =	5.4090	(0.0870)
Avg % Abs =	5.4440	(0.0627)
STD DEV =	0.0306	(0.0214)
REL STD DEV =	0.563	(34.088)

CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	10.0840	(-0.0130)
Sample #2 =	9.9560	(0.1380)
Sample #3 =	9.9410	(0.1560)
Sample #4 =	9.9060	(0.1700)
Avg % Abs =	9.9343	(0.1547)
STD DEV =	0.0257	(0.0160)
REL STD DEV =	0.258	(10.372)

\*\*\*\* AUTO CAL DATA \*\*\*\*

CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.123

Std Dev = 0.03 Rel Std Dev = 22.65

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.907

Std Dev = 0.01 Rel Std Dev = 0.64

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.996

Std Dev = 0.02 Rel Std Dev = 0.89

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.761

Std Dev = 0.01 Rel Std Dev = 0.25

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.444

Std Dev = 0.03 Rel Std Dev = 1.56

Zero Order Coef = -325.24

First Order Coef = 2462.37

Second Order Coef = 40.86

Standard Deviation = 22.245350

CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.123

Std Dev = 0.01 Rel Std Dev = 4.61

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.567

Std Dev = 0.01 Rel Std Dev = 0.66

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.612

Std Dev = 0.02 Rel Std Dev = 0.67

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.914

Std Dev = 0.02 Rel Std Dev = 0.26

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 9.934

Std Dev = 0.03 Rel Std Dev = 0.26

Zero Order Coef = -154.92

First Order Coef = 1294.49

Second Order Coef = 15.90

Standard Deviation = 24.075262

opt. Bench, Cal. Adj.

Post Stability Checks

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0004
0.040	0.041	-0.0008
0.100	0.100	0.0002
0.200	0.200	0.0002
0.300	0.300	-0.0001

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	-0.0002
0.100	0.099	0.0007
0.200	0.201	-0.0006
0.300	0.300	0.0002

Sol Value = 0.080 g/210L \*\*\*  
 Fit value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 \*\*\*\*\* CHANNEL 1  
 Sample #1 = 2918.00  
 Sample #2 = 3007.00  
 Sample #3 = 3016.00  
 Sample #4 = 2982.00  
 Average Result = 3001.6667  
 STD DEV = 17.6163  
 REL STD DEV = 0.587

\*\*\*\*\*  
 \*\*\*\*\* CHANNEL 2  
 Sample #1 = 3295.00  
 Sample #2 = 3320.00  
 Sample #3 = 3350.00  
 Sample #4 = 3342.00  
 Average Result = 3337.3333  
 STD DEV = 15.5349  
 REL STD DEV = 0.4651

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1025  
 3 um H2O Adjust (ng/l\*10,000) = 838  
 9 um H2O Adjust (ng/l\*10,000) = 472  
 \*\*\*\* AUTO CAL PASS

FHP  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000781  
 01/19/2021  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:00
Control Test	0.049	15:00
Air Blank	0.000	15:01
Control Test	0.049	15:01
Air Blank	0.000	15:02
Control Test	0.049	15:03
Air Blank	0.000	15:03
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FHP  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000781  
 01/19/2021  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:10
Control Test	0.202	15:10
Air Blank	0.000	15:11
Control Test	0.198	15:12
Air Blank	0.000	15:12
Control Test	0.199	15:13
Air Blank	0.000	15:13
Control Test Stats		
Average	0.1997	
Std Dev	0.0021	
Rel Std Dev(%)	1.0426	

Operator's Signature

FHP  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000781  
 01/19/2021  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:05
Control Test	0.080	15:05
Air Blank	0.000	15:06
Control Test	0.079	15:07
Air Blank	0.000	15:07
Control Test	0.079	15:08
Air Blank	0.000	15:09
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature

FHP  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000781  
 01/19/2021  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:14
Control Test	0.078	15:15
Air Blank	0.000	15:15
Control Test	0.079	15:16
Air Blank	0.000	15:16
Control Test	0.079	15:16
Air Blank	0.000	15:17
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

wet

Dry