

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000873, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000873</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI PD</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>09/30/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:55</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/30/2021

Date

David E Reyes-Rivera

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD
Time of Inspection: 09:55

Date of Inspection: 09/30/2021

Serial Number: 80-000873
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.197	0.078
0.000	0.048	0.078	0.197	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.048	0.079	0.197	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.078	0.197	0.078

Standard Deviations	0.0005	0.0005	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Stephen Rey Quinn

DAVID STEPHEN REY QUINN

Signature and Printed Name

09/30/2021
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities 2	80-000873	Miami Police Department	9/30/2021	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000873 09/30/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:53</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:56</td></tr> <tr><td>Control Test</td><td>0.048</td><td>07:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:57</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:53	Control Test	0.049	07:54	Air Blank	0.000	07:54	Control Test	0.049	07:55	Air Blank	0.000	07:56	Control Test	0.048	07:56	Air Blank	0.000	07:57	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000873 09/30/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:58</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:59</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:00</td></tr> <tr><td>Control Test</td><td>0.077</td><td>08:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.4741</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:58	Control Test	0.079	07:59	Air Blank	0.000	07:59	Control Test	0.079	08:00	Air Blank	0.000	08:00	Control Test	0.077	08:01	Air Blank	0.000	08:02	Control Test Stats			Average	0.0783		Std Dev	0.0012		Rel Std Dev(%)	1.4741		<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000873 09/30/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:03</td></tr> <tr><td>Control Test</td><td>0.198</td><td>08:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:04</td></tr> <tr><td>Control Test</td><td>0.198</td><td>08:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:05</td></tr> <tr><td>Control Test</td><td>0.197</td><td>08:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	08:03	Control Test	0.198	08:04	Air Blank	0.000	08:04	Control Test	0.198	08:05	Air Blank	0.000	08:05	Control Test	0.197	08:06	Air Blank	0.000	08:07	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000873 09/30/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:08</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:08</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:09</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	08:08	Control Test	0.080	08:08	Air Blank	0.000	08:08	Control Test	0.079	08:09	Air Blank	0.000	08:09	Control Test	0.079	08:10	Air Blank	0.000	08:10	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:53																																																																																																																																																	
Control Test	0.049	07:54																																																																																																																																																	
Air Blank	0.000	07:54																																																																																																																																																	
Control Test	0.049	07:55																																																																																																																																																	
Air Blank	0.000	07:56																																																																																																																																																	
Control Test	0.048	07:56																																																																																																																																																	
Air Blank	0.000	07:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:58																																																																																																																																																	
Control Test	0.079	07:59																																																																																																																																																	
Air Blank	0.000	07:59																																																																																																																																																	
Control Test	0.079	08:00																																																																																																																																																	
Air Blank	0.000	08:00																																																																																																																																																	
Control Test	0.077	08:01																																																																																																																																																	
Air Blank	0.000	08:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	1.4741																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:03																																																																																																																																																	
Control Test	0.198	08:04																																																																																																																																																	
Air Blank	0.000	08:04																																																																																																																																																	
Control Test	0.198	08:05																																																																																																																																																	
Air Blank	0.000	08:05																																																																																																																																																	
Control Test	0.197	08:06																																																																																																																																																	
Air Blank	0.000	08:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:08																																																																																																																																																	
Control Test	0.080	08:08																																																																																																																																																	
Air Blank	0.000	08:08																																																																																																																																																	
Control Test	0.079	08:09																																																																																																																																																	
Air Blank	0.000	08:09																																																																																																																																																	
Control Test	0.079	08:10																																																																																																																																																	
Air Blank	0.000	08:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		

MIAMI PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000873
09/30/2021
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SORT(0iff)) = 6.164
2: Rate (Liters/min) = 15
SORT(0iff)) = 11.043
3: Rate (Liters/min) = 30
SORT(0iff)) = 20.566
Dependent Data Scale Factor = 10000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 669
Rounded Intercept = -491127
Correlation = 0.99761

Flow Calibration	
SN:	80-000873
Agency:	Miami Police Department
Date:	9/30/2021
By:	DERR <i>[Signature]</i>



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000873, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000873</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI PD</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>09/21/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:30</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/21/2021

Date

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 January 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD
Time of Inspection: 16:30

Date of Inspection: 09/21/2021

Serial Number: 80-000873
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.079	0.197	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.048	0.079	0.197	0.079


Standard Deviations	0.0004	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

09/21/2021
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000873	Miami Police Department	9/21/2021	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000873 09/21/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:02</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:05</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	14:02	Control Test	0.048	14:03	Air Blank	0.000	14:04	Control Test	0.049	14:04	Air Blank	0.000	14:05	Control Test	0.049	14:05	Air Blank	0.000	14:06	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000873 09/21/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:07</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:09</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:10</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	14:07	Control Test	0.079	14:08	Air Blank	0.000	14:09	Control Test	0.078	14:09	Air Blank	0.000	14:10	Control Test	0.078	14:11	Air Blank	0.000	14:11	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000873 09/21/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test</td><td>0.196</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1967</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2936</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	14:12	Control Test	0.196	14:13	Air Blank	0.000	14:13	Control Test	0.197	14:14	Air Blank	0.000	14:15	Control Test	0.197	14:15	Air Blank	0.000	14:16	Control Test Stats			Average	0.1967		Std Dev	0.0006		Rel Std Dev(%)	0.2936		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000873 09/21/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:19</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	14:18	Control Test	0.080	14:18	Air Blank	0.000	14:18	Control Test	0.079	14:19	Air Blank	0.000	14:19	Control Test	0.079	14:20	Air Blank	0.000	14:20	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:02																																																																																																																																																	
Control Test	0.048	14:03																																																																																																																																																	
Air Blank	0.000	14:04																																																																																																																																																	
Control Test	0.049	14:04																																																																																																																																																	
Air Blank	0.000	14:05																																																																																																																																																	
Control Test	0.049	14:05																																																																																																																																																	
Air Blank	0.000	14:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:07																																																																																																																																																	
Control Test	0.079	14:08																																																																																																																																																	
Air Blank	0.000	14:09																																																																																																																																																	
Control Test	0.078	14:09																																																																																																																																																	
Air Blank	0.000	14:10																																																																																																																																																	
Control Test	0.078	14:11																																																																																																																																																	
Air Blank	0.000	14:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:12																																																																																																																																																	
Control Test	0.196	14:13																																																																																																																																																	
Air Blank	0.000	14:13																																																																																																																																																	
Control Test	0.197	14:14																																																																																																																																																	
Air Blank	0.000	14:15																																																																																																																																																	
Control Test	0.197	14:15																																																																																																																																																	
Air Blank	0.000	14:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1967																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2936																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:18																																																																																																																																																	
Control Test	0.080	14:18																																																																																																																																																	
Air Blank	0.000	14:18																																																																																																																																																	
Control Test	0.079	14:19																																																																																																																																																	
Air Blank	0.000	14:19																																																																																																																																																	
Control Test	0.079	14:20																																																																																																																																																	
Air Blank	0.000	14:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		