



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade Police Department

S/N 80-000881

Florida Department of Law Enforcement

Date In 3/1/2021

DI Completion Date

Ship P/U H/D CMI EE

Intake By <u>DERR</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____	Quality Checks By <u>DER</u> Date <u>3/1/2021</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>215</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.105</u> (.139 - .169) 36 mm <u>0.136</u> (.156 - .190) 53 mm <u>0.207</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 01/30/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	202010A 10/05/2022	0.080	SD3968	202010B 10/05/2022	0.200	SD3969	202010D 10/06/2022	0.080 DGS	N/A	AG003005 01/30/2022	Flow Calibration By <u>DER</u> Date <u>3/1/2021</u> Flow Column # <u>ATP101</u> <input checked="" type="checkbox"/> 5L/min – 17mm <input checked="" type="checkbox"/> 15L/min – 53mm <input checked="" type="checkbox"/> 30L/min – 103mm <input checked="" type="checkbox"/> R-Value <u>215/208/207</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.128/0.132/105</u> (.139 - .169) 36 mm <u>0.148/0.148/125</u> (.156 - .190) 53 mm <u>0.218/0.214/195</u> (.228 - .278) 103 mm <u>0.492/0.500/480</u> (.447 - .547) Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ DI Temp. Checks By <u>DERR</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.85C</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>
Simulator	Serial #	Lot #/Exp															
0.050	SD3967	202010A 10/05/2022															
0.080	SD3968	202010B 10/05/2022															
0.200	SD3969	202010D 10/06/2022															
0.080 DGS	N/A	AG003005 01/30/2022															

Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By _____ Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.050</td> <td></td> </tr> <tr> <td>0.080</td> <td></td> </tr> <tr> <td>0.200</td> <td></td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Form 51</u>	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000																																																													
Interferent																																																													
0.050																																																													
0.080																																																													
0.200																																																													

Notes/Suggested Service: <u>Attempted 3 flow calibrations each time the instrument failed the verification. Returning the instrument back to CMI. DERR</u> _____ _____ _____ _____ _____	<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____
---	--

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000881	Miami-Dade Police Department	3/1/2021	DERR <i>DELL</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000881 03/01/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:05</td></tr> <tr><td>Control Test</td><td>0.048</td><td>07:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:06</td></tr> <tr><td>Control Test</td><td>0.047</td><td>07:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:07</td></tr> <tr><td>Control Test</td><td>0.048</td><td>07:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:08</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0477</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2112</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>DELL</i></p>	Test	g/210L	Time	Air Blank	0.000	07:05	Control Test	0.048	07:05	Air Blank	0.000	07:06	Control Test	0.047	07:07	Air Blank	0.000	07:07	Control Test	0.048	07:08	Air Blank	0.000	07:08	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel Std Dev(%)	1.2112		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000881 03/01/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:09</td></tr> <tr><td>Control Test</td><td>0.078</td><td>07:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:11</td></tr> <tr><td>Control Test</td><td>0.078</td><td>07:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:12</td></tr> <tr><td>Control Test</td><td>0.077</td><td>07:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:13</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>DELL</i></p>	Test	g/210L	Time	Air Blank	0.000	07:09	Control Test	0.078	07:10	Air Blank	0.000	07:11	Control Test	0.078	07:11	Air Blank	0.000	07:12	Control Test	0.077	07:12	Air Blank	0.000	07:13	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000881 03/01/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:14</td></tr> <tr><td>Control Test</td><td>0.196</td><td>07:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:16</td></tr> <tr><td>Control Test</td><td>0.194</td><td>07:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:17</td></tr> <tr><td>Control Test</td><td>0.196</td><td>07:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:18</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.1953</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5911</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>DELL</i></p>	Test	g/210L	Time	Air Blank	0.000	07:14	Control Test	0.196	07:15	Air Blank	0.000	07:16	Control Test	0.194	07:16	Air Blank	0.000	07:17	Control Test	0.196	07:17	Air Blank	0.000	07:18	Control Test Stats			Average	0.1953		Std Dev	0.0012		Rel Std Dev(%)	0.5911		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000881 03/01/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:20</td></tr> <tr><td>Control Test</td><td>0.081</td><td>07:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:21</td></tr> <tr><td>Control Test</td><td>0.081</td><td>07:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:21</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:22</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.4374</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>DELL</i></p>	Test	g/210L	Time	Air Blank	0.000	07:20	Control Test	0.081	07:20	Air Blank	0.000	07:21	Control Test	0.081	07:21	Air Blank	0.000	07:21	Control Test	0.079	07:22	Air Blank	0.000	07:22	Control Test Stats			Average	0.0803		Std Dev	0.0012		Rel Std Dev(%)	1.4374	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:05																																																																																																																																																	
Control Test	0.048	07:05																																																																																																																																																	
Air Blank	0.000	07:06																																																																																																																																																	
Control Test	0.047	07:07																																																																																																																																																	
Air Blank	0.000	07:07																																																																																																																																																	
Control Test	0.048	07:08																																																																																																																																																	
Air Blank	0.000	07:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0477																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2112																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:09																																																																																																																																																	
Control Test	0.078	07:10																																																																																																																																																	
Air Blank	0.000	07:11																																																																																																																																																	
Control Test	0.078	07:11																																																																																																																																																	
Air Blank	0.000	07:12																																																																																																																																																	
Control Test	0.077	07:12																																																																																																																																																	
Air Blank	0.000	07:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:14																																																																																																																																																	
Control Test	0.196	07:15																																																																																																																																																	
Air Blank	0.000	07:16																																																																																																																																																	
Control Test	0.194	07:16																																																																																																																																																	
Air Blank	0.000	07:17																																																																																																																																																	
Control Test	0.196	07:17																																																																																																																																																	
Air Blank	0.000	07:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1953																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:20																																																																																																																																																	
Control Test	0.081	07:20																																																																																																																																																	
Air Blank	0.000	07:21																																																																																																																																																	
Control Test	0.081	07:21																																																																																																																																																	
Air Blank	0.000	07:21																																																																																																																																																	
Control Test	0.079	07:22																																																																																																																																																	
Air Blank	0.000	07:22																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	1.4374																																																																																																																																																		

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000881
03/01/2021
Software: 8100.27

Flow Rate Calibration*****

- 1: Rate (Liters/min) = 5
SQRT(Diff)) = 7.680
- 2: Rate (Liters/min) = 15
SQRT(Diff)) = 12.082
- 3: Rate (Liters/min) = 30
SQRT(Diff)) = 21.723

Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 681
Rounded Intercept = -744473
Correlation = 0.99528

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000881
03/01/2021
Software: 8100.27

Flow Rate Calibration*****

- 1: Rate (Liters/min) = 5
SQRT(Diff)) = 7.414
- 2: Rate (Liters/min) = 15
SQRT(Diff)) = 12.164
- 3: Rate (Liters/min) = 30
SQRT(Diff)) = 21.539

Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 682
Rounded Intercept = -726550
Correlation = 0.99742

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000881
03/01/2021
Software: 8100.27

Flow Rate Calibration*****

- 1: Rate (Liters/min) = 5
SQRT(Diff)) = 9.000
- 2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.789
- 3: Rate (Liters/min) = 30
SQRT(Diff)) = 21.234

Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 753
Rounded Intercept = -1033157
Correlation = 0.98212

Flow Calibration	
SN:	80-000881
Agency:	Miami Dade Police Department
Date:	3/1/2021
By:	DERR 

