INSTRUMENT PROCESSING SHEET

FDLE
Florida Department of Law Enforcement

Agency Hollywood PD S/N 80-001063

Date In 9/13/2021 DI Completion Date 9/13/2021 ☐Ship ■P/U ☐H/D ☐CMI ☐EE

Intake By MH

■ Annual
☐ Registration
☐ Return from CMI / EE

Visual Inspection:
■ Case ■ Handle
■ Keyboard ■ Dry Gas Shelf
■ Feet ■ Breath Tube
■ Ports ■ Screws Tight

Other Equipment/ Accessories:
■ Power cord ☐ Printer Cable
☐ Static Bag ☐ 12V DC Cable

Notes:

Quality Checks By MH Date 09/13/2021

■ Breath Tube Screen
■ Replace External O-Rings
■ Instrument Set Up Verified
■ R-Value 220
■ Flow Verification (L/s)
Flow Column # ATP 106
32 mm 0.160 (.139 - .169)
36 mm 0.175 (.156 - .190)
53 mm 0.242 (.228 - .278)
103 mm 0.503 (.447 - .547)
■ Barometric Pressure Check
Gauge ID # 28199
■ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	MP4863	202010A 10/05/2022
0.080	MP4864	202010B 10/05/2022
0.200	MP5097	202010D 10/06/2022
0.080 DGS	N/A	AG026705 09/23/2022

Flow Calibration By Date

Flow Column #
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value
☐ Post Calibration Verification (L/s)
Flow Column #
32 mm (.139 - .169)
36 mm (.156 - .190)
53 mm (.228 - .278)
103 mm (.447 - .547)

Maintenance By

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☐ Other

DI Temp. Checks By MH

■ Lab Temp °C 21.37
External Digital Therm. ID#: 381196
■ 34°C +-.2 Serial #: MP4863
■ 34°C +-.2 Serial #: MP4864
■ 34°C +-.2 Serial #: MP5097

Calibration Adjustment By

Barometric Pressure Gauge ID #

Simulator	Serial #	Lot #	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection By MH

Barometric Pressure ID# 68639
Gauge 1019 Instrument 1018
Mouth Alcohol Solution Lot # 2021-B
Acetone Stock Solution Lot # 2020-A

Simulator	Serial Number
0.000	SD1014
Interferent	SD1015
0.050	MP4863
0.080	MP4864
0.200	MP5097

Attachments

■ Form 41
■ Stability Checks
■ Calibration Certificate
☐ Calibration Adjustment
☐ Post-Stability Checks
☐ Flow Calibration
☐ Form 40
☐ Other

Notes/Suggested Service:

■ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
■ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**
■ **Conduct an Agency Inspection Before Evidentiary Use**

David Eliezer Reyes-Rivera Digitally signed by David Eliezer Reyes-Rivera Date: 2021.09.13 14:12:08 -04'00'
Tech Review / Date

2021.09.13 14:16:16-04'00'
Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLLYWOOD PD
Time of Inspection: 11:55 Date of Inspection: 09/13/2021

Serial Number: 80-001063
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.197	0.080

Standard Deviations	0.0000	0.0004	0.0006	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL D HAUGHEY
Signature and Printed Name

09/13/2021
Date

FDLE/ATP Form 41 –Revised August 2005

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001063	Hollywood PD	09/13/21	MJK

0.05g/210L

0.047 to 0.053 ☑

HOLLYWOOD PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001063
09/13/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:06
Control Test	0.048	10:07
Air Blank	0.000	10:07
Control Test	0.049	10:08
Air Blank	0.000	10:08
Control Test	0.049	10:09
Air Blank	0.000	10:10

Control Test Stats

Average	0.0487
Std Dev	0.0006
Rel Std Dev(%)	1.1863

MJK
Operator's Signature

0.08g/210L

0.077 to 0.083 ☑

HOLLYWOOD PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001063
09/13/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.079	10:12
Air Blank	0.000	10:12
Control Test	0.078	10:13
Air Blank	0.000	10:13
Control Test	0.079	10:14
Air Blank	0.000	10:15

Control Test Stats

Average	0.0787
Std Dev	0.0006
Rel Std Dev(%)	0.7339

MJK
Operator's Signature

0.20g/210L

0.194 to 0.206 ☑

HOLLYWOOD PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001063
09/13/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:16
Control Test	0.196	10:17
Air Blank	0.000	10:17
Control Test	0.196	10:18
Air Blank	0.000	10:18
Control Test	0.196	10:19
Air Blank	0.000	10:20

Control Test Stats

Average	0.1960
Std Dev	0.0000
Rel Std Dev(%)	0.0000

MJK
Operator's Signature

DGS 0.08g/210L

0.077 to 0.083 ☑

HOLLYWOOD PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001063
09/13/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:22
Control Test	0.079	10:23
Air Blank	0.000	10:23
Control Test	0.079	10:23
Air Blank	0.000	10:24
Control Test	0.079	10:24
Air Blank	0.000	10:25

Control Test Stats

Average	0.0790
Std Dev	0.0000
Rel Std Dev(%)	0.0000

MJK
Operator's Signature

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001063 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001063		UNCERTAINTY* ±
Owning Agency:	HOLLYWOOD PD	0.050 g/ 210 L	0.005
Calibration Date:	09/13/2021	0.080 g/ 210 L	0.004
Calibration Time:	11:55	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full,
without written approval of the Florida Department of
Law Enforcement Alcohol Testing Program.

09/13/2021
Date

MICHAEL D HAUGHEY,
Department Inspector