



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Troop E



S/N 80-001121

Florida Department of
Law Enforcement

Date In 11/4/2021 DI Completion Date 11/4/2021

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By DERR	Quality Checks	By DER	Date 11/4/2021	Flow Calibration	By	Date																																																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>187</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																														
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705 09/23/2022</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202010A 10/05/2022	0.080	MP6287	202010B 10/05/2022	0.200	MP6288	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022	Maintenance By DERR ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Updated dial out number</u> DI Temp. Checks By DERR ☒ Lab Temp °C <u>21.50C</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>MP6286</u> ☒ 34°C +/- .2 Serial #: <u>MP6287</u> ☒ 34°C +/- .2 Serial #: <u>MP6288</u>																																																
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP6286	202010A 10/05/2022																																																																	
0.080	MP6287	202010B 10/05/2022																																																																	
0.200	MP6288	202010D 10/06/2022																																																																	
0.080 DGS	N/A	AG026705 09/23/2022																																																																	
Calibration Adjustment By _____ Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By DERR Barometric Pressure ID# <u>28199</u> Gauge <u>1015</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP6284</td></tr> <tr><td>Interferent</td><td>MP6285</td></tr> <tr><td>0.050</td><td>MP6286</td></tr> <tr><td>0.080</td><td>MP6287</td></tr> <tr><td>0.200</td><td>MP6288</td></tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____						Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	MP6284																																																																		
Interferent	MP6285																																																																		
0.050	MP6286																																																																		
0.080	MP6287																																																																		
0.200	MP6288																																																																		
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Israel Soto Digitally signed by Israel Soto Date: 2021.11.05 07:30:15 +0400 05 07:37:17 Tech Review / Date Admin Review / Date																																																																	



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001121, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001121</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP TROOP E MIAMI</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>11/04/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:58</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/04/2021

Date

David E Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 14:58

Date of Inspection: 11/04/2021

Serial Number: 80-001121
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.081
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.080	0.198	0.081


Standard Deviations	0.0003	0.0005	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

11/04/2021
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001121	Florida Highway Patrol Troop E	11/4/2021	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																			
FHP TROOP E MIAW Intoxilyzer - Alcohol Analyzer Model 8000 11/04/2021 Software: 8100.27 SN 80-001121 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr><tr><td>Control Test</td><td>0.049</td><td>11:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr><tr><td>Control Test</td><td>0.049</td><td>11:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr><tr><td>Control Test</td><td>0.049</td><td>11:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr><tr><td>Control Test</td><td>0.049</td><td>11:20</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:17	Control Test	0.049	11:17	Air Blank	0.000	11:18	Control Test	0.049	11:19	Air Blank	0.000	11:19	Control Test	0.049	11:20	Air Blank	0.000	11:20	Control Test	0.049	11:20	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FHP TROOP E MIAW Intoxilyzer - Alcohol Analyzer Model 8000 11/04/2021 Software: 8100.27 SN 80-001121 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0793</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr></tbody></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:22	Control Test	0.079	11:23	Air Blank	0.000	11:23	Control Test	0.080	11:24	Air Blank	0.000	11:24	Control Test	0.079	11:25	Air Blank	0.000	11:26	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		FHP TROOP E MIAW Intoxilyzer - Alcohol Analyzer Model 8000 11/04/2021 Software: 8100.27 SN 80-001121 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:27</td></tr><tr><td>Control Test</td><td>0.199</td><td>11:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr><tr><td>Control Test</td><td>0.199</td><td>11:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1987</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2966</td><td></td></tr></tbody></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:27	Control Test	0.199	11:28	Air Blank	0.000	11:29	Control Test	0.199	11:29	Air Blank	0.000	11:30	Control Test	0.198	11:31	Air Blank	0.000	11:31	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2966		FHP TROOP E MIAW Intoxilyzer - Alcohol Analyzer Model 8000 11/04/2021 Software: 8100.27 SN 80-001121 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr><tr><td>Control Test</td><td>0.081</td><td>11:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></tbody></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:33	Control Test	0.080	11:33	Air Blank	0.000	11:34	Control Test	0.080	11:34	Air Blank	0.000	11:35	Control Test	0.081	11:35	Air Blank	0.000	11:35	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	11:17																																																																																																																																																				
Control Test	0.049	11:17																																																																																																																																																				
Air Blank	0.000	11:18																																																																																																																																																				
Control Test	0.049	11:19																																																																																																																																																				
Air Blank	0.000	11:19																																																																																																																																																				
Control Test	0.049	11:20																																																																																																																																																				
Air Blank	0.000	11:20																																																																																																																																																				
Control Test	0.049	11:20																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.0490																																																																																																																																																					
Std Dev	0.0000																																																																																																																																																					
Rel Std Dev(%)	0.0000																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	11:22																																																																																																																																																				
Control Test	0.079	11:23																																																																																																																																																				
Air Blank	0.000	11:23																																																																																																																																																				
Control Test	0.080	11:24																																																																																																																																																				
Air Blank	0.000	11:24																																																																																																																																																				
Control Test	0.079	11:25																																																																																																																																																				
Air Blank	0.000	11:26																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.0793																																																																																																																																																					
Std Dev	0.0006																																																																																																																																																					
Rel Std Dev(%)	0.7277																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	11:27																																																																																																																																																				
Control Test	0.199	11:28																																																																																																																																																				
Air Blank	0.000	11:29																																																																																																																																																				
Control Test	0.199	11:29																																																																																																																																																				
Air Blank	0.000	11:30																																																																																																																																																				
Control Test	0.198	11:31																																																																																																																																																				
Air Blank	0.000	11:31																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.1987																																																																																																																																																					
Std Dev	0.0006																																																																																																																																																					
Rel Std Dev(%)	0.2966																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	11:33																																																																																																																																																				
Control Test	0.080	11:33																																																																																																																																																				
Air Blank	0.000	11:34																																																																																																																																																				
Control Test	0.080	11:34																																																																																																																																																				
Air Blank	0.000	11:35																																																																																																																																																				
Control Test	0.081	11:35																																																																																																																																																				
Air Blank	0.000	11:35																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.0803																																																																																																																																																					
Std Dev	0.0006																																																																																																																																																					
Rel Std Dev(%)	0.7187																																																																																																																																																					