



INSTRUMENT PROCESSING SHEET

Agency Volusia CSO

S/N 80-001132

Florida Department of
Law Enforcement

Date In 8/30/2021

DI Completion Date 9/14/2021

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 9/14/2021	Flow Calibration	By	Date																																																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>FDLE loaner box B-02</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>146</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																														
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5092</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705 09/23/2022</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A 10/05/2022	0.080	MP5093	202010B 10/05/2022	0.200	MP5094	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																																
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP5092	202010A 10/05/2022																																																																	
0.080	MP5093	202010B 10/05/2022																																																																	
0.200	MP5094	202010D 10/06/2022																																																																	
0.080 DGS	N/A	AG026705 09/23/2022																																																																	
				DI Temp. Checks By TDG _____ ☒ Lab Temp °C <u>21.18</u> External Digital Therm. ID#: <u>381198</u> ☒ 34°C +/- .2 Serial #: <u>MP5092</u> ☒ 34°C +/- .2 Serial #: <u>MP5093</u> ☒ 34°C +/- .2 Serial #: <u>MP5094</u>																																																															
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By TDG _____ Barometric Pressure ID# <u>28199</u> Gauge <u>1016</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>SD3963</td></tr> <tr><td>Interferent</td><td>SD1017</td></tr> <tr><td>0.050</td><td>MP5092</td></tr> <tr><td>0.080</td><td>MP5093</td></tr> <tr><td>0.200</td><td>MP5094</td></tr> </tbody> </table>				Simulator	Serial Number	0.000	SD3963	Interferent	SD1017	0.050	MP5092	0.080	MP5093	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	SD3963																																																																		
Interferent	SD1017																																																																		
0.050	MP5092																																																																		
0.080	MP5093																																																																		
0.200	MP5094																																																																		
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																															
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																															
David Eliezer Reyes-Rivera Digitally signed by David Eliezer Reyes-Rivera Date: 2021.09.14 13:45:47 -04'00'				 14 15:52:43 09/14/2021																																																															
Tech Review / Date				Admin. Review / Date																																																															

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001132	Volusia CSO	09/14/2021	TDG MK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓ Volusia County S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001132 09/14/2021 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:32</td></tr> <tr><td>Control Test</td><td>0.050</td><td>09:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:34</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:35</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:36</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </table> Operator's Signature _____	Air Blank	0.000	09:32	Control Test	0.050	09:33	Air Blank	0.000	09:34	Control Test	0.049	09:34	Air Blank	0.000	09:35	Control Test	0.049	09:36	Air Blank	0.000	09:36	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		Volusia County S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001132 09/14/2021 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:40</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:41</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:42</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature _____	Air Blank	0.000	09:40	Control Test	0.079	09:40	Air Blank	0.000	09:41	Control Test	0.079	09:41	Air Blank	0.000	09:42	Control Test	0.079	09:43	Air Blank	0.000	09:43	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		Volusia County S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001132 09/14/2021 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:52</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:53</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:55</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </table> Operator's Signature _____	Air Blank	0.000	09:51	Control Test	0.198	09:52	Air Blank	0.000	09:52	Control Test	0.198	09:53	Air Blank	0.000	09:53	Control Test	0.199	09:54	Air Blank	0.000	09:55	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		Volusia County S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001132 09/14/2021 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:56</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:58</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:59</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature _____	Air Blank	0.000	09:56	Control Test	0.079	09:57	Air Blank	0.000	09:57	Control Test	0.079	09:57	Air Blank	0.000	09:58	Control Test	0.079	09:58	Air Blank	0.000	09:59	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	09:32																																																																																																																																					
Control Test	0.050	09:33																																																																																																																																					
Air Blank	0.000	09:34																																																																																																																																					
Control Test	0.049	09:34																																																																																																																																					
Air Blank	0.000	09:35																																																																																																																																					
Control Test	0.049	09:36																																																																																																																																					
Air Blank	0.000	09:36																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0493																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1703																																																																																																																																						
Air Blank	0.000	09:40																																																																																																																																					
Control Test	0.079	09:40																																																																																																																																					
Air Blank	0.000	09:41																																																																																																																																					
Control Test	0.079	09:41																																																																																																																																					
Air Blank	0.000	09:42																																																																																																																																					
Control Test	0.079	09:43																																																																																																																																					
Air Blank	0.000	09:43																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	09:51																																																																																																																																					
Control Test	0.198	09:52																																																																																																																																					
Air Blank	0.000	09:52																																																																																																																																					
Control Test	0.198	09:53																																																																																																																																					
Air Blank	0.000	09:53																																																																																																																																					
Control Test	0.199	09:54																																																																																																																																					
Air Blank	0.000	09:55																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1983																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2911																																																																																																																																						
Air Blank	0.000	09:56																																																																																																																																					
Control Test	0.079	09:57																																																																																																																																					
Air Blank	0.000	09:57																																																																																																																																					
Control Test	0.079	09:57																																																																																																																																					
Air Blank	0.000	09:58																																																																																																																																					
Control Test	0.079	09:58																																																																																																																																					
Air Blank	0.000	09:59																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						

Comments:

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT – INTOXILYZER 8000

Agency: VOLUSIA COUNTY S.O.
Time of Inspection: 12:36

Date of Inspection: 09/14/2021

Serial Number: 80-001132
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.079

Standard Deviations	0.0000	0.0000	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5



Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D. Gutschow TAYLOR D. GUTSCHOW
Signature and Printed Name

09/14/2021
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001132, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001132</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>VOLUSIA COUNTY S.O.</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>09/14/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:36</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/14/2021

Date



TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality