FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Daytona Beach PD S/N 80-001135

Date In 4/13/2021 DI Completion Date 4/15/2021 ■Ship ❑P/U ❑H/D ❑CMI ❑EE

Intake By TDG

■ Annual
❑ Registration
❑ Return from CMI / EE

Visual Inspection:
■ Case ■ Handle
■ Keyboard ■ Dry Gas Shelf
■ Feet ■ Breath Tube
■ Ports ■ Screws Tight

Other Equipment/ Accessories:
❑ Power cord ❑ Printer Cable
■ Static Bag ❑ 12V DC Cable

Notes:

Quality Checks By TDG Date 4/15/2021

■ Breath Tube Screen
■ Replace External O-Rings
■ Instrument Set Up Verified
■ R-Value 222
■ Flow Verification (L/s)
Flow Column # ATP101
32 mm 0.160 (.139 - .169)
36 mm 0.179 (.156 - .190)
53 mm 0.238 (.228 - .278)
103 mm 0.492 (.447 - .547)
■ Barometric Pressure Check
Gauge ID # 68639
■ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	MP5092	202010A
		10/05/2022
0.080	MP5093	202010B
		10/05/2022
0.200	MP5094	202010D
		10/06/2022
0.080 DGS	N/A	AG026705
		09/23/2022

Flow Calibration By Date

Flow Column #
❑ 5L/min – 17mm
❑ 15L/min – 53mm
❑ 30L/min – 103mm
❑ R-Value
❑ Post Calibration Verification (L/s)
Flow Column #
32 mm (.139 - .169)
36 mm (.156 - .190)
53 mm (.228 - .278)
103 mm (.447 - .547)

Maintenance By

❑ Battery Replacement
❑ Dry Gas Regulator Replacement
❑ Breath Tube Replacement
❑ Other

DI Temp. Checks By TDG

■ Lab Temp °C 21.81
External Digital Therm. ID#: 300504
■ 34°C +-.2 Serial #: MP5092
■ 34°C +-.2 Serial #: MP5093
■ 34°C +-.2 Serial #: MP5094

Calibration Adjustment By

Barometric Pressure Gauge ID #

Simulator	Serial #	Lot #	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

❑ Post Calibration Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection By TDG

Barometric Pressure ID# 28199
Gauge 1013 Instrument 1012
Mouth Alcohol Solution Lot # 2019-B
Acetone Stock Solution Lot # 2020-A

Simulator	Serial Number
0.000	SD3963
Interferent	SD1017
0.050	MP5092
0.080	MP5093
0.200	MP5094

Attachments

■ Form 41	❑ Post-Stability Checks
■ Stability Checks	❑ Flow Calibration
■ Calibration Certificate	❑ Form 40
❑ Calibration Adjustment	❑ Other

Notes/Suggested Service:

■ **Instrument Complies with Chapter 11D-8, FAC**
❑ **Instrument Does Not Comply with Chapter 11D-8, FAC**
■ **Return to/Place into Evidentiary Use**
❑ **Remain Out of Evidentiary Use**
■ **Conduct an Agency Inspection Before Evidentiary Use**

Michael D Haughey 2021.04.16 13:51:33 -04'00'
Tech Review / Date

2021.04.21 09:18:41 -04'00'
Admin Review / Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001135	Daytona Beach PD	04/15/2021	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L	
0.047 to 0.053 ✓	0.077 to 0.083 ✓	0.194 to 0.206 ✓	0.077 to 0.083 ✓	≤0.003 of Wet ✓

Test	g/210L	Time
Air Blank	0.000	11:02
Control Test	0.050	11:03
Air Blank	0.000	11:04
Control Test	0.050	11:04
Air Blank	0.000	11:05
Control Test	0.050	11:06
Air Blank	0.000	11:06
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MG

Operator's Signature

DAYTONA BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001135
04/15/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:10
Control Test	0.080	11:11
Air Blank	0.000	11:12
Control Test	0.080	11:12
Air Blank	0.000	11:13
Control Test	0.080	11:14
Air Blank	0.000	11:14
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MG

Operator's Signature

DAYTONA BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001135
04/15/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:18
Control Test	0.201	11:18
Air Blank	0.000	11:19
Control Test	0.200	11:20
Air Blank	0.000	11:20
Control Test	0.201	11:21
Air Blank	0.000	11:22
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

MG

Operator's Signature

DGS

DAYTONA BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001135
04/15/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:24
Control Test	0.081	11:24
Air Blank	0.000	11:25
Control Test	0.081	11:25
Air Blank	0.000	11:26
Control Test	0.082	11:26
Air Blank	0.000	11:27
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

MG

Operator's Signature

Comments:

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DAYTONA BEACH P.D. Serial Number: 80-001135
Time of Inspection: 13:51 Date of Inspection: 04/15/2021 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.079	0.200	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.079	0.201	0.082
0.000	0.050	0.079	0.200	0.082
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.079	0.200	0.081
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.201	0.081
Standard Deviations	0.0003	0.0005	0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor Gutschow TAYLOR D GUTSCHOW
Signature and Printed Name

04/15/2021
Date

FDLE/ATP Form 41 –Revised August 2005

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001135 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001135	UNCERTAINTY* ±	
Owning Agency:	DAYTONA BEACH P.D.	0.050 g/ 210 L	0.005
Calibration Date:	04/15/2021	0.080 g/ 210 L	0.004
Calibration Time:	13:51	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full,
without written approval of the Florida Department of
Law Enforcement Alcohol Testing Program.

04/15/2021
Date

TAYLOR D GUTSCHOW,
Department Inspector

Return Material Authorization

Ship to: ☑ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Justin Dunne on 11/29/2021

Items Returned: Instrument ☑ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001135

Bill To Address:
Daytona Beach Police Department
Attn: Justin Dunne

Ship to Address:
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
Attn: Alcohol Testing Program
4700 Terminal Drive, Suite 1
Fort Myers, FL 33907

Reason for Return:

Displays "Err: 012, Contact CMI". Uploaded records prior to sending to repair. No records remain on the instrument.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to $_____.

☑ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Justin Dunne

Phone #: 386-316-5594 Email: DunneJustin@DBPD.US

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Daytona Beach PD S/N 80-001135

Date In 11/29/2021 DI Completion Date _____ ■Ship ☐P/U ☐H/D ■CMI ☐EE

Intake By TDG

- ■ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ■ Case ■ Handle
- ■ Keyboard ■ Dry Gas Shelf
- ■ Feet ■ Breath Tube
- ■ Ports ■ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ■ Static Bag ☐ 12V DC Cable

Notes: AI Justin Dunne reports an Error 12 and a second error that occurs when he enters the agency code ("Error- Check File Agency"). Agency was unable to upload. I uploaded via direct connect on 11/29/2021.

Quality Checks By ____ Date ____

- ☐ Breath Tube Screen
- ☐ Replace External O-Rings
- ☐ Instrument Set Up Verified
- ☐ R-Value _____
- ☐ Flow Verification (L/s)

Flow Column # _____

32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

- ☐ Barometric Pressure Check

Gauge ID # _____

- ☐ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050		
0.080		
0.200		
0.080 DGS	N/A	

Flow Calibration By ____ Date ____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

Maintenance By ____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

DI Temp. Checks By ____

- ☐ Lab Temp °C _____

External Digital Therm. ID#: _____

- ☐ 34°C +-.2 Serial #: _____
- ☐ 34°C +-.2 Serial #: _____
- ☐ 34°C +-.2 Serial #: _____

Calibration ~~Adjustment~~ By ____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial #	Lot #	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

- ☐ Post Calibration Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection By ____

Barometric Pressure ID# _____

Gauge _____ Instrument _____

Mouth Alcohol Solution Lot # _____

Acetone Stock Solution Lot # _____

Simulator	Serial Number
0.000	
Interferent	
0.050	
0.080	
0.200	

Attachments

- ☐ Form 41
- ☐ Stability Checks
- ☐ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ■ Other Form 51

Notes/Suggested Service: Instrument still displays Error 12 after direct-connect upload. Will send to CMI. (TDG)

- ☐ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☐ **Return to/Place into Evidentiary Use**
- ■ **Remain Out of Evidentiary Use**
- ☐ **Conduct an Agency Inspection Before Evidentiary Use**

Israel Soto Digitally signed by Israel Soto Date: 2022.03.21 14:06:20 -04'00'

Tech Review / Date _____ Admin Review / Date