



INSTRUMENT PROCESSING SHEET

Agency Hiialeah Gardens Police Department

S/N 80-001275

Florida Department of Law Enforcement

Date In 3/18/2021

DI Completion Date 3/18/2021

Ship

P/U

H/D

CMI

EE

Intake By <u>DERR</u>		Quality Checks By <u>DER</u> Date <u>3/18/2021</u>		Flow Calibration By <u>DER</u> Date <u>3/18/2021</u>																
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>115</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.132</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.226</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks		Flow Column # <u>ATP101</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>115</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547)																
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705 09/23/2022</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD3967	202010A 10/05/2022	0.080	SD3968	202010B 10/05/2022	0.200	SD3969	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022	Maintenance By <u>DERR</u> <input checked="" type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																		
0.050	SD3967	202010A 10/05/2022																		
0.080	SD3968	202010B 10/05/2022																		
0.200	SD3969	202010D 10/06/2022																		
0.080 DGS	N/A	AG026705 09/23/2022																		
				DI Temp. Checks By <u>DERR</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.09C</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>																

Calibration Adjustment By <u>DERR</u>			
Barometric Pressure Gauge <u>1016</u> ID # <u>68639</u>			
Simulator	Serial #	Lot #	Expiration
0.000	MP5095	N/A	N/A
0.040	MP5098	20060	2/10/2022
0.100	MP5099	20420	09/22/2022
0.200	MP5100	20160	03/18/2022
0.300	MP5101	20030	01/21/2022
0.080 DGS	N/A	00521080A2	03/05/2023
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks			
Simulator	Serial #	Lot #	Expiration
0.050	SD3967	202010A	10/05/2022
0.080	SD3968	202010B	10/05/2022
0.200	SD3969	202010D	10/06/2022
0.080 DGS	N/A	AG026705	09/23/2022

Department Inspection By <u>DERR</u>	
Barometric Pressure ID# <u>28199</u>	
Gauge <u>1016</u> Instrument <u>1016</u>	
Mouth Alcohol Solution Lot # <u>2020-A</u>	
Acetone Stock Solution Lot # <u>2020-A</u>	
Simulator	Serial Number
0.000	SD3965
Interferent	SD1015
0.050	SD3967
0.080	SD3968
0.200	SD3969

Attachments	
<input checked="" type="checkbox"/> Form 41	<input checked="" type="checkbox"/> Post-Stability Checks
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration
<input checked="" type="checkbox"/> Calibration Certificate	<input checked="" type="checkbox"/> Form 40
<input checked="" type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____

Notes/Suggested Service: Optical calibration to bring Barometric values closer to nominal. DERR

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC	
<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC	
<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use	
<input type="checkbox"/> Remain Out of Evidentiary Use	
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	
2021.03.18	
<i>Michael D. Hargrave</i> 2021.03.18 12:38:42 -04'00'	<i>[Signature]</i> 2021.03.18 9 13:22:19 -04'00'
Tech Review / Date	Admin Review / Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001275, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001275</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HIALEAH GARDENS PD</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>03/18/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:00</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/18/2021

Date


DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH GARDENS PD
Time of Inspection: 12:00

Date of Inspection: 03/18/2021

Serial Number: 80-001275
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.080	0.200	0.080
0.000	0.050	0.080	0.199	0.079
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.199	0.080
Standard Deviations	0.0005	0.0004	0.0006	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

DAVID E REYES-RIVERA

03/18/2021
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-001275	Hialeah Gardens Police Department	3/18/2021	DERR <i>MLL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>HIALEAH GARDENS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001275 03/18/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:08</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MLL</i></p>	Test	g/210L	Time	Air Blank	0.000	10:06	Control Test	0.048	10:07	Air Blank	0.000	10:07	Control Test	0.049	10:08	Air Blank	0.000	10:08	Control Test	0.049	10:09	Air Blank	0.000	10:10	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>HIALEAH GARDENS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001275 03/18/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:11</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MLL</i></p>	Test	g/210L	Time	Air Blank	0.000	10:11	Control Test	0.079	10:11	Air Blank	0.000	10:12	Control Test	0.079	10:13	Air Blank	0.000	10:13	Control Test	0.079	10:14	Air Blank	0.000	10:14	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>HIALEAH GARDENS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001275 03/18/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:16</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:17</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5812</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MLL</i></p>	Test	g/210L	Time	Air Blank	0.000	10:16	Control Test	0.198	10:16	Air Blank	0.000	10:17	Control Test	0.200	10:18	Air Blank	0.000	10:18	Control Test	0.198	10:19	Air Blank	0.000	10:19	Control Test Stats			Average	0.1987		Std Dev	0.0012		Rel Std Dev(%)	0.5812		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>HIALEAH GARDENS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001275 03/18/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MLL</i></p>	Test	g/210L	Time	Air Blank	0.000	10:21	Control Test	0.080	10:21	Air Blank	0.000	10:22	Control Test	0.079	10:22	Air Blank	0.000	10:23	Control Test	0.080	10:23	Air Blank	0.000	10:24	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:06																																																																																																																																																	
Control Test	0.048	10:07																																																																																																																																																	
Air Blank	0.000	10:07																																																																																																																																																	
Control Test	0.049	10:08																																																																																																																																																	
Air Blank	0.000	10:08																																																																																																																																																	
Control Test	0.049	10:09																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:11																																																																																																																																																	
Control Test	0.079	10:11																																																																																																																																																	
Air Blank	0.000	10:12																																																																																																																																																	
Control Test	0.079	10:13																																																																																																																																																	
Air Blank	0.000	10:13																																																																																																																																																	
Control Test	0.079	10:14																																																																																																																																																	
Air Blank	0.000	10:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:16																																																																																																																																																	
Control Test	0.198	10:16																																																																																																																																																	
Air Blank	0.000	10:17																																																																																																																																																	
Control Test	0.200	10:18																																																																																																																																																	
Air Blank	0.000	10:18																																																																																																																																																	
Control Test	0.198	10:19																																																																																																																																																	
Air Blank	0.000	10:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5812																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:21																																																																																																																																																	
Control Test	0.080	10:21																																																																																																																																																	
Air Blank	0.000	10:22																																																																																																																																																	
Control Test	0.079	10:22																																																																																																																																																	
Air Blank	0.000	10:23																																																																																																																																																	
Control Test	0.080	10:23																																																																																																																																																	
Air Blank	0.000	10:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001275	Hialeah Gardens Police Department	3/18/2021	DERR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>HIALEAH GARDENS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001275 03/18/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:57</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:58</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:00</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	08:57	Control Test	0.049	08:58	Air Blank	0.000	08:58	Control Test	0.048	08:59	Air Blank	0.000	09:00	Control Test	0.049	09:00	Air Blank	0.000	09:01	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel. Std Dev(%)	1.1863		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>HIALEAH GARDENS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001275 03/18/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:02</td></tr> <tr><td>Control Test</td><td>0.077</td><td>09:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:03</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:04</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	09:02	Control Test	0.077	09:03	Air Blank	0.000	09:03	Control Test	0.078	09:04	Air Blank	0.000	09:04	Control Test	0.078	09:05	Air Blank	0.000	09:06	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel. Std Dev(%)	0.7434		<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>HIALEAH GARDENS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001275 03/18/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr> <tr><td>Control Test</td><td>0.192</td><td>09:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:08</td></tr> <tr><td>Control Test</td><td>0.193</td><td>09:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:09</td></tr> <tr><td>Control Test</td><td>0.193</td><td>09:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1927</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2997</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	09:07	Control Test	0.192	09:07	Air Blank	0.000	09:08	Control Test	0.193	09:08	Air Blank	0.000	09:09	Control Test	0.193	09:10	Air Blank	0.000	09:10	Control Test Stats			Average	0.1927		Std Dev	0.0006		Rel. Std Dev(%)	0.2997		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>HIALEAH GARDENS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001275 03/18/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:12</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:13</td></tr> <tr><td>Control Test</td><td>0.077</td><td>09:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:14</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	09:12	Control Test	0.078	09:12	Air Blank	0.000	09:13	Control Test	0.077	09:13	Air Blank	0.000	09:14	Control Test	0.078	09:14	Air Blank	0.000	09:14	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel. Std Dev(%)	0.7434	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:57																																																																																																																																																	
Control Test	0.049	08:58																																																																																																																																																	
Air Blank	0.000	08:58																																																																																																																																																	
Control Test	0.048	08:59																																																																																																																																																	
Air Blank	0.000	09:00																																																																																																																																																	
Control Test	0.049	09:00																																																																																																																																																	
Air Blank	0.000	09:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:02																																																																																																																																																	
Control Test	0.077	09:03																																																																																																																																																	
Air Blank	0.000	09:03																																																																																																																																																	
Control Test	0.078	09:04																																																																																																																																																	
Air Blank	0.000	09:04																																																																																																																																																	
Control Test	0.078	09:05																																																																																																																																																	
Air Blank	0.000	09:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:07																																																																																																																																																	
Control Test	0.192	09:07																																																																																																																																																	
Air Blank	0.000	09:08																																																																																																																																																	
Control Test	0.193	09:08																																																																																																																																																	
Air Blank	0.000	09:09																																																																																																																																																	
Control Test	0.193	09:10																																																																																																																																																	
Air Blank	0.000	09:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1927																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2997																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:12																																																																																																																																																	
Control Test	0.078	09:12																																																																																																																																																	
Air Blank	0.000	09:13																																																																																																																																																	
Control Test	0.077	09:13																																																																																																																																																	
Air Blank	0.000	09:14																																																																																																																																																	
Control Test	0.078	09:14																																																																																																																																																	
Air Blank	0.000	09:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7434																																																																																																																																																		

HIALEAH GARDENS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001275
03/18/2021
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
Sqrt(Diff)) = 5.914
2: Rate (Liters/min) = 15
Sqrt(Diff)) = 11.398
3: Rate (Liters/min) = 30
Sqrt(Diff)) = 21.023

Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 642
Rounded Intercept = -433738
Correlation = 0.99912

Flow Calibration	
SN:	80-0001275
Agency:	Hialeah Gardens Police Department
Date:	3/18/2021
By:	DERR <i>[Signature]</i>

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH GARDENS PD
Time of Inspection: 08:46

Date of Inspection: 03/18/2021

Serial Number: 80-001275
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

DAVID E REYES-RIVERA

03/18/2021
 Date