



INSTRUMENT PROCESSING SHEET

Agency Jacksonville SOS/N 80-001279Florida Department of
Law EnforcementDate In 12/3/2021 DI Completion Date 12/3/2021
☒ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By <u>RAW</u>	Quality Checks	By <u>RAW</u> Date <u>12/3/21</u>	Flow Calibration	By	Date														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Instrument coming in for <u>Department Inspection prior to</u> <u>being sent to CMI for yearly</u> <u>preventative maintenance per</u> <u>agency contract.-RAW, 12/3/21</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>214</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.250</u> (.228 - .278) 103 mm <u>.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A 10-05-22</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202010B 10-05-22</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D 10-06-22</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403 05-14-23</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	202010A 10-05-22	0.080	MP5089	202010B 10-05-22	0.200	MP5090	202010D 10-06-22	0.080 DGS	N/A	AG113403 05-14-23	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____	
Simulator	Serial #	Lot #/Exp																		
0.050	MP5088	202010A 10-05-22																		
0.080	MP5089	202010B 10-05-22																		
0.200	MP5090	202010D 10-06-22																		
0.080 DGS	N/A	AG113403 05-14-23																		
				DI Temp. Checks By <u>IS</u> ☒ Lab Temp °C <u>23.58</u> External Digital Therm. ID#: <u>381189</u> ☒ 34°C +/- .2 Serial #: <u>MP5088</u> ☒ 34°C +/- .2 Serial #: <u>MP5089</u> ☒ 34°C +/- .2 Serial #: <u>MP5090</u>																

Calibration Adjustment	By	Department Inspection	By <u>RAW</u>																																																												
Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Barometric Pressure ID# <u>30793</u> Gauge <u>1016</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP5086																																																														
Interferent	MP5087																																																														
0.050	MP5088																																																														
0.080	MP5089																																																														
0.200	MP5090																																																														
		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																																																													

Notes/Suggested Service: Admin Review
Amended Shipping location and
Evidentiary use sections - RAW, 12/3/21

- ☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC
☒ Return to/Place into Evidentiary Use
☒ Remain Out of Evidentiary Use RAW, 12/3/21
☒ Conduct an Agency Inspection Before Evidentiary Use

 David Eliezer
 Reyes-Rivera

 Digitally signed by David
 Reyes-Rivera
 Date: 2021.12.03 08:59:36
 -05'00'

 5 13:48:03
 -05'00'

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 16:06

Date of Inspection: 12/03/2021

Serial Number: 80-001279
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.051	0.080	0.203	0.080
0.000	0.050	0.080	0.203	0.080
0.000	0.050	0.080	0.203	0.080
0.000	0.051	0.080	0.203	0.080
0.000	0.051	0.080	0.202	0.080
0.000	0.050	0.080	0.203	0.080
0.000	0.050	0.081	0.203	0.080
0.000	0.051	0.081	0.203	0.080
0.000	0.051	0.081	0.203	0.080
0.000	0.051	0.081	0.203	0.080

Standard Deviations	0.0005	0.0005	0.0003	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 RICHARD A WILLIAMS
Signature and Printed Name

12/03/2021
Date

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001279
12/03/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:56
Control Test	0.050	13:56
Air Blank	0.000	13:57
Control Test	0.049	13:57
Air Blank	0.000	13:58
Control Test	0.050	13:59
Air Blank	0.000	13:59
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

.05

RAW

Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001279
12/03/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:03
Control Test	0.080	14:04
Air Blank	0.000	14:05
Control Test	0.080	14:05
Air Blank	0.000	14:06
Control Test	0.080	14:06
Air Blank	0.000	14:07
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

.08
wet

RAW

Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001279
12/03/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:08
Control Test	0.203	14:09
Air Blank	0.000	14:09
Control Test	0.203	14:10
Air Blank	0.000	14:11
Control Test	0.202	14:11
Air Blank	0.000	14:12
Control Test Stats		
Average	0.2027	
Std Dev	0.0006	
Rel Std Dev(%)	0.2849	

.20

RAW

Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001279
12/03/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:16
Control Test	0.080	14:17
Air Blank	0.000	14:17
Control Test	0.080	14:18
Air Blank	0.000	14:18
Control Test	0.080	14:18
Air Blank	0.000	14:19
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

.08
DGS

RAW

Operator's Signature

Stability Checks

80-001279

12/3/21

RAW



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001279, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001279</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>JACKSONVILLE SO</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>12/03/2021</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>16:06</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Digitally signed by Richard A Williams
Date: 2021.12.03 16:37:44 -05'00'

Richard A Williams

12/03/2021

Date

RICHARD A WILLIAMS,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality