

Return Material Authorization

Ship to: CMI, Inc.
 Enforcement Electronics

Shipment to repair facility authorized by: Kellie Melville on 08/23/2021

Items Returned: Instrument Supplies Other Describe: _____
Instrument Model: Intoxilyzer 8000 Serial Number: 80-001418

Bill To Address: <u>Orange County SO</u> _____ _____ _____	Ship to Address: <u>FDLE ATP</u> <u>Attn: Michael Haughey</u> <u>4700 Terminal Drive Suite 1</u> <u>Fort Myers FL 33907</u> _____ _____
-------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------

Reason for Return:
Error 012 code displayed.

Please choose one of the following options:

1. I _____, authorize all repairs.

2. I _____, authorize repairs up to \$ _____.

3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Kellie Melville
Phone #: 407-836-0804 Email: Kelly.Melville@ocfl.net
ATP Contact Name: Michael Haughey ATP Email: michaelhaughey@fdle.state.fl.us

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Orange Co. 80
Time of Inspection:

Date of Inspection:

Serial Number: 80-001418
Software:

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations			
---------------------	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

Instrument did not receive 2021 Department Inspection due to being at repair. Instrument sent to repair directly by agency. *from FMZOC on 08/23/21. -BAW, 01/12/22*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Handwritten Signature]

Signature and Printed Name

12/29/21
Date