

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: St. Petersburg PD
Time of Inspection: n/a

Date of Inspection: n/a

Serial Number: 80-001653
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Standard Deviations				
---------------------	--	--	--	--


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

Was unable to inspect the instrument. During the Quality Checks, it was observed that the instrument could not draw samples from the wet-bath simulators. Instrument was sent to repair and did not return within the calendar year.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

Taylor D. Gutschow

 1/18/2022
 Date

Return Material Authorization

Ship to: CMI, Inc.
 Enforcement Electronics

Shipment to repair facility authorized by: Sean McCullough on 11/5/2021

Items Returned: Instrument Supplies Other Describe: _____
Instrument Model: Intoxilyzer 8000 Serial Number: 80-001653

Bill To Address: <u>St. Petersburg Police Department</u> <u>Attn: Sean McCullough</u> <u>1300 First Avenue North</u> <u>St. Petersburg, FL 33705</u>	Ship to Address: <u>Florida Department of Law Enforcement</u> <u>Fort Myers Regional Operations Center</u> <u>Attn: Taylor Gutschow (Alcohol Testing)</u> <u>4700 Terminal Drive, Suite 1</u> <u>Fort Myers, FL 33907</u>
---	---

Reason for Return:
Instrument will not draw samples from wet-bath simulators but WILL draw samples from the dry gas cylinder. No bubbles seen in simulator liquid. Simulator issue ruled out. Instrument was able to draw wet bath samples during the Agency Inspection several days ago, so the issue is believed to have arisen during transit. All records uploaded to COBRA. No records remain on the instrument.

Please choose one of the following options:

1. I _____, authorize all repairs.

2. I _____, authorize repairs up to \$_____.

3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Sean McCullough
Phone #: 727-401-1391 Email: Sean.McCullough@stpete.org
ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency St. Petersburg PD

S/N 80-001653

Florida Department of Law Enforcement

Date In 11/1/2021

DI Completion Date

Ship P/U H/D CMI EE

Intake	By TDG	Quality Checks	By TDG	Date 11/4/2021	Flow Calibration	By	Date															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Double boxed		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value 195 <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # ATP101 32 mm 0.164 (.139 - .169) 36 mm 0.183 (.156 - .190) 53 mm 0.250 (.228 - .278) 103 mm 0.527 (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # 68639 <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																	
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5092</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705 09/23/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A 10/05/2022	0.080	MP4864	202010B 10/05/2022	0.200	MP5094	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022			Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ DI Temp. Checks By _____ <input type="checkbox"/> Lab Temp °C _____ External Digital Therm. ID#: _____ <input type="checkbox"/> 34°C +/- .2 Serial #: _____ <input type="checkbox"/> 34°C +/- .2 Serial #: _____ <input type="checkbox"/> 34°C +/- .2 Serial #: _____		
Simulator	Serial #	Lot #/Exp																				
0.050	MP5092	202010A 10/05/2022																				
0.080	MP4864	202010B 10/05/2022																				
0.200	MP5094	202010D 10/06/2022																				
0.080 DGS	N/A	AG026705 09/23/2022																				

Calibration Adjustment	By	Department Inspection	By																																																												
Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td></tr> <tr><td>Interferent</td><td></td></tr> <tr><td>0.050</td><td></td></tr> <tr><td>0.080</td><td></td></tr> <tr><td>0.200</td><td></td></tr> </tbody> </table>	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200		
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000																																																															
Interferent																																																															
0.050																																																															
0.080																																																															
0.200																																																															
Notes/Suggested Service: Instrument will not draw samples from wet bath simulators. Cannot inspect. Will send to repair facility (TDG).		Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other Form 51																																																													
		<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													
		Tech Review / Date _____ Admin Review / Date _____																																																													

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1653	St. Petersburg PD	11/04/2021	TDG ML

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet X																																																																																																																																																
<p>ST. PETERSBURG PD INSTRUMENT ID: 4083287 Model: 8000 SN: 80-101653 11/04/2021 Software: 8110.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>ML Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:00	Control Test	0.000	11:01	Air Blank	0.000	11:01	Control Test	0.000	11:02	Air Blank	0.000	11:02	Control Test	0.000	11:03	Air Blank	0.000	11:03	Control Test Stats			Average	0.0000		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>ST. PETERSBURG PD INSTRUMENT ID: 4083287 Model: 8000 SN: 80-101653 11/04/2021 Software: 8110.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:09</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>ML Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:09	Control Test	0.000	11:10	Air Blank	0.000	11:11	Control Test	0.000	11:11	Air Blank	0.000	11:12	Control Test	0.000	11:12	Air Blank	0.000	11:13	Control Test Stats			Average	0.0000		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>ST. PETERSBURG PD INSTRUMENT ID: 4083287 Model: 8000 SN: 80-101653 11/04/2021 Software: 8110.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>ML Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:17	Control Test	0.000	11:18	Air Blank	0.000	11:18	Control Test	0.000	11:19	Air Blank	0.000	11:19	Control Test	0.000	11:20	Air Blank	0.000	11:20	Control Test Stats			Average	0.0000		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>ST. PETERSBURG PD INSTRUMENT ID: 4083287 Model: 8000 SN: 80-101653 11/04/2021 Software: 8110.27</p> <p style="text-align: center;">DGS</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test</td><td>0.082</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr> <tr><td>Control Test</td><td>0.081</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.2346</td><td></td></tr> </tbody> </table> <p>ML Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:21	Control Test	0.082	11:21	Air Blank	0.000	11:22	Control Test	0.080	11:22	Air Blank	0.000	11:23	Control Test	0.081	11:23	Air Blank	0.000	11:24	Control Test Stats			Average	0.0810		Std Dev	0.0010		Rel. Std Dev(%)	1.2346	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:00																																																																																																																																																	
Control Test	0.000	11:01																																																																																																																																																	
Air Blank	0.000	11:01																																																																																																																																																	
Control Test	0.000	11:02																																																																																																																																																	
Air Blank	0.000	11:02																																																																																																																																																	
Control Test	0.000	11:03																																																																																																																																																	
Air Blank	0.000	11:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0000																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:09																																																																																																																																																	
Control Test	0.000	11:10																																																																																																																																																	
Air Blank	0.000	11:11																																																																																																																																																	
Control Test	0.000	11:11																																																																																																																																																	
Air Blank	0.000	11:12																																																																																																																																																	
Control Test	0.000	11:12																																																																																																																																																	
Air Blank	0.000	11:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0000																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:17																																																																																																																																																	
Control Test	0.000	11:18																																																																																																																																																	
Air Blank	0.000	11:18																																																																																																																																																	
Control Test	0.000	11:19																																																																																																																																																	
Air Blank	0.000	11:19																																																																																																																																																	
Control Test	0.000	11:20																																																																																																																																																	
Air Blank	0.000	11:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0000																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:21																																																																																																																																																	
Control Test	0.082	11:21																																																																																																																																																	
Air Blank	0.000	11:22																																																																																																																																																	
Control Test	0.080	11:22																																																																																																																																																	
Air Blank	0.000	11:23																																																																																																																																																	
Control Test	0.081	11:23																																																																																																																																																	
Air Blank	0.000	11:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel. Std Dev(%)	1.2346																																																																																																																																																		

Comments: Instrument not pulling samples from sims. No bubbles observed during wet-bath control tests. Sims worked fine on multiple other instruments on bench. Instrument was able to pull wet-bath samples during the 10/28/2021 Agency Inspection. ML 11/04/2021

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT – INTOXILYZER 8000

Agency: St. Petersburg PD
Time of Inspection: 11:30

Date of Inspection: 11/04/2021

Serial Number: 80-001653
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:
I was unable to inspect instrument 80-001653. During the Quality Checks, it was observed that the instrument could not draw samples from the wet-bath simulators. The instrument was successfully able to do so during the Agency Inspection conducted several days prior. The instrument will be sent to repair.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.
I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor Gutschow
Taylor Gutschow

Signature and Printed Name

11/04/2021
Date