



INSTRUMENT PROCESSING SHEET

Agency Palm Beach County SO

S/N 80-001741

Florida Department of
Law Enforcement

Date In 3/9/2021

DI Completion Date 3/9/2021

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By MH	Quality Checks	By MH	Date <u>03/09/2021</u>	Flow Calibration	By	Date																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>217</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 01/30/2022</td> </tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050	MP4863	202010A 10/05/2022	0.080	MP4864	202010B 10/05/2022	0.200	MP5097	202010D 10/06/2022	0.080 DGS	N/A	AG003005 01/30/2022	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																											
Simulator	Serial #	Lot #/Exp																																													
0.050	MP4863	202010A 10/05/2022																																													
0.080	MP4864	202010B 10/05/2022																																													
0.200	MP5097	202010D 10/06/2022																																													
0.080 DGS	N/A	AG003005 01/30/2022																																													
		DI Temp. Checks By MH _____ ☒ Lab Temp °C <u>23.26</u> External Digital Therm. ID#: <u>300504</u> ☒ 34°C +/- .2 Serial #: <u>MP4863</u> ☒ 34°C +/- .2 Serial #: <u>MP4864</u> ☒ 34°C +/- .2 Serial #: <u>MP5097</u>																																													
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____				Department Inspection By MH _____ Barometric Pressure ID# <u>28663</u> Gauge <u>1031</u> Instrument <u>1026</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u>																																											
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	SD1014																																														
Interferent	SD1015																																														
0.050	MP4863																																														
0.080	MP4864																																														
0.200	MP5097																																														
☐ Post Calibration Adjustment Stability Checks				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																											
Notes/Suggested Service: <u>Tech Review: Added Date to Stability</u> <u>Form MH 3/9/21</u> _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2021.03. Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2021.03.09 13:11:12 -05'00' 10 08:47:36																																											
Tech Review / Date				Admin Review / Date																																											

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO
Time of Inspection: 11:55

Date of Inspection: 03/09/2021

Serial Number: 80-001741
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.080	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.080	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079


Standard Deviations	0.0004	0.0004	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 MICHAEL D HAUGHEY
Signature and Printed Name
03/09/2021
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001774	Palm Beach County, SD	80-001774	ML

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒

PALYN SEARCH CO 50
Intoxilyzer - Alcotest Analyzer
Model 8000
03/09/2021
SN 80-001741
Software: 8100.27

PALM BEACH CO SO
Intoxilyzer - Alconol Analyzer
Model 8000
03/03/2021
SN 80-001741
Software: 8100.27

PALM BEACH CO 50
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-001741
03/09/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:25
Control Test	0.050	09:26
Air Blank	0.000	09:26
Control Test	0.050	09:27
Air Blank	0.000	09:27
Control Test	0.050	09:28
Air Blank	0.000	09:29
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/2.0L	Time
Air Blank	0.000	09:30
Control Test	0.090	09:31
Air Blank	0.000	09:31
Control Test	0.090	09:32
Air Blank	0.000	09:33
Control Test	0.080	09:33
Air Blank	0.000	09:34
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/20L	Time
Air Blank	0.000	09:36
Control Test	0.998	09:37
Air Blank	0.000	09:37
Control Test	0.197	09:38
Air Blank	0.000	09:38
Control Test	0.197	09:39
Air Blank	0.000	09:40
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	


Test	g2JUL	Time
Air Blank	0.000	09:43
Control Test	0.079	09:43
Air Blank	0.000	09:44
Control Test	0.079	09:44
Air Blank	0.000	09:45
Control Test	0.079	09:45
Air Blank	0.000	09:45
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Signature _____

Operator = Signature

AK

Signature _____



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001741, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001741</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PALM BEACH CO SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>03/09/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:55</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/09/2021

Date

MICHAEL D HAUGHEY,
Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1