

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:30

Date of Inspection: 07/23/2021

Serial Number: 80-003409
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.079	0.197	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.080	0.198	0.078
0.000	0.049	0.080	0.197	0.078
0.000	0.050	0.080	0.197	0.078
0.000	0.049	0.080	0.197	0.079


Standard Deviations	0.0006	0.0005	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY
Signature and Printed Name

07/23/2021
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-003409	FHPD	07/23/21	AK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒

See External Print-Out

FL HIGHWAY PATROL
Intoxilyzer - Alcotest Analyzer
Model 8000
07/23/2021
SN 80-003409
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:11
Control Test	0.078	09:12
Air Blank	0.000	09:12
Control Test	0.078	09:13
Air Blank	0.000	09:14
Control Test	0.077	09:14
Air Blank	0.000	09:15
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

Operator's Signature *AK*

FL HIGHWAY PATROL
Intoxilyzer - Alcotest Analyzer
Model 8000
07/23/2021
SN 80-003409
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:19
Control Test	0.196	09:20
Air Blank	0.000	09:20
Control Test	0.196	09:21
Air Blank	0.000	09:21
Control Test	0.197	09:22
Air Blank	0.000	09:22
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

Operator's Signature *AK*

FL HIGHWAY PATROL
Intoxilyzer - Alcotest Analyzer
Model 8000
07/23/2021
SN 80-003409
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:24
Control Test	0.078	09:24
Air Blank	0.000	09:25
Control Test	0.079	09:25
Air Blank	0.000	09:26
Control Test	0.078	09:26
Air Blank	0.000	09:26
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Operator's Signature *AK*

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003409
07/23/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:06
Control Test	0.048	09:07
Air Blank	0.000	09:07
Control Test	0.048	09:08
Air Blank	0.000	09:09
Control Test	0.048	09:09
Air Blank	0.000	09:10
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003409, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003409</u>	UNCERTAINTY * $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	
Calibration Date:	<u>07/23/2021</u>	0.050 g/ 210 L 0.005
Calibration Time:	<u>11:30</u>	0.080 g/ 210 L 0.004
		0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after an adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/23/2021

Date


MICHAEL D HAUGHEY,
Department Inspector

FDLE/ATP Form 69 January 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Time of Inspection: 08:49

Date of Inspection: 07/23/2021

Serial Number: 80-003409

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No
		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED. AI BYPASS.

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL D HAUGHEY

Signature and Printed Name

07/23/2021

Date