



INSTRUMENT PROCESSING SHEET

Agency Osceola County SO

S/N 80-004587

Florida Department of
Law EnforcementDate In 1/22/2021DI Completion Date 1/27/2021☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By MH	Quality Checks	By MH	Date	1/27/2021	Flow Calibration	By	Date																																																									
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>216</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 106</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks																																																															
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP4863</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP4864</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP5097</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG003005 01/30/2022</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	202010A 10/05/2022	0.080	MP4864	202010B 10/05/2022	0.200	MP5097	202010D 10/06/2022	0.080 DGS	N/A	AG003005 01/30/2022																																																
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP4863	202010A 10/05/2022																																																															
0.080	MP4864	202010B 10/05/2022																																																															
0.200	MP5097	202010D 10/06/2022																																																															
0.080 DGS	N/A	AG003005 01/30/2022																																																															
						Flow Column # ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value ☐ Post Calibration Verification (L/s) Flow Column # 32 mm (.139 - .169) 36 mm (.156 - .190) 53 mm (.228 - .278) 103 mm (.447 - .547)																																																											
						Maintenance ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other																																																											
						DI Temp. Checks ☒ Lab Temp °C <u>21.47</u> External Digital Therm. ID#: <u>300504</u> ☒ 34°C +/-2 Serial #: <u>MP4863</u> ☒ 34°C +/-2 Serial #: <u>MP4864</u> ☒ 34°C +/-2 Serial #: <u>MP5097</u>																																																											
Calibration Adjustment Barometric Pressure Gauge ID # <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> Notes/Suggested Service:		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Department Inspection Barometric Pressure ID# <u>28663</u> Gauge <u>1019</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>SD1014</td></tr><tr><td>Interferent</td><td>SD1015</td></tr><tr><td>0.050</td><td>MP4863</td></tr><tr><td>0.080</td><td>MP4864</td></tr><tr><td>0.200</td><td>MP5097</td></tr></tbody></table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use David Eliezer Reyes-Rivera 2021-01-27 15:06:20 -05'00' Tech Review / Date Admin Review / Date	Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097		
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	SD1014																																																																
Interferent	SD1015																																																																
0.050	MP4863																																																																
0.080	MP4864																																																																
0.200	MP5097																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 12:58

Date of Inspection: 01/27/2021

Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.079	0.199	0.078
0.000	0.049	0.079	0.200	0.078


Standard Deviations	0.0004	0.0004	0.0004	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

01/27/2021
Date

DERR
Digitally signed
by DERR
Date: 2021.01.27
15:04:42 -05'00'

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-004587	Osceola County SO	8/27/2021	MM

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.077 to 0.083 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
OSCEOLA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2021 Software: 8100.27 SN 80-004587 Test g/210L Time Air Blank 0.000 10:00 Control Test 0.049 10:01 Air Blank 0.000 10:02 Control Test 0.049 10:02 Air Blank 0.000 10:03 Control Test 0.049 10:03 Air Blank 0.000 10:04 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000	OSCEOLA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2021 Software: 8100.27 SN 80-004587 Test g/210L Time Air Blank 0.000 10:06 Control Test 0.079 10:07 Air Blank 0.000 10:08 Control Test 0.080 10:08 Air Blank 0.000 10:09 Control Test 0.080 10:10 Air Blank 0.000 10:10 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247	OSCEOLA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2021 Software: 8100.27 SN 80-004587 Test g/210L Time Air Blank 0.000 10:12 Control Test 0.200 10:13 Air Blank 0.000 10:13 Control Test 0.200 10:14 Air Blank 0.000 10:14 Control Test 0.200 10:15 Air Blank 0.000 10:16 Control Test Stats Average 0.2000 Std Dev 0.0000 Rel Std Dev(%) 0.0000	OSCEOLA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2021 Software: 8100.27 SN 80-004587 Test g/210L Time Air Blank 0.000 10:17 Control Test 0.079 10:17 Air Blank 0.000 10:18 Control Test 0.080 10:18 Air Blank 0.000 10:19 Control Test 0.079 10:19 Air Blank 0.000 10:20 Control Test Stats Average 0.0793 Std Dev 0.0005 Rel Std Dev(%) 0.7277

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature

Digitally signed
 by DERR
 Date: 2021.01.27
 15:03:58 -05'00'



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-004587, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-004587</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>OSCEOLA COUNTY S.O.</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>01/27/2021</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:58</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2021
Issuing Authority: Alcohol Testing Program

01/27/2021

Date


MICHAEL D'HAUGHEY,
Department Inspector

Service • Integrity • Respect • Quality

Page 1 of 1

DERR
Digitally signed
by DERR
Date: 2021.01.27
15:03:12 -05'00'