

Return Material Authorization

Ship to: CMI, Inc.
 Enforcement Electronics

Shipment to repair facility authorized by: William Jerard on 04-02-2021

Items Returned: Instrument Supplies Other Describe: _____
Instrument Model: Intoxilyzer 8000 Serial Number: 80-005057

Bill To Address:
Okaloosa County Sheriff's Office

Ship to Address:
Alcohol Testing Program FDLE
Tallahassee, Florida

Reason for Return:
R-value below 100, instrument needs flow sensor replacement.

Please choose one of the following options:

- 1. I _____, authorize all repairs.
- 2. I _____, authorize repairs up to \$_____.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: William Jerard
Phone #: 850-609-4947 Email: bjerard@sheriff-okaloosa.org

ATP Contact Name: Israel Soto ATP Email: israelsoto@fdle.state.fl.us

