

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-005113	Hillsborough CSO	05/11/2021	TDG <i>ML</i>

0.05g/210L 0.047 to 0.053		0.08g/210L 0.077 to 0.083		0.20g/210L 0.194 to 0.206		DGS 0.08g/210L 0.077 to 0.083		≤0.003 of Wet																																																																																																																																			
✓		✓		✓		✓		✓																																																																																																																																			
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 05/11/2021 Software: 8100.27 Test Time <table><tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr><tr><td>Control Test</td><td>0.048</td><td>12:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr><tr><td>Control Test</td><td>0.048</td><td>12:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr><tr><td>Control Test</td><td>0.048</td><td>12:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:08</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0480</td><td></td></tr><tr><td>Std Dev</td><td>0.0003</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> Operator's Signature <i>ML</i>		Air Blank	0.000	12:04	Control Test	0.048	12:05	Air Blank	0.000	12:05	Control Test	0.048	12:06	Air Blank	0.000	12:07	Control Test	0.048	12:07	Air Blank	0.000	12:08	Control Test Stats			Average	0.0480		Std Dev	0.0003		Rel Std Dev(%)	0.0000		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 05/11/2021 Software: 8100.27 Test Time <table><tr><td>Air Blank</td><td>0.000</td><td>12:11</td></tr><tr><td>Control Test</td><td>0.078</td><td>12:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:12</td></tr><tr><td>Control Test</td><td>0.077</td><td>12:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:13</td></tr><tr><td>Control Test</td><td>0.077</td><td>12:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0773</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7456</td><td></td></tr></table> Operator's Signature <i>ML</i>		Air Blank	0.000	12:11	Control Test	0.078	12:12	Air Blank	0.000	12:12	Control Test	0.077	12:13	Air Blank	0.000	12:13	Control Test	0.077	12:14	Air Blank	0.000	12:15	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7456		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 05/11/2021 Software: 8100.27 Test Time <table><tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr><tr><td>Control Test</td><td>0.194</td><td>12:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr><tr><td>Control Test</td><td>0.194</td><td>12:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr><tr><td>Control Test</td><td>0.194</td><td>12:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1940</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> Operator's Signature <i>ML</i>		Air Blank	0.000	12:17	Control Test	0.194	12:18	Air Blank	0.000	12:18	Control Test	0.194	12:19	Air Blank	0.000	12:20	Control Test	0.194	12:20	Air Blank	0.000	12:21	Control Test Stats			Average	0.1940		Std Dev	0.0000		Rel Std Dev(%)	0.0000		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 05/11/2021 Software: 8100.27 Test Time <table><tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:23</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:24</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> Operator's Signature <i>ML</i>		Air Blank	0.000	12:22	Control Test	0.079	12:23	Air Blank	0.000	12:23	Control Test	0.079	12:23	Air Blank	0.000	12:24	Control Test	0.079	12:24	Air Blank	0.000	12:25	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	12:04																																																																																																																																									
Control Test	0.048	12:05																																																																																																																																									
Air Blank	0.000	12:05																																																																																																																																									
Control Test	0.048	12:06																																																																																																																																									
Air Blank	0.000	12:07																																																																																																																																									
Control Test	0.048	12:07																																																																																																																																									
Air Blank	0.000	12:08																																																																																																																																									
Control Test Stats																																																																																																																																											
Average	0.0480																																																																																																																																										
Std Dev	0.0003																																																																																																																																										
Rel Std Dev(%)	0.0000																																																																																																																																										
Air Blank	0.000	12:11																																																																																																																																									
Control Test	0.078	12:12																																																																																																																																									
Air Blank	0.000	12:12																																																																																																																																									
Control Test	0.077	12:13																																																																																																																																									
Air Blank	0.000	12:13																																																																																																																																									
Control Test	0.077	12:14																																																																																																																																									
Air Blank	0.000	12:15																																																																																																																																									
Control Test Stats																																																																																																																																											
Average	0.0773																																																																																																																																										
Std Dev	0.0006																																																																																																																																										
Rel Std Dev(%)	0.7456																																																																																																																																										
Air Blank	0.000	12:17																																																																																																																																									
Control Test	0.194	12:18																																																																																																																																									
Air Blank	0.000	12:18																																																																																																																																									
Control Test	0.194	12:19																																																																																																																																									
Air Blank	0.000	12:20																																																																																																																																									
Control Test	0.194	12:20																																																																																																																																									
Air Blank	0.000	12:21																																																																																																																																									
Control Test Stats																																																																																																																																											
Average	0.1940																																																																																																																																										
Std Dev	0.0000																																																																																																																																										
Rel Std Dev(%)	0.0000																																																																																																																																										
Air Blank	0.000	12:22																																																																																																																																									
Control Test	0.079	12:23																																																																																																																																									
Air Blank	0.000	12:23																																																																																																																																									
Control Test	0.079	12:23																																																																																																																																									
Air Blank	0.000	12:24																																																																																																																																									
Control Test	0.079	12:24																																																																																																																																									
Air Blank	0.000	12:25																																																																																																																																									
Control Test Stats																																																																																																																																											
Average	0.0790																																																																																																																																										
Std Dev	0.0000																																																																																																																																										
Rel Std Dev(%)	0.0000																																																																																																																																										

Comments:

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 14:23

Date of Inspection: 05/11/2021

Serial Number: 80-005113
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.048	0.078	0.193	0.079
0.000	0.049	0.078	0.194	0.078
0.000	0.048	0.078	0.194	0.078
0.000	0.048	0.078	0.194	0.078
0.000	0.049	0.078	0.194	0.078
0.000	0.049	0.078	0.194	0.078
0.000	0.049	0.078	0.194	0.078
0.000	0.049	0.078	0.194	0.078
0.000	0.049	0.078	0.194	0.078
0.000	0.049	0.078	0.194	0.078

Standard Deviations	0.0004	0.0000	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D. GUTSCHOW

Signature and Printed Name

05/11/2021
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005113, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005113</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>05/11/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:23</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/11/2021

Date

TAYLOR D GUTSCHOW,

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to:

☐ CMI, Inc.

☒ Enforcement Electronics

Shipment to repair facility authorized by: Roger Skipper on 03/01/2021

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-005113

Bill To Address:
Hillsborough County Sheriff's Office

Attn: Roger Skipper

1201 Orient Road

Tampa, FL 33619

Ship to Address:
Florida Department of Law Enforcement

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Instrument's R-Value was less than 100.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Roger Skipper

Phone #: 813-267-7154

Email: rskipper@teamhcsso.com

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency Hillsborough CSO

S/N 80-005113

Florida Department of
Law Enforcement

Date In 2/26/2021

DI Completion Date _____

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☒ EE

Intake	By <u>TDG</u>	Quality Checks	By <u>TDG</u> Date <u>3/1/2021</u>	Flow Calibration	By _____	Date _____																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>82</u> ☐ Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks		Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
				Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																														
				DI Temp. Checks By _____ ☐ Lab Temp °C _____ External Digital Therm. ID#: _____ ☐ 34°C +/- .2 Serial #: _____ ☐ 34°C +/- .2 Serial #: _____ ☐ 34°C +/- .2 Serial #: _____																														
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:25%;">Simulator</th> <th style="width:25%;">Serial #</th> <th style="width:25%;">Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> </tr> </table>							Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A														
Simulator	Serial #	Lot #/Exp																																
0.050																																		
0.080																																		
0.200																																		
0.080 DGS	N/A																																	
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:25%;">Simulator</th> <th style="width:25%;">Serial #</th> <th style="width:25%;">Lot #</th> <th style="width:25%;">Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>							Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A		
Simulator	Serial #	Lot #	Expiration																															
0.000		N/A	N/A																															
0.040																																		
0.100																																		
0.200																																		
0.300																																		
0.080 DGS	N/A																																	
☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:25%;">Simulator</th> <th style="width:25%;">Serial #</th> <th style="width:25%;">Lot #</th> <th style="width:25%;">Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>							Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A										
Simulator	Serial #	Lot #	Expiration																															
0.050																																		
0.080																																		
0.200																																		
0.080 DGS	N/A																																	
Notes/Suggested Service: <u>R-Value was 82. Per SOP, I will</u> <u>send to a repair facility. AI prefers EE. All records were</u> <u>uploaded prior to shipping out. (TDG 3/1/2021)</u> _____ _____ _____ _____ _____																																		
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:50%;">Attachments</th> <th style="width:50%;">By _____</th> </tr> <tr> <td> ☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment </td> <td> ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____ </td> </tr> </table>							Attachments	By _____	☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____																								
Attachments	By _____																																	
☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____																																	
☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place Into Evidentiary Use ☐ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																		
<table style="width:100%;"> <tr> <td style="width:50%; text-align: center;">Tech Review / Date _____</td> <td style="width:50%; text-align: center;">Admin Review / Date _____</td> </tr> </table>							Tech Review / Date _____	Admin Review / Date _____																										
Tech Review / Date _____	Admin Review / Date _____																																	

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 09:52

Date of Inspection: 03/01/2021

Serial Number: 80-005113
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI. COMPLIANCE NOT DETERMINED.

Not determined *MG*
3/1/2021

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

Signature and Printed Name

TAYLOR D GUTSCHOW

03/01/2021
Date