



INSTRUMENT PROCESSING SHEET

Agency Sarasota County SOS/N 80-005936

Florida Department of Law Enforcement

Date In 9/20/2021 DI Completion Date 9/21/2021 Ship P/U H/D CMI EE

Intake By MH

Annual
 Registration
 Return from CMI / EE

Visual Inspection:
 Case Handle
 Keyboard Dry Gas Shelf
 Feet Breath Tube
 Ports Screws Tight

Other Equipment/ Accessories:
 Power cord Printer Cable
 Static Bag 12V DC Cable

Notes: _____

Quality Checks By MH Date 09/21/2021

Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 198
 Flow Verification (L/s)
 Flow Column # ATP 106
 32 mm 0.144 (.139 - .169)
 36 mm 0.160 (.156 - .190)
 53 mm 0.234 (.228 - .278)
 103 mm 0.496 (.447 - .547)
 Barometric Pressure Check
 Gauge ID # 28199
 Stability Checks

Flow Calibration By _____ Date _____

Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32 mm _____ (.139 - .169)
 36 mm _____ (.156 - .190)
 53 mm _____ (.228 - .278)
 103 mm _____ (.447 - .547)

Simulator	Serial #	Lot #/Exp
0.050	MP4863	202010A
		10/05/2022
0.080	MP4864	202010B
		10/05/2022
0.200	MP5097	202010D
		10/06/2022
0.080 DGS	N/A	AG026705
		09/23/2022

Maintenance By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

DI Temp. Checks By MH

Lab Temp °C 21.33
 External Digital Therm. ID#: 381198
 34°C +/- .2 Serial #: MP4863
 34°C +/- .2 Serial #: MP4864
 34°C +/- .2 Serial #: MP5097

Calibration Adjustment By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial #	Lot #	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection By MH

Barometric Pressure ID# 68639
 Gauge 1016 Instrument 1017
 Mouth Alcohol Solution Lot # 2021-B
 Acetone Stock Solution Lot # 2020-A


Simulator	Serial Number
0.000	SD1014
Interferent	SD1015
0.050	MP4863
0.080	MP4864
0.200	MP5097

Attachments

Form 41
 Stability Checks
 Calibration Certificate
 Calibration Adjustment

Post-Stability Checks
 Flow Calibration
 Form 40
 Other _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Israel Soto Digitally signed by Israel Soto
Date: 2021.09.21 15:03:58
+0400  13:51:25
 Tech Review / Date _____ Admin Review / Date _____

Notes/Suggested Service: _____

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 12:36

Date of Inspection: 09/21/2021

Serial Number: 80-005936
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079


Standard Deviations	0.0005	0.0000	0.0006	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY
Signature and Printed Name

09/21/2021
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-005936	Sarasota County SO	09/21/2021	MK

0.05g/210L 0.047 to 0.053

0.08g/210L 0.077 to 0.083

0.20g/210L 0.194 to 0.206

DGS 0.08g/210L 0.077 to 0.083

SARASOTA COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-005936
09/21/2021
Software: 8100.27

SARASOTA COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-005936
09/21/2021
Software: 8100.27

SARASOTA COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-005936
09/21/2021
Software: 8100.27

SARASOTA COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-005936
09/21/2021
Software: 8100.27

Test g/210L Time
Air Blank 0.000 09:26
Control Test 0.049 09:27
Air Blank 0.000 09:27
Control Test 0.048 09:28
Air Blank 0.000 09:29
Control Test 0.048 09:29
Air Blank 0.000 09:30
Control Test Status
Average 0.0483
Std Dev 0.0005
Rel Std Dev(%) 1.1945

Test g/210L Time
Air Blank 0.000 09:31
Control Test 0.079 09:32
Air Blank 0.000 09:32
Control Test 0.079 09:33
Air Blank 0.000 09:34
Control Test 0.078 09:34
Air Blank 0.000 09:35
Control Test Status
Average 0.0787
Std Dev 0.0005
Rel Std Dev(%) 0.7339

Test g/210L Time
Air Blank 0.000 09:36
Control Test 0.196 09:37
Air Blank 0.000 09:38
Control Test 0.196 09:38
Air Blank 0.000 09:39
Control Test 0.196 09:40
Air Blank 0.000 09:40
Control Test Status
Average 0.1960
Std Dev 0.0000
Rel Std Dev(%) 0.0000


Test g/210L Time
Air Blank 0.000 09:42
Control Test 0.079 09:42
Air Blank 0.000 09:43
Control Test 0.080 09:43
Air Blank 0.000 09:43
Control Test 0.079 09:43
Air Blank 0.000 09:44
Control Test Status
Average 0.0793
Std Dev 0.0005
Rel Std Dev(%) 0.7277

Operator's signature *MK*

Operator's signature *MK*

Operator's signature *MK*

Operator's signature *MK*



Florida Department of Law Enforcement
 Alcohol Testing Program
 4700 Terminal Drive, Suite 1
 Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005936, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005936</u>	UNCERTAINTY* ±	
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>09/21/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:36</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3). The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/21/2021 Date

 MICHAEL D HAUGHEY,
 Department Inspector

FDLE/ATP Form 69 January 2021
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to: CMI, Inc.

Enforcement Electronics

Shipment to repair facility authorized by: Michael Feltovic on 10/19/2021

Items Returned: Instrument Supplies Other Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-005936

Bill To Address:
Sarasota County Sheriff's Office
Attn: Michael Feltovic
6010 Cattleridge Blvd
Sarasota, Florida 34232

Ship to Address:
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
Attn: Alcohol Testing Program
4700 Terminal Drive, Suite 1
Fort Myers, FL 33907

Reason for Return:

The printer port needs to be replaced.

Please choose one of the following options:

- 1. I _____, authorize all repairs.
- 2. I _____, authorize repairs up to \$ _____.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Michael Feltovic

Phone #: 941-861-4076 Email: Michael.Feltovic@sarasotasheriff.org

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us