



INSTRUMENT PROCESSING SHEET

Agency Monroe County SOS/N 80-006471

Florida Department of Law Enforcement

Date In 01/19/2021 DI Completion Date 01/19/2021 Ship P/U H/D CMI EE

Intake Performed By <u>MH</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>MH</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>205</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 106</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.480</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	--	--

Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2021.01.25 14:19:27 -05'00'	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____
--	---

Simulator	Serial #	Lot #/Exp
0.050	MP4863	202010A 10/05/2022
0.080	MP4864	202010B 10/05/2022
0.200	MP5097	202010D 10/06/2022
0.080 DGS	N/A	AG003005 01/30/2022

Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>MH</u> Barometric Pressure ID# <u>28663</u> Gauge <u>1027</u> Instrument <u>1027</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD1014																																																												
Interferent	SD1015																																																												
0.050	MP4863																																																												
0.080	MP4864																																																												
0.200	MP5097																																																												

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____
--

Notes/Suggested Service: _____ _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use David Eliezer Reyes-Rivera Digitally signed by David Eliezer Reyes-Rivera Date: 2021.01.21 10:02:54 -05'00' 2021.01.25 14:18:55 Tech Review / Date _____ Admin Review / Date _____
---	---

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY S.O.
Time of Inspection: 11:29

Date of Inspection: 01/19/2021

Serial Number: 80-006471
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.079	0.199	0.078


Standard Deviations	0.0004	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




MICHAEL D HAUGHEY

Signature and Printed Name





01/19/2021
Date

DERR Digitally signed
by DERR
Date: 2021.01.21
10:02:12 -05'00'

 2021.01.
25
14:17:47
-05'00'

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006471	Monroe County SO	01/19/2021	MLK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input type="checkbox"/>	0.077 to 0.083 <input type="checkbox"/>	0.194 to 0.206 <input type="checkbox"/>	0.077 to 0.083 <input type="checkbox"/>

<p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p>	<p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p>	<p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p>	<p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p> <p>MONROE COUNTY S.O. Intoxilyzer - Alconol Analyzer Model 8000 01/19/2021 Software: 8100.27</p>																																																																																																																																																
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:57</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:58</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:59</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:00</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:57	Control Test	0.048	08:57	Air Blank	0.000	08:58	Control Test	0.048	08:59	Air Blank	0.000	08:59	Control Test	0.049	09:00	Air Blank	0.000	09:00	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:01</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:02</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:04</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:01	Control Test	0.078	09:02	Air Blank	0.000	09:02	Control Test	0.078	09:03	Air Blank	0.000	09:04	Control Test	0.078	09:04	Air Blank	0.000	09:05	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:08</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:09</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:07	Control Test	0.198	09:07	Air Blank	0.000	09:08	Control Test	0.198	09:09	Air Blank	0.000	09:09	Control Test	0.198	09:10	Air Blank	0.000	09:10	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:13</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:14</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:14</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:15</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:13	Control Test	0.078	09:13	Air Blank	0.000	09:14	Control Test	0.078	09:14	Air Blank	0.000	09:14	Control Test	0.078	09:15	Air Blank	0.000	09:15	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:57																																																																																																																																																	
Control Test	0.048	08:57																																																																																																																																																	
Air Blank	0.000	08:58																																																																																																																																																	
Control Test	0.048	08:59																																																																																																																																																	
Air Blank	0.000	08:59																																																																																																																																																	
Control Test	0.049	09:00																																																																																																																																																	
Air Blank	0.000	09:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:01																																																																																																																																																	
Control Test	0.078	09:02																																																																																																																																																	
Air Blank	0.000	09:02																																																																																																																																																	
Control Test	0.078	09:03																																																																																																																																																	
Air Blank	0.000	09:04																																																																																																																																																	
Control Test	0.078	09:04																																																																																																																																																	
Air Blank	0.000	09:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:07																																																																																																																																																	
Control Test	0.198	09:07																																																																																																																																																	
Air Blank	0.000	09:08																																																																																																																																																	
Control Test	0.198	09:09																																																																																																																																																	
Air Blank	0.000	09:09																																																																																																																																																	
Control Test	0.198	09:10																																																																																																																																																	
Air Blank	0.000	09:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:13																																																																																																																																																	
Control Test	0.078	09:13																																																																																																																																																	
Air Blank	0.000	09:14																																																																																																																																																	
Control Test	0.078	09:14																																																																																																																																																	
Air Blank	0.000	09:14																																																																																																																																																	
Control Test	0.078	09:15																																																																																																																																																	
Air Blank	0.000	09:15																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Operator's Signature: 	Operator's Signature: 	Operator's Signature: 	Operator's Signature: 																																																																																																																																																



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006471, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006471</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MONROE COUNTY S.O.</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>01/19/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:29</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/19/2021

Date

MICHAEL D HAUGHEY,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

DERR
Digitally signed
by DERR
Date: 2021.01.21
100100.4500

2021.01.
25
14:17:08
-05'00'