



INSTRUMENT PROCESSING SHEET

Agency Manatee County SOS/N 80-006565Florida Department of
Law EnforcementDate In 3/19/2021DI Completion Date 3/24/2021☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By MH	Quality Checks	By MH	Date <u>03/24/2021</u>	Flow Calibration	By	Date																											
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>207</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP4863</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP4864</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP5096</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG026705 09/23/2022</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP4863	202010A 10/05/2022	0.080	MP4864	202010B 10/05/2022	0.200	MP5096	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____														
Simulator	Serial #	Lot #/Exp																																
0.050	MP4863	202010A 10/05/2022																																
0.080	MP4864	202010B 10/05/2022																																
0.200	MP5096	202010D 10/06/2022																																
0.080 DGS	N/A	AG026705 09/23/2022																																
		DI Temp. Checks By MH _____ ☒ Lab Temp °C <u>22.08</u> External Digital Therm. ID#: <u>300504</u> ☒ 34°C +/- .2 Serial #: <u>MP4863</u> ☒ 34°C +/- .2 Serial #: <u>MP4864</u> ☒ 34°C +/- .2 Serial #: <u>MP5096</u>																																
Calibration Adjustment By _____		Department Inspection By MH _____																																
Barometric Pressure Gauge ID # _____		Barometric Pressure ID# <u>68639</u>																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1016</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u>				
Simulator	Serial #	Lot #	Expiration																															
0.000		N/A	N/A																															
0.040																																		
0.100																																		
0.200																																		
0.300																																		
0.080 DGS	N/A																																	
☐ Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>SD1014</td></tr><tr><td>Interferent</td><td>SD1015</td></tr><tr><td>0.050</td><td>MP4863</td></tr><tr><td>0.080</td><td>MP4864</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>					Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5096																
Simulator	Serial Number																																	
0.000	SD1014																																	
Interferent	SD1015																																	
0.050	MP4863																																	
0.080	MP4864																																	
0.200	MP5096																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____												
Simulator	Serial #	Lot #	Expiration																															
0.050																																		
0.080																																		
0.200																																		
0.080 DGS	N/A																																	
Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2021.03.																																
		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2021.03.26 13:33:42 -0400 31 14:15:22 Tech Review / Date Admin Review / Date																																

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 12:28

Date of Inspection: 03/24/2021

Serial Number: 80-006565
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.051	0.082	0.201	0.081
0.000	0.051	0.081	0.201	0.081
0.000	0.051	0.081	0.201	0.081
0.000	0.051	0.081	0.201	0.080
0.000	0.051	0.081	0.201	0.081
0.000	0.051	0.081	0.201	0.081
0.000	0.051	0.080	0.201	0.080
0.000	0.051	0.081	0.201	0.081
0.000	0.051	0.081	0.201	0.080
0.000	0.051	0.081	0.201	0.080


Standard Deviations	0.0000	0.0004	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

03/24/2021
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006565	Manatee County SO	03/24/21	MX

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒

MANATEE COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-006565
03/24/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.051	10:06
Air Blank	0.000	10:06
Control Test	0.051	10:07
Air Blank	0.000	10:07
Control Test	0.051	10:08
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MX

Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-006565
03/24/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.081	10:11
Air Blank	0.000	10:12
Control Test	0.081	10:13
Air Blank	0.000	10:13
Control Test	0.081	10:14
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MX

Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-006565
03/24/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:16
Control Test	0.202	10:17
Air Blank	0.000	10:17
Control Test	0.202	10:18
Air Blank	0.000	10:19
Control Test	0.202	10:19
Air Blank	0.000	10:20
Control Test Stats		
Average	0.2020	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MX


Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-006565
03/24/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:22
Control Test	0.081	10:22
Air Blank	0.000	10:22
Control Test	0.080	10:23
Air Blank	0.000	10:23
Control Test	0.081	10:24
Air Blank	0.000	10:24
Control Test	0.081	10:24
Air Blank	0.000	10:25
Control Test Stats		
Average	0.0808	
Std Dev	0.0005	
Rel Std Dev(%)	0.6192	

MX

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006565, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006565</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MANATEE COUNTY SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>03/24/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:28</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/24/2021

Date


MICHAEL D HAUGHEY,
Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1