



# INSTRUMENT PROCESSING SHEET

Agency Holmes Beach PDS/N 80-006692Florida Department of  
Law EnforcementDate In 10/22/2021 DI Completion Date 11/8/2021 Ship  P/U  H/D  CMI  EE

Intake	By TDG	Quality Checks	By TDG	Date 11/8/2021	Flow Calibration	By	Date																																																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>177</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks				Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
					<b>Maintenance</b> By _____																																																														
					<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																																																														
					<b>DI Temp. Checks</b> By TDG _____																																																														
					<input checked="" type="checkbox"/> Lab Temp °C <u>20.36</u> External Digital Therm. ID#: <u>381198</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5092</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5094</u>																																																														
<b>Calibration Adjustment</b>				<b>Department Inspection</b>																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28199</u> Gauge <u>1021</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>SD3963</td></tr> <tr><td>Interferent</td><td>SD1017</td></tr> <tr><td>0.050</td><td>MP5092</td></tr> <tr><td>0.080</td><td>MP4864</td></tr> <tr><td>0.200</td><td>MP5094</td></tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____				Simulator	Serial Number	0.000	SD3963	Interferent	SD1017	0.050	MP5092	0.080	MP4864	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	SD3963																																																																		
Interferent	SD1017																																																																		
0.050	MP5092																																																																		
0.080	MP4864																																																																		
0.200	MP5094																																																																		
				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																															
				Israel Soto <small>Digitally signed by Israel Soto Date: 2021.11.10 11:48:52 +05'00'</small> Tech Review / Date _____ Admin Review / Date <u>2021.11.10 14:23:42</u>																																																															

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 6692	Holmes Beach PD	11/08/2021	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083

<p>HOLMES BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 11/08/2021 SN 80-006692 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:33</p> <p>Control Test 0.049 10:34</p> <p>Air Blank 0.000 10:35</p> <p>Control Test 0.050 10:35</p> <p>Air Blank 0.000 10:36</p> <p>Control Test 0.050 10:36</p> <p>Air Blank 0.000 10:37</p> <p>Control Test Stats</p> <p>Average 0.0497</p> <p>Std Dev 0.0006</p> <p>Rel. Std Dev(%) 1.1625</p>	<p>HOLMES BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 11/08/2021 SN 80-006692 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:42</p> <p>Control Test 0.079 10:42</p> <p>Air Blank 0.000 10:43</p> <p>Control Test 0.080 10:43</p> <p>Air Blank 0.000 10:44</p> <p>Control Test 0.080 10:44</p> <p>Air Blank 0.000 10:45</p> <p>Control Test Stats</p> <p>Average 0.0797</p> <p>Std Dev 0.0006</p> <p>Rel. Std Dev(%) 0.7247</p>	<p>HOLMES BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 11/08/2021 SN 80-006692 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:50</p> <p>Control Test 0.199 10:51</p> <p>Air Blank 0.000 10:51</p> <p>Control Test 0.199 10:52</p> <p>Air Blank 0.000 10:53</p> <p>Control Test 0.198 10:53</p> <p>Air Blank 0.000 10:54</p> <p>Control Test Stats</p> <p>Average 0.1987</p> <p>Std Dev 0.0006</p> <p>Rel. Std Dev(%) 0.2906</p>	<p>HOLMES BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 11/08/2021 SN 80-006692 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:56</p> <p>Control Test 0.078 10:57</p> <p>Air Blank 0.000 10:57</p> <p>Control Test 0.078 10:58</p> <p>Air Blank 0.000 10:58</p> <p>Control Test 0.078 10:58</p> <p>Air Blank 0.000 10:59</p> <p>Control Test Stats</p> <p>Average 0.0780</p> <p>Std Dev 0.0000</p> <p>Rel. Std Dev(%) 0.0000</p>
Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>

*DGS*

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES BEACH PD  
Time of Inspection: 13:20

Date of Inspection: 11/08/2021

Serial Number: 80-006692  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.078	0.198	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.077

Standard Deviations	0.0000	0.0003	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/08/2021  
Date



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 4700 Terminal Drive, Suite 1  
 Ft. Myers, FL 33907

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006692, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006692</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HOLMES BEACH PD</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>11/08/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:20</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
 \*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).  
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.  
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/08/2021

Date

TAYLOR D GUTSCHOW,  
 Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality