



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006924, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006924</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>05/27/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:17</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/27/2021

Date

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 12:17

Date of Inspection: 05/27/2021

Serial Number: 80-006924
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.047	0.078	0.195	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.049	0.079	0.197	0.079
0.000	0.048	0.080	0.197	0.080
0.000	0.048	0.079	0.197	0.079


Standard Deviations	0.0006	0.0004	0.0009	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

DAVID E REYES-RIVERA

05/27/2021
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-006924	Broward County Sheriff's Office	5/27/2021	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 05/27/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr><tr><td>Control Test</td><td>0.048</td><td>10:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr><tr><td>Control Test</td><td>0.050</td><td>10:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>2.0408</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	10:20	Control Test	0.048	10:21	Air Blank	0.000	10:21	Control Test	0.049	10:22	Air Blank	0.000	10:22	Control Test	0.050	10:23	Air Blank	0.000	10:24	Control Test Stats			Average	0.0490		Std Dev	0.0010		Rel Std Dev(%)	2.0408		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 05/27/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	10:25	Control Test	0.078	10:25	Air Blank	0.000	10:26	Control Test	0.078	10:27	Air Blank	0.000	10:27	Control Test	0.078	10:28	Air Blank	0.000	10:28	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 05/27/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr><tr><td>Control Test</td><td>0.195</td><td>10:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr><tr><td>Control Test</td><td>0.197</td><td>10:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1967</td><td></td></tr><tr><td>Std Dev</td><td>0.0015</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7767</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	10:30	Control Test	0.195	10:31	Air Blank	0.000	10:32	Control Test	0.198	10:32	Air Blank	0.000	10:33	Control Test	0.197	10:34	Air Blank	0.000	10:34	Control Test Stats			Average	0.1967		Std Dev	0.0015		Rel Std Dev(%)	0.7767		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 05/27/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	10:35	Control Test	0.079	10:35	Air Blank	0.000	10:36	Control Test	0.078	10:36	Air Blank	0.000	10:37	Control Test	0.079	10:37	Air Blank	0.000	10:38	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:20																																																																																																																																																	
Control Test	0.048	10:21																																																																																																																																																	
Air Blank	0.000	10:21																																																																																																																																																	
Control Test	0.049	10:22																																																																																																																																																	
Air Blank	0.000	10:22																																																																																																																																																	
Control Test	0.050	10:23																																																																																																																																																	
Air Blank	0.000	10:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	2.0408																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:25																																																																																																																																																	
Control Test	0.078	10:25																																																																																																																																																	
Air Blank	0.000	10:26																																																																																																																																																	
Control Test	0.078	10:27																																																																																																																																																	
Air Blank	0.000	10:27																																																																																																																																																	
Control Test	0.078	10:28																																																																																																																																																	
Air Blank	0.000	10:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:30																																																																																																																																																	
Control Test	0.195	10:31																																																																																																																																																	
Air Blank	0.000	10:32																																																																																																																																																	
Control Test	0.198	10:32																																																																																																																																																	
Air Blank	0.000	10:33																																																																																																																																																	
Control Test	0.197	10:34																																																																																																																																																	
Air Blank	0.000	10:34																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1967																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	0.7767																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:35																																																																																																																																																	
Control Test	0.079	10:35																																																																																																																																																	
Air Blank	0.000	10:36																																																																																																																																																	
Control Test	0.078	10:36																																																																																																																																																	
Air Blank	0.000	10:37																																																																																																																																																	
Control Test	0.079	10:37																																																																																																																																																	
Air Blank	0.000	10:38																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
 Operator's Signature	 Operator's Signature	 Operator's Signature	 Operator's Signature																																																																																																																																																

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4340 (0.0150)
Sample #2 = 1.4530 (0.0100)
Sample #3 = 1.4850 (-0.0040)
Sample #4 = 1.4160 (0.0180)
Avg % Abs = 1.4513 (0.0080)
STD DEV = 0.0345 (0.0111)
REL STD DEV = 2.379 (139.194)

Model 8000
05/27/2021
09:25:35

BROWARD COUNTY SO

Intoxilizer - Alcohol Analyzer

Auto Calibration

Max Power Res Value = 82

Auto Range Res Value = 55

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12599, Sum Io = 12783

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0480 (0.0010)
Sample #2 = 0.1170 (0.0040)
Sample #3 = 0.0250 (0.0040)
Sample #4 = 0.0760 (0.0050)
Avg % Abs = 0.0727 (0.0423)
STD DEV = 0.0461 (0.0333)
REL STD DEV = 63.427 (78.642)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0790 (0.0240)
Sample #2 = 0.1380 (0.0180)
Sample #3 = 0.1250 (0.0230)
Sample #4 = 0.1290 (0.0160)
Avg % Abs = 0.1307 (0.0190)
STD DEV = 0.0067 (0.0036)
REL STD DEV = 5.096 (18.977)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12576, Sum Io = 12778

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7550 (0.0020)
Sample #2 = 0.7800 (0.0230)
Sample #3 = 0.8090 (0.0070)
Sample #4 = 0.7360 (0.0530)
Avg % Abs = 0.7750 (0.0277)
STD DEV = 0.0368 (0.0234)
REL STD DEV = 4.743 (84.406)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.4210 (0.0000)
Sample #2 = 3.3760 (0.0240)
Sample #3 = 3.4440 (0.0010)
Sample #4 = 3.3880 (0.0250)
Avg % Abs = 3.4027 (0.0167)
STD DEV = 0.0363 (0.0136)
REL STD DEV = 1.067 (81.462)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12561, Sum Io = 12768

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5830 (-0.0230)
Sample #2 = 3.5740 (0.0000)
Sample #3 = 3.5710 (0.0110)
Sample #4 = 3.5430 (0.0390)
Avg % Abs = 3.5827 (0.0167)
STD DEV = 0.0171 (0.0201)
REL STD DEV = 0.480 (120.648)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.5220 (-0.0250)
Sample #2 = 6.4750 (0.0040)
Sample #3 = 6.4870 (0.0050)
Sample #4 = 6.4940 (0.0240)
Avg % Abs = 6.4853 (0.0110)
STD DEV = 0.0096 (0.0113)
REL STD DEV = 0.148 (102.449)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12565, Sum Io = 12772

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.6610 (-0.0100)
Sample #2 = 1.8030 (0.0190)
Sample #3 = 1.8520 (-0.0060)
Sample #4 = 1.8000 (0.0350)
Avg % Abs = 1.8183 (0.0160)
STD DEV = 0.0252 (0.0207)
REL STD DEV = 1.606 (129.150)

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12555, Sum Io = 12766

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.145.00
Sample #2 = 3.190.00
Sample #3 = 3.193.00
Sample #4 = 3.125.00
Average Result = 3.169.3333
STD DEV = 38.4231
REL STD DEV = 1.212

Channel 2 >>>>
Sample #1 = 3267.00
Sample #2 = 3278.00
Sample #3 = 3301.00
Sample #4 = 3240.00
Average Result = 3273.0000
STD DEV = 30.8058
REL STD DEV = 0.941

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.3430 (0.0090)
Sample #2 = 9.3850 (0.0010)
Sample #3 = 9.3160 (0.0150)
Sample #4 = 9.3520 (0.0160)
Avg % Abs = 9.3510 (0.0107)
STD DEV = 0.0395 (0.0094)
REL STD DEV = 0.369 (78.623)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.3430 (0.0090)
Sample #2 = 9.3850 (0.0010)
Sample #3 = 9.3160 (0.0150)
Sample #4 = 9.3520 (0.0160)
Avg % Abs = 9.3510 (0.0107)
STD DEV = 0.0395 (0.0094)
REL STD DEV = 0.369 (78.623)

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.3430 (0.0090)
Sample #2 = 9.3850 (0.0010)
Sample #3 = 9.3160 (0.0150)
Sample #4 = 9.3520 (0.0160)
Avg % Abs = 9.3510 (0.0107)
STD DEV = 0.0395 (0.0094)
REL STD DEV = 0.369 (78.623)

Sol Value = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.073
Std Dev = 0.05 Rel Std Dev = 63.43
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.778
Std Dev = 0.01 Rel Std Dev = 1.19
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.818
Std Dev = 0.03 Rel Std Dev = 1.61
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.563
Std Dev = 0.02 Rel Std Dev = 0.48
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.145
Std Dev = 0.01 Rel Std Dev = 0.15
Zero Order Coef = -156.86
First Order Coef = 2597.43
Second Order Coef = 39.67
Standard Deviation = 55.118656

Optical Calibration

SN: 80-006924
Agency: Broward County SO
Date: 5/27/2021
Quadratic Fit: +/- 0.002g/210L
By: DERR



Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006924	Broward County Sheriff's Office	5/27/2021	DERR <i>MD</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/27/2021 Software: 8100.27 SN 80-006924 Test g/210L Time Air Blank 0.000 08:23 Control Test 0.046 08:24 Air Blank 0.000 08:25 Control Test 0.048 08:25 Air Blank 0.000 08:26 Control Test 0.046 08:27 Air Blank 0.000 08:27 Control Test Stats Average 0.0467 Std Dev 0.0012 Rel Std Dev(%) 2.4744	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/27/2021 Software: 8100.27 SN 80-006924 Test g/210L Time Air Blank 0.000 08:28 Control Test 0.077 08:29 Air Blank 0.000 08:30 Control Test 0.076 08:30 Air Blank 0.000 08:31 Control Test 0.077 08:32 Air Blank 0.000 08:32 Control Test Stats Average 0.0767 Std Dev 0.0006 Rel Std Dev(%) 0.7531	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/27/2021 Software: 8100.27 SN 80-006924 Test g/210L Time Air Blank 0.000 08:33 Control Test 0.193 08:34 Air Blank 0.000 08:34 Control Test 0.192 08:35 Air Blank 0.000 08:36 Control Test 0.193 08:36 Air Blank 0.000 08:37 Control Test Stats Average 0.1927 Std Dev 0.0006 Rel Std Dev(%) 0.2997	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/27/2021 Software: 8100.27 SN 80-006924 Test g/210L Time Air Blank 0.000 08:38 Control Test 0.077 08:38 Air Blank 0.000 08:39 Control Test 0.078 08:39 Air Blank 0.000 08:40 Control Test 0.077 08:40 Air Blank 0.000 08:41 Control Test Stats Average 0.0773 Std Dev 0.0006 Rel Std Dev(%) 0.7466
<i>MD</i> Operator's Signature	<i>MD</i> Operator's Signature	<i>MD</i> Operator's Signature	<i>MD</i> Operator's Signature

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006924
05/27/2021
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
 SQRT(Diff)) = 5.656
2: Rate (Liters/min) = 15
 SQRT(Diff)) = 10.582
3: Rate (Liters/min) = 30
 SQRT(Diff)) = 20.344
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 255
Rounded Slope = 656
Rounded Intercept = -390491
Correlation = 0.99734

Flow Calibration	
SN:	80-0006924
Agency:	Broward County Sheriff's Office
Date:	5/27/2021
By:	DERR <i>[Signature]</i>



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
1700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006924, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006924</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>05/27/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:17</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

05/27/2021

Date

DAVID E REYES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 12:17

Date of Inspection: 05/27/2021

Serial Number: 80-006924
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:76026705 Exp: 09/23/2022
0.000	0.047	0.078	0.195	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.049	0.079	0.197	0.079
0.000	0.048	0.080	0.197	0.080
0.000	0.048	0.079	0.197	0.079


Standard Deviations	0.0006	0.0004	0.0009	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that, I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVER

Signature and Printed Name

05/27/2021
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-006924	Broward County Sheriff's Office	5/27/2021	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 05/27/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr><tr><td>Control Test</td><td>0.048</td><td>10:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr><tr><td>Control Test</td><td>0.050</td><td>10:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>2.0408</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	10:20	Control Test	0.048	10:21	Air Blank	0.000	10:21	Control Test	0.049	10:22	Air Blank	0.000	10:22	Control Test	0.050	10:23	Air Blank	0.000	10:24	Control Test Stats			Average	0.0490		Std Dev	0.0010		Rel Std Dev(%)	2.0408		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 05/27/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	10:25	Control Test	0.078	10:25	Air Blank	0.000	10:26	Control Test	0.078	10:27	Air Blank	0.000	10:27	Control Test	0.078	10:28	Air Blank	0.000	10:28	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 05/27/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr><tr><td>Control Test</td><td>0.195</td><td>10:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr><tr><td>Control Test</td><td>0.197</td><td>10:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1967</td><td></td></tr><tr><td>Std Dev</td><td>0.0015</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7767</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	10:30	Control Test	0.195	10:31	Air Blank	0.000	10:32	Control Test	0.198	10:32	Air Blank	0.000	10:33	Control Test	0.197	10:34	Air Blank	0.000	10:34	Control Test Stats			Average	0.1967		Std Dev	0.0015		Rel Std Dev(%)	0.7767		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 05/27/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	10:35	Control Test	0.079	10:35	Air Blank	0.000	10:36	Control Test	0.078	10:36	Air Blank	0.000	10:37	Control Test	0.079	10:37	Air Blank	0.000	10:38	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:20																																																																																																																																																	
Control Test	0.048	10:21																																																																																																																																																	
Air Blank	0.000	10:21																																																																																																																																																	
Control Test	0.049	10:22																																																																																																																																																	
Air Blank	0.000	10:22																																																																																																																																																	
Control Test	0.050	10:23																																																																																																																																																	
Air Blank	0.000	10:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	2.0408																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:25																																																																																																																																																	
Control Test	0.078	10:25																																																																																																																																																	
Air Blank	0.000	10:26																																																																																																																																																	
Control Test	0.078	10:27																																																																																																																																																	
Air Blank	0.000	10:27																																																																																																																																																	
Control Test	0.078	10:28																																																																																																																																																	
Air Blank	0.000	10:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:30																																																																																																																																																	
Control Test	0.195	10:31																																																																																																																																																	
Air Blank	0.000	10:32																																																																																																																																																	
Control Test	0.198	10:32																																																																																																																																																	
Air Blank	0.000	10:33																																																																																																																																																	
Control Test	0.197	10:34																																																																																																																																																	
Air Blank	0.000	10:34																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1967																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	0.7767																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:35																																																																																																																																																	
Control Test	0.079	10:35																																																																																																																																																	
Air Blank	0.000	10:36																																																																																																																																																	
Control Test	0.078	10:36																																																																																																																																																	
Air Blank	0.000	10:37																																																																																																																																																	
Control Test	0.079	10:37																																																																																																																																																	
Air Blank	0.000	10:38																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
 Operator's Signature	 Operator's Signature	 Operator's Signature	 Operator's Signature																																																																																																																																																

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.131

Std Dev = 0.01 Rel Std Dev = 5.10

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.464

Std Dev = 0.02 Rel Std Dev = 1.30

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.403

Std Dev = 0.04 Rel Std Dev = 1.07

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.485

Std Dev = 0.01 Rel Std Dev = 0.15

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 9.351

Std Dev = 0.03 Rel Std Dev = 0.37

Zero Order Coef = -173.18

First Order Coef = 1390.14

Second Order Coef = 16.61

Standard Deviation = 12.578669

Solution Status Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0007
0.040	0.040	0.0003
0.100	0.099	0.0014
0.200	0.202	-0.0016
0.300	0.299	0.0006

Solution Status Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	0.0002
0.100	0.100	0.0003
0.200	0.200	-0.0004
0.300	0.300	0.0002

Optical Calibration	
SN:	80-006924
Agency: Broward County SO	
Date:	5/27/2021
Quadratic Fit: +/- 0.002g/210L	
By:	DERR <i>[Signature]</i>

09-25-35
05/27/2021
BROWARD COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
SN 80-006924
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

Auto Calibration
Max Power Res Value = 82
Auto Range Res Value = 55

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l ****
Samples Taken = 4, Discarded = 1
Sum Io = 12589, 9um Io = 12783

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

Optical Calibration	
SN:	80-006924
Agency: Broward County SO	
Date:	5/27/2021
Quadratic Fit: +/- 0.002g/210L	
By:	DERR

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

05/27/2021
09-25-35
05/27/2021
09-25-35

Standard Deviation = 55.118556

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006924
05/27/2021
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff)) = 5.656
2: Rate (Liters/min) = 15
SQRT(Diff)) = 10.582
3: Rate (Liters/min) = 30
SQRT(Diff)) = 20.344
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 656
Rounded Intercept = -380491
Correlation = 0.99734

Flow Calibration	
SN:	80-0006924
Agency: Broward County Sheriff's Office	
Date:	5/27/2021
By:	DERR <i>del</i>