



INSTRUMENT PROCESSING SHEET

Agency St. John's County SOS/N 80-007366Florida Department of
Law EnforcementDate In 01-08-2021 DI Completion Date 01-11-2021☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>IS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2021.01.12 23:40:05 -05'00'	Quality Checks Performed By <u>IS</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>184</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A 10-05-2022</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> <td>202010B 10-05-2022</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D 10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102 04-20-2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202010A 10-05-2022	0.080	MP5086	202010B 10-05-2022	0.200	MP5090	202010D 10-06-2022	0.080 DGS	N/A	AG011102 04-20-2022	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>IS</u> ☒ Lab Temp °C <u>21.49</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C +/- .2 Serial #: <u>MP5088</u> ☒ 34°C +/- .2 Serial #: <u>MP5086</u> ☒ 34°C +/- .2 Serial #: <u>MP5090</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP5088	202010A 10-05-2022																																																											
0.080	MP5086	202010B 10-05-2022																																																											
0.200	MP5090	202010D 10-06-2022																																																											
0.080 DGS	N/A	AG011102 04-20-2022																																																											
Calibration Adjustment Performed By <u>IS</u> Barometric Pressure Gauge <u>1023</u> ID # <u>28421</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>20060</td> <td>02-10-2022</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20190</td> <td>04-06-2022</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>03-18-2022</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>01-21-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>06-05-2021</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A</td> <td>10-05-2022</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> <td>202010B</td> <td>10-05-2022</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D</td> <td>10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102</td> <td>04-20-2022</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	02-10-2022	0.100	MP5083	20190	04-06-2022	0.200	MP5084	20160	03-18-2022	0.300	MP5085	20030	01-21-2022	0.080 DGS	N/A	08819080A1	06-05-2021	Simulator	Serial Number	Lot Number	Expiration	0.050	MP5088	202010A	10-05-2022	0.080	MP5086	202010B	10-05-2022	0.200	MP5090	202010D	10-06-2022	0.080 DGS	N/A	AG011102	04-20-2022	Department Inspection Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1021</u> Instrument <u>1023</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G11621</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____	Simulator	Serial Number	0.000	G11621	Interferent	MP5087	0.050	MP5088	0.080	MP5086	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	MP5091	N/A	N/A																																																										
0.040	MP5082	20060	02-10-2022																																																										
0.100	MP5083	20190	04-06-2022																																																										
0.200	MP5084	20160	03-18-2022																																																										
0.300	MP5085	20030	01-21-2022																																																										
0.080 DGS	N/A	08819080A1	06-05-2021																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	MP5088	202010A	10-05-2022																																																										
0.080	MP5086	202010B	10-05-2022																																																										
0.200	MP5090	202010D	10-06-2022																																																										
0.080 DGS	N/A	AG011102	04-20-2022																																																										
Simulator	Serial Number																																																												
0.000	G11621																																																												
Interferent	MP5087																																																												
0.050	MP5088																																																												
0.080	MP5086																																																												
0.200	MP5090																																																												
Notes/Suggested Service: <u>0.05 stability check was not nominal, checked simulator lid and reran stability, value still not nominal, performed optical bench cal adjust.</u> _____ _____ _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use David Eliezer Reyes-Rivera Digitally signed by David Eliezer Reyes-Rivera Date: 2021.01.11 13:02:09 -05'00' Tech Review / Date _____ Admin Review / Date <u>2021.01.12 23:39:21</u>																																																											

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST JOHNS SO
Time of Inspection: 12:16

Date of Inspection: 01/11/2021

Serial Number: 80-007366
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
Standard Deviations	0.0004	0.0000	0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

01/11/2021
Date

DERR
Digitally signed
by DERR
Date: 2021.01.11
13:01:33 -05'00'

Stability Checks

ST JOHNS SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
01/11/2021
Software: 8100.27

ST JOHNS SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
01/11/2021
Software: 8100.27

ST JOHNS SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
01/11/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.047	08:36
Air Blank	0.000	08:37
Control Test	0.046	08:38
Air Blank	0.000	08:38
Control Test	0.046	08:39
Air Blank	0.000	08:40
Control Test Stats		
Average	0.0463	
Std Dev	0.0006	
Rel Std Dev(%)	1.2461	

Test	g/210L	Time
Air Blank	0.000	08:41
Control Test	0.046	08:41
Air Blank	0.000	08:42
Control Test	0.046	08:42
Air Blank	0.000	08:43
Control Test	0.046	08:44
Air Blank	0.000	08:44
Control Test Stats		
Average	0.0460	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	08:46
Control Test	0.077	08:46
Air Blank	0.000	08:47
Control Test	0.077	08:48
Air Blank	0.000	08:48
Control Test	0.077	08:49
Air Blank	0.000	08:49
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet


Operator's Signature


Operator's Signature


Operator's Signature

ST JOHNS SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
01/11/2021
Software: 8100.27

ST JOHNS SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
01/11/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.198	08:51
Air Blank	0.000	08:52
Control Test	0.197	08:53
Air Blank	0.000	08:53
Control Test	0.197	08:54
Air Blank	0.000	08:54
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

Test	g/210L	Time
Air Blank	0.000	08:55
Control Test	0.078	08:56
Air Blank	0.000	08:56
Control Test	0.078	08:56
Air Blank	0.000	08:57
Control Test	0.078	08:57
Air Blank	0.000	08:58
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Dry


Operator's Signature


Operator's Signature

DERR
Digitally signed
by DERR
Date: 2021.01.11
13:01:00 -05'00'

2021.01.11
2
23:37:58
-05'00'



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007366, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-007366	UNCERTAINTY* $\pm$
Owning Agency:	ST JOHNS SO	0.050 g/ 210 L 0.004
Calibration Date:	01/11/2021	0.080 g/ 210 L 0.005
Calibration Time:	12:16	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/11/2021

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Digitally signed
by DERR
Date: 2021.01.11
13:00:21 -05'00'

2021.01.
12
23:37:21
-05'00'

Opt. Bench Cal. Adj.

SA

ST JOHNS SD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
01/11/2021 09:30:35

Auto Calibration
Max Power Res Value = 86
Auto Range Res Value = 58

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12773, 9um Io = 13277

CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.1290	(-0.0040)
Sample #2	0.1480	(0.0470)
Sample #3	0.0980	(0.1050)
Sample #4	0.1110	(0.1310)
Avg % Abs	0.1190	(0.0943)
STD DEV	0.0259	(0.0430)
REL STD DEV	21.800	(45.587)

CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.1870	(-0.0100)
Sample #2	0.2100	(0.0000)
Sample #3	0.1870	(0.0160)
Sample #4	0.1840	(0.0340)
Avg % Abs	0.1937	(0.0167)
STD DEV	0.0142	(0.0170)
REL STD DEV	7.345	(102.059)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12752, 9um Io = 13274

CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.8140	(-0.0220)
Sample #2	0.8180	(0.0210)
Sample #3	0.8280	(0.0340)
Sample #4	0.8030	(0.0450)
Avg % Abs	0.8163	(0.0333)
STD DEV	0.0126	(0.0120)
REL STD DEV	1.541	(36.042)

CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.5300	(-0.0070)
Sample #2	1.4940	(0.0420)
Sample #3	1.5230	(0.0430)
Sample #4	1.5020	(0.0390)
Avg % Abs	1.5063	(0.0413)
STD DEV	0.0150	(0.0021)
REL STD DEV	0.994	(5.036)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12741, 9um Io = 13269

CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.8270	(-0.0140)
Sample #2	1.8360	(0.0160)
Sample #3	1.8000	(0.0380)
Sample #4	1.8330	(0.0340)
Avg % Abs	1.8230	(0.0293)
STD DEV	0.0200	(0.0117)
REL STD DEV	1.096	(39.951)

CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.4490	(-0.0010)
Sample #2	3.4340	(0.0320)
Sample #3	3.4010	(0.0570)
Sample #4	3.4240	(0.0470)
Avg % Abs	3.4197	(0.0453)
STD DEV	0.0169	(0.0126)
REL STD DEV	0.495	(27.757)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12735, 9um Io = 13265

CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.4910	(-0.0050)
Sample #2	3.4600	(0.0360)
Sample #3	3.4870	(0.0320)
Sample #4	3.4540	(0.0630)
Avg % Abs	3.4670	(0.0437)
STD DEV	0.0176	(0.0169)
REL STD DEV	0.507	(38.616)

CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	6.5230	(-0.0100)
Sample #2	6.4580	(0.0470)
Sample #3	6.4690	(0.0590)
Sample #4	6.4490	(0.0790)
Avg % Abs	6.4587	(0.0617)
STD DEV	0.0100	(0.0162)
REL STD DEV	0.155	(26.215)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12729, 9um Io = 13263

CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	5.0980	(-0.0160)
Sample #2	5.0600	(0.0380)
Sample #3	5.0090	(0.0660)
Sample #4	5.0160	(0.0590)
Avg % Abs	5.0283	(0.0543)
STD DEV	0.0276	(0.0146)
REL STD DEV	0.550	(26.819)

CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	9.4170	(-0.0130)
Sample #2	9.3160	(0.0930)
Sample #3	9.2920	(0.1170)
Sample #4	9.2740	(0.1220)
Avg % Abs	9.2940	(0.1107)
STD DEV	0.0211	(0.0155)
REL STD DEV	0.227	(14.008)

***** AUTO CAL DATA *****

CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.119
Std Dev = 0.03 Rel Std Dev = 21.80
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.816
Std Dev = 0.01 Rel Std Dev = 1.54
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.823
Std Dev = 0.02 Rel Std Dev = 1.10
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.467
Std Dev = 0.02 Rel Std Dev = 0.51
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.028
Std Dev = 0.03 Rel Std Dev = 0.55
Zero Order Coef = -323.75
First Order Coef = 2709.56
Second Order Coef = 38.79
Standard Deviation = 11.850574

CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.194
Std Dev = 0.01 Rel Std Dev = 7.34
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.506
Std Dev = 0.01 Rel Std Dev = 0.99
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.420
Std Dev = 0.02 Rel Std Dev = 0.49
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.459
Std Dev = 0.01 Rel Std Dev = 0.16
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.294
Std Dev = 0.02 Rel Std Dev = 0.23
Zero Order Coef = -265.58
First Order Coef = 1410.59
Second Order Coef = 16.62
Standard Deviation = 10.689440

2021.01.
12
23:36:34
-05'00'

DERR

Digitally signed
by DERR
Date: 2021.01.11
12:59:42 -05'00'

opt. Bench Cal. Adj.

Post Stability Checks

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0000
0.040	0.040	-0.0002
0.100	0.100	0.0004
0.200	0.200	-0.0003
0.300	0.300	0.0001

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	0.0002
0.100	0.100	0.0002
0.200	0.200	-0.0003
0.300	0.300	0.0001

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3345.00
Sample #2 = 3393.00
Sample #3 = 3375.00
Sample #4 = 3382.00
Average Result = 3383.3333
STD DEV = 9.0738
REL STD DEV = 0.268

***** CHANNEL 2

Sample #1 = 3177.00
Sample #2 = 3164.00
Sample #3 = 3185.00
Sample #4 = 3172.00
Average Result = 3173.6667
STD DEV = 10.5987
REL STD DEV = 0.334

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1023

3 um H2O Adjust (mg/l $\times$ 10,000) = 426

9 um H2O Adjust (mg/l $\times$ 10,000) = 636

**** AUTO CAL PASS

ST JOHNS SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
01/11/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:16
Control Test	0.049	10:17
Air Blank	0.000	10:18
Control Test	0.049	10:18
Air Blank	0.000	10:19
Control Test	0.048	10:20
Air Blank	0.000	10:20
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature

ST JOHNS SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
01/11/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:26
Control Test	0.201	10:26
Air Blank	0.000	10:27
Control Test	0.200	10:28
Air Blank	0.000	10:28
Control Test	0.199	10:29
Air Blank	0.000	10:29
Control Test Stats		
Average	0.2000	
Std Dev	0.0010	
Rel Std Dev(%)	0.5000	

Operator's Signature

ST JOHNS SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
01/11/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:21
Control Test	0.080	10:22
Air Blank	0.000	10:22
Control Test	0.079	10:23
Air Blank	0.000	10:24
Control Test	0.079	10:24
Air Blank	0.000	10:25
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature

ST JOHNS SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
01/11/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:31
Control Test	0.081	10:31
Air Blank	0.000	10:32
Control Test	0.080	10:32
Air Blank	0.000	10:33
Control Test	0.081	10:33
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Operator's Signature

Dry

2021.01.1
2 23:35:33
-05'00'

DERR
Digitally signed
by DERR
Date: 2021.01.11
12:59:08 -05'00'