



INSTRUMENT PROCESSING SHEET

Agency Broward County Sheriff's Office

S/N 80-007434

Florida Department of
Law Enforcement

Date In 12/18/2020 DI Completion Date 01/05/2021

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>DERR</u> ☒ Annual ☒ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☒ 12V DC Cable Notes: _____ 	Quality Checks Performed By <u>DERR</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>237</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 01/30/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	202010A 10/05/2022	0.080	SD3968	202010B 10/05/2022	0.200	SD3969	202010D 10/06/2022	0.080 DGS	N/A	AG003005 01/30/2022	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>DERR</u> ☒ Lab Temp °C <u>22.11C</u> External Digital Therm. ID#: <u>300918</u> ☒ 34°C +/- .2 Serial #: <u>SD3967</u> ☒ 34°C +/- .2 Serial #: <u>SD3968</u> ☒ 34°C +/- .2 Serial #: <u>SD3969</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	SD3967	202010A 10/05/2022																																																
0.080	SD3968	202010B 10/05/2022																																																
0.200	SD3969	202010D 10/06/2022																																																
0.080 DGS	N/A	AG003005 01/30/2022																																																
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2021.01.07 13:40:07 -05'00'	Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ 		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial Number	Lot Number	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Department Inspection Performed By <u>DERR</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1017</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 47</u>		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969	☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>Michael D. Haughey</u> 2021.01.05 14:45:26 -05'00' <u>CR</u> 2021.01.07 13:39:34 Tech Review / Date _____ Adm n Review / Date _____																																				
Simulator	Serial Number																																																	
0.000	SD3965																																																	
Interferent	SD3966																																																	
0.050	SD3967																																																	
0.080	SD3968																																																	
0.200	SD3969																																																	



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007434, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007434</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/05/2021</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>03:08</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/05/2021

Date

David Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 03:08

Date of Inspection: 01/05/2021

Serial Number: 80-007434
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.049	0.081	0.202	0.080
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.081	0.203	0.081
0.000	0.049	0.081	0.203	0.080
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.081	0.203	0.081
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.082	0.203	0.080
0.000	0.050	0.082	0.203	0.080

Standard Deviations	0.0004	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

01/05/2021
Date

MX
2021.0
1.07
13:38:3
0
-05'00"

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007434	Broward County Sheriff's Office	01/05/2021	DERR DEEL

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒																																																																																																																																																
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/05/2021 Software: 8100.27 SN 80-007434 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>01:06</td></tr><tr><td>Control Test</td><td>0.047</td><td>01:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:07</td></tr><tr><td>Control Test</td><td>0.048</td><td>01:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:08</td></tr><tr><td>Control Test</td><td>0.049</td><td>01:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:10</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0480</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>2.0833</td><td></td></tr></table> Operator's Signature <i>DEEL</i>	Test	g/210L	Time	Air Blank	0.000	01:06	Control Test	0.047	01:07	Air Blank	0.000	01:07	Control Test	0.048	01:08	Air Blank	0.000	01:08	Control Test	0.049	01:09	Air Blank	0.000	01:10	Control Test Stats			Average	0.0480		Std Dev	0.0010		Rel Std Dev(%)	2.0833		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/05/2021 Software: 8100.27 SN 80-007434 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>01:11</td></tr><tr><td>Control Test</td><td>0.079</td><td>01:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:12</td></tr><tr><td>Control Test</td><td>0.079</td><td>01:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:13</td></tr><tr><td>Control Test</td><td>0.080</td><td>01:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:15</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0793</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr></table> Operator's Signature <i>DEEL</i>	Test	g/210L	Time	Air Blank	0.000	01:11	Control Test	0.079	01:12	Air Blank	0.000	01:12	Control Test	0.079	01:13	Air Blank	0.000	01:13	Control Test	0.080	01:14	Air Blank	0.000	01:15	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/05/2021 Software: 8100.27 SN 80-007434 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>01:16</td></tr><tr><td>Control Test</td><td>0.199</td><td>01:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:17</td></tr><tr><td>Control Test</td><td>0.200</td><td>01:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:18</td></tr><tr><td>Control Test</td><td>0.201</td><td>01:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:20</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2000</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.5000</td><td></td></tr></table> Operator's Signature <i>DEEL</i>	Test	g/210L	Time	Air Blank	0.000	01:16	Control Test	0.199	01:17	Air Blank	0.000	01:17	Control Test	0.200	01:18	Air Blank	0.000	01:18	Control Test	0.201	01:19	Air Blank	0.000	01:20	Control Test Stats			Average	0.2000		Std Dev	0.0010		Rel Std Dev(%)	0.5000		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/05/2021 Software: 8100.27 SN 80-007434 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>01:21</td></tr><tr><td>Control Test</td><td>0.080</td><td>01:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:22</td></tr><tr><td>Control Test</td><td>0.080</td><td>01:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:23</td></tr><tr><td>Control Test</td><td>0.081</td><td>01:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>01:24</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></table> Operator's Signature <i>DEEL</i>	Test	g/210L	Time	Air Blank	0.000	01:21	Control Test	0.080	01:21	Air Blank	0.000	01:22	Control Test	0.080	01:22	Air Blank	0.000	01:23	Control Test	0.081	01:23	Air Blank	0.000	01:24	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	01:06																																																																																																																																																	
Control Test	0.047	01:07																																																																																																																																																	
Air Blank	0.000	01:07																																																																																																																																																	
Control Test	0.048	01:08																																																																																																																																																	
Air Blank	0.000	01:08																																																																																																																																																	
Control Test	0.049	01:09																																																																																																																																																	
Air Blank	0.000	01:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	2.0833																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	01:11																																																																																																																																																	
Control Test	0.079	01:12																																																																																																																																																	
Air Blank	0.000	01:12																																																																																																																																																	
Control Test	0.079	01:13																																																																																																																																																	
Air Blank	0.000	01:13																																																																																																																																																	
Control Test	0.080	01:14																																																																																																																																																	
Air Blank	0.000	01:15																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	01:16																																																																																																																																																	
Control Test	0.199	01:17																																																																																																																																																	
Air Blank	0.000	01:17																																																																																																																																																	
Control Test	0.200	01:18																																																																																																																																																	
Air Blank	0.000	01:18																																																																																																																																																	
Control Test	0.201	01:19																																																																																																																																																	
Air Blank	0.000	01:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2000																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	01:21																																																																																																																																																	
Control Test	0.080	01:21																																																																																																																																																	
Air Blank	0.000	01:22																																																																																																																																																	
Control Test	0.080	01:22																																																																																																																																																	
Air Blank	0.000	01:23																																																																																																																																																	
Control Test	0.081	01:23																																																																																																																																																	
Air Blank	0.000	01:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007434

OWNING AGENCY: Broward County Sheriff's Office

DATE OF DEPARTMENT INSPECTION: 01/05/2021

AGENCY INSPECTOR: Joshua Sapp

ADDRESS: 200 NW 27ST AVE

CITY, STATE, ZIP: Ft. Lauderdale, FL 33311

TELEPHONE NUMBER: (954) 321-4849

FAX NUMBER: _____

EMAIL ADDRESS (if available):

Joshua_Sapp@sheriff.org

For Program Office Use Only:



Registration Issued

Instrument Added to Evidentiary Instrument Database

Instrument Added to Monthly Statistics Database

Contact Information Added to Instrument Database