



INSTRUMENT PROCESSING SHEET

Agency Baker County Sheriff's Office

S/N 80-007439

Florida Department of
Law Enforcement

Date In 5/6/2021

DI Completion Date 6/1/2021

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>IS</u> <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	Quality Checks By <u>RAW</u> Date <u>05-10-21</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>225</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A 10-5-22</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202010B 10-5-22</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D 10-6-22</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102 04-20-22</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202010A 10-5-22	0.080	MP5089	202010B 10-5-22	0.200	MP5090	202010D 10-6-22	0.080 DGS	N/A	AG011102 04-20-22	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ DI Temp. Checks By <u>RAW</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.60</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5090</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP5088	202010A 10-5-22																																																											
0.080	MP5089	202010B 10-5-22																																																											
0.200	MP5090	202010D 10-6-22																																																											
0.080 DGS	N/A	AG011102 04-20-22																																																											
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: <u>Original registration to FHP and switched to Baker County SO. Bypassed AI to operate instrument.- RAW</u> _____ _____ _____ _____ _____	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>RAW</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1019</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table> Attachments <div style="display: flex; justify-content: space-between;"> <div> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment </div> <div> <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Form 47</u> </div> </div> <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between;"> <div> David Eliezer Reyes-Rivera <small>Digitally signed by David Eliezer Reyes-Rivera Date: 2021.06.08 15:05:02 -0400</small> </div> <div style="text-align: right;"> <u>2021.06.</u> <u>14</u> <u>16:17:07</u> <u>04:00'</u> </div> </div> <div style="display: flex; justify-content: space-between;"> <div>Tech Review / Date _____</div> <div>Admin Review / Date _____</div> </div>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP5086																																																												
Interferent	MP5087																																																												
0.050	MP5088																																																												
0.080	MP5089																																																												
0.200	MP5090																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 15:18

Date of Inspection: 06/01/2021

Serial Number: 80-007439
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.049	0.080	0.199	0.079
0.000	0.048	0.080	0.198	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.048	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.079
Standard Deviations	0.0004	0.0000	0.0003	0.0000


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

A F / M A: No Sample Provided. TIMED OUT. Walked away and did not provide sample within 3 minute window. Repeated test, OK. -RAW.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 RICHARD A WILLIAMS
Signature and Printed Name

06/01/2021
Date

Stability Checks

BAKER COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-007439

06/01/2021

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:42
Control Test	0.049	12:43
Air Blank	0.000	12:43
Control Test	0.049	12:44
Air Blank	0.000	12:44
Control Test	0.049	12:45
Air Blank	0.000	12:46
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.05

RAW

Operator's Signature

BAKER COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-007439

06/01/2021

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:49
Control Test	0.079	12:49
Air Blank	0.000	12:50
Control Test	0.079	12:50
Air Blank	0.000	12:51
Control Test	0.079	12:52
Air Blank	0.000	12:52
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.08
wet

RAW

Operator's Signature

BAKER COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-007439

06/01/2021

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:09
Control Test	0.199	13:10
Air Blank	0.000	13:11
Control Test	0.199	13:11
Air Blank	0.000	13:12
Control Test	0.199	13:13
Air Blank	0.000	13:13
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.20

RAW

Operator's Signature

BAKER COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-007439

06/01/2021

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:18
Control Test	0.079	13:18
Air Blank	0.000	13:19
Control Test	0.080	13:19
Air Blank	0.000	13:20
Control Test	0.080	13:20
Air Blank	0.000	13:21
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

0.08

DGS

RAW

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
05/10/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:04
Control Test	0.049	15:04
Air Blank	0.000	15:05
Control Test	0.049	15:06
Air Blank	0.000	15:06
Control Test	0.049	15:07
Air Blank	0.000	15:07
Control Test Stats		
Average	0.1490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.05

PAW

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
05/10/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:12
Control Test	0.080	15:13
Air Blank	0.000	15:13
Control Test	0.080	15:14
Air Blank	0.000	15:14
Control Test	0.079	15:15
Air Blank	0.000	15:16
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

0.08

wet

PAW

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
05/10/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:19
Control Test	0.199	15:20
Air Blank	0.000	15:20
Control Test	0.198	15:21
Air Blank	0.000	15:21
Control Test	0.199	15:22
Air Blank	0.000	15:23
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

0.20

PAW

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
05/10/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:26
Control Test	0.080	15:26
Air Blank	0.000	15:27
Control Test	0.080	15:27
Air Blank	0.000	15:27
Control Test	0.080	15:28
Air Blank	0.000	15:28
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.08 DGS

PAW

Operator's Signature

Agency changed.
Redid stabilities.(6/11)

PAW

Stability Checks
80-007439

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 12:24

Date of Inspection: 06/01/2021

Serial Number: 80-007439

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

Time-Date changed. BYPASSED TO OPERATE INSTRUMENT.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name RICHARD A WILLIAMS

06/01/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 14:55

Date of Inspection: 05/10/2021

Serial Number: 80-007439

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE, NOT A COMPLIANCE CHECK

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

RICHARD A WILLIAMS

05/10/2021
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007439, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007439</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>BAKER COUNTY SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>06/01/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:18</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/01/2021

Date

Richard A Williams Digitally signed by Richard A Williams
Date: 2021.06.07 16:14:46 -04'00'

RICHARD A WILLIAMS,
Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007439

OWNING AGENCY: Baker County Sheriff's Office

DATE OF DEPARTMENT INSPECTION: 6/1/2021

AGENCY INSPECTOR: Derek Watson

ADDRESS: 1 Sheriff's Office Drive

CITY, STATE, ZIP: Macclenny, Florida, 32063

TELEPHONE NUMBER: 904-259-3311

FAX NUMBER: N/A

EMAIL ADDRESS (if available): dwatson@bakerso.com

For Program Office Use Only:





Registration Issued

Instrument Added to Evidentiary Instrument Database

Instrument Added to Monthly Statistics Database

Contact Information Added to Instrument Database



Agency **Florida Highway Patrol**

S/N 80-007439

Date In 4/2/2021 DI Completion Date 4/5/2021

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

FDLE/ATP Form 48 January 2021
Issuing Authority: Alcohol Testing Program

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 11:36

Date of Inspection: 04/05/2021

Serial Number: 80-007439

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.051	0.081	0.201	0.080
0.000	0.050	0.081	0.201	0.081
0.000	0.050	0.081	0.201	0.081
0.000	0.051	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.051	0.081	0.201	0.081
0.000	0.050	0.081	0.201	0.080
0.000	0.050	0.081	0.202	0.080
0.000	0.050	0.081	0.201	0.080
0.000	0.051	0.081	0.201	0.081


Standard Deviations	0.0005	0.0004	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

04/05/2021
Date

stability checks

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
04/05/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:08
Control Test	0.050	09:09
Air Blank	0.000	09:09
Control Test	0.050	09:10
Air Blank	0.000	09:11
Control Test	0.050	09:11
Air Blank	0.000	09:12
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
04/05/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:13
Control Test	0.081	09:14
Air Blank	0.000	09:14
Control Test	0.081	09:15
Air Blank	0.000	09:15
Control Test	0.080	09:16
Air Blank	0.000	09:17
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
04/05/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:18
Control Test	0.200	09:18
Air Blank	0.000	09:19
Control Test	0.200	09:20
Air Blank	0.000	09:20
Control Test	0.200	09:21
Air Blank	0.000	09:21
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet



Operator's Signature



Operator's Signature



Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
04/05/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:24
Control Test	0.080	09:24
Air Blank	0.000	09:25
Control Test	0.081	09:25
Air Blank	0.000	09:26
Control Test	0.081	09:26
Air Blank	0.000	09:26
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Dry



Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007439, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007439</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>04/05/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:36</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Digitally signed by Israel Soto
Date: 2021.04.05 12:56:40
-04'00'

Israel Soto

04/05/2021

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007439

OWNING AGENCY: Florida Highway Patrol

DATE OF DEPARTMENT INSPECTION: 04-05-2021

AGENCY INSPECTOR: Susan Barge

ADDRESS: 6030 CR 2321

CITY, STATE, ZIP: Panama City, FL. 32404

TELEPHONE NUMBER: 352-620-4701

FAX NUMBER: _____

EMAIL ADDRESS (if available): Susanbarge@flhsmv.gov

For Program Office Use Only:

- ec 4/21/21
- ☒ Registration Issued
 - ☒ Instrument Added to Evidentiary Instrument Database
 - ☒ Instrument Added to Monthly Statistics Database
 - ☒ Contact Information Added to Instrument Database