



INSTRUMENT PROCESSING SHEET

Agency Desoto County Sheriff's OfficeS/N 80-007509Florida Department of
Law EnforcementDate In 10/29/2021 DI Completion Date 11/1/2021☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By DERR	Quality Checks	By DER	Date 11/1/2021	Flow Calibration	By	Date																																					
☒ Annual ☒ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>200</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP6287</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP6288</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG026705 09/23/2022</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202010A 10/05/2022	0.080	MP6287	202010B 10/05/2022	0.200	MP6288	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																								
Simulator	Serial #	Lot #/Exp																																										
0.050	MP6286	202010A 10/05/2022																																										
0.080	MP6287	202010B 10/05/2022																																										
0.200	MP6288	202010D 10/06/2022																																										
0.080 DGS	N/A	AG026705 09/23/2022																																										
		DI Temp. Checks By DERR ☒ Lab Temp °C <u>21.82C</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>MP6286</u> ☒ 34°C +/- .2 Serial #: <u>MP6287</u> ☒ 34°C +/- .2 Serial #: <u>MP6288</u>																																										
Calibration Adjustment By _____		Department Inspection By DERR _____																																										
Barometric Pressure Gauge ID # _____		Barometric Pressure ID# <u>28199</u> Gauge <u>1021</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																									
0.000		N/A	N/A																																									
0.040																																												
0.100																																												
0.200																																												
0.300																																												
0.080 DGS	N/A																																											
Simulator	Serial Number																																											
0.000	MP6284																																											
Interferent	MP6285																																											
0.050	MP6286																																											
0.080	MP6287																																											
0.200	MP6288																																											
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Form 47</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																						
Simulator	Serial #	Lot #	Expiration																																									
0.050																																												
0.080																																												
0.200																																												
0.080 DGS	N/A																																											
Notes/Suggested Service: Printed the 0.05 stabilities on the external printer by mistake. DERR 11/1/2021		Israel Soto Tech Review / Date <u>12:21:44</u> Admin Review / Date _____																																										



ANAB
ANSI National Accreditation Board
ACCREDITED
ISO/IEC 17025
FORENSIC CALIBRATION
LABORATORY

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007509, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007509</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>DESOTO COUNTY S.O.</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>11/01/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:57</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/01/2021

Date

David E Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 January 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.

Time of Inspection: 12:57

Date of Inspection: 11/01/2021

Serial Number: 80-007509

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.080	0.197	0.080
0.000	0.049	0.080	0.197	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.079
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.080

Standard Deviations	0.0003	0.0004	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

11/01/2021
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007509	Desoto County Sheriff's Office	11/1/2021	DERR <i>the</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒																																																																																																																																																
DESOTO COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr><tr><td>Control Test</td><td>0.050</td><td>10:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr><tr><td>Control Test Status</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></tbody></table>  Operator's Signature Printed on external Printer by mistake bear	Test	g/210L	Time	Air Blank	0.000	10:28	Control Test	0.050	10:29	Air Blank	0.000	10:29	Control Test	0.049	10:30	Air Blank	0.000	10:31	Control Test	0.049	10:31	Air Blank	0.000	10:32	Control Test Status			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		DESOTO COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr><tr><td>Control Test Status</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:33	Control Test	0.079	10:34	Air Blank	0.000	10:34	Control Test	0.079	10:35	Air Blank	0.000	10:36	Control Test	0.079	10:36	Air Blank	0.000	10:37	Control Test Status			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		DESOTO COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:40</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr><tr><td>Control Test Status</td><td></td><td></td></tr><tr><td>Average</td><td>0.1980</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:38	Control Test	0.198	10:39	Air Blank	0.000	10:39	Control Test	0.198	10:40	Air Blank	0.000	10:40	Control Test	0.198	10:41	Air Blank	0.000	10:42	Control Test Status			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		DESOTO COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:44</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:44</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:45</td></tr><tr><td>Control Test Status</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:43	Control Test	0.080	10:43	Air Blank	0.000	10:44	Control Test	0.079	10:44	Air Blank	0.000	10:44	Control Test	0.080	10:45	Air Blank	0.000	10:45	Control Test Status			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:28																																																																																																																																																	
Control Test	0.050	10:29																																																																																																																																																	
Air Blank	0.000	10:29																																																																																																																																																	
Control Test	0.049	10:30																																																																																																																																																	
Air Blank	0.000	10:31																																																																																																																																																	
Control Test	0.049	10:31																																																																																																																																																	
Air Blank	0.000	10:32																																																																																																																																																	
Control Test Status																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:33																																																																																																																																																	
Control Test	0.079	10:34																																																																																																																																																	
Air Blank	0.000	10:34																																																																																																																																																	
Control Test	0.079	10:35																																																																																																																																																	
Air Blank	0.000	10:36																																																																																																																																																	
Control Test	0.079	10:36																																																																																																																																																	
Air Blank	0.000	10:37																																																																																																																																																	
Control Test Status																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:38																																																																																																																																																	
Control Test	0.198	10:39																																																																																																																																																	
Air Blank	0.000	10:39																																																																																																																																																	
Control Test	0.198	10:40																																																																																																																																																	
Air Blank	0.000	10:40																																																																																																																																																	
Control Test	0.198	10:41																																																																																																																																																	
Air Blank	0.000	10:42																																																																																																																																																	
Control Test Status																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:43																																																																																																																																																	
Control Test	0.080	10:43																																																																																																																																																	
Air Blank	0.000	10:44																																																																																																																																																	
Control Test	0.079	10:44																																																																																																																																																	
Air Blank	0.000	10:44																																																																																																																																																	
Control Test	0.080	10:45																																																																																																																																																	
Air Blank	0.000	10:45																																																																																																																																																	
Control Test Status																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency:

Time of Inspection: 10:02

Date of Inspection: 11/01/2021

Serial Number: 80-007509

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____


Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA
Signature and Printed Name

11/01/2021
Date



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer8000

SERIAL NUMBER: 80-007509

OWNING AGENCY: Desoto County Sheriff's Office

DATE OF DEPARTMENT INSPECTION: 11/1/2021

AGENCY INSPECTOR: Deputy Joshua Pitts

ADDRESS: 208 East Cypress St

CITY, STATE, ZIP: Arcadia, Florida 34266

TELEPHONE NUMBER: (863) 993-4710

FAX NUMBER: _____

EMAIL ADDRESS (if available): Joshuapitts@desotosheriff.org

For Program Office Use Only:

 **Registration Issued**
 **Instrument Added to Evidentiary Instrument Database**
 **Instrument Added to Monthly Statistics Database**
 **Contact Information Added to Instrument Database**