



Florida Department of  
Law Enforcement

## DEPARTMENT INSPECTOR FIELD NOTES

Agency: Dade City Police Department Instrument Serial Number: 80-001073 and 80-005477

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
<input type="checkbox"/> Registration <input type="checkbox"/> Request for Registration Attached <input type="checkbox"/> Annual <input type="checkbox"/> Inspection After Repair <input type="checkbox"/> Other: _____  <input type="checkbox"/> Barometric Pressure: _____ Gauge _____ Instrument  <input type="checkbox"/> Instrument Set Up Verified	<input type="checkbox"/> Agency Inspection Notification <input checked="" type="checkbox"/> Agency Contact <input type="checkbox"/> Other: _____  Date of Notification/Contact: <u>7/16/2021</u>  Agency Inspector: <u>Dean Baldwin</u>	The following records were audited: <input checked="" type="checkbox"/> Agency Inspection Reports <input checked="" type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input checked="" type="checkbox"/> Instrument Registration <input checked="" type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis  <input type="checkbox"/> Other: _____ <input type="checkbox"/> Other: _____
Comments: <u>Compliant with Chapter 11D-8, FAC.</u>    		

Instrument/Area:	Equipment:	Supplies:
<input checked="" type="checkbox"/> Clean/Dry <input checked="" type="checkbox"/> Secure <input checked="" type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input checked="" type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input checked="" type="checkbox"/> Class A Glassware	<input checked="" type="checkbox"/> Distilled/Deionized Water <input checked="" type="checkbox"/> Mouth Alcohol Solution <input checked="" type="checkbox"/> Acetone Stock Solution <input type="checkbox"/> Alcohol Reference Solution <input checked="" type="checkbox"/> Dry Gas Standard <input checked="" type="checkbox"/> Mouth Pieces
Comments: <u>Unable to audit Alcohol Reference Solutions. Agency Inspector was out of stock and awaiting receipt of a refill order. Discussed options for the Agency</u> <u>Inspection later this month if the solutions do not arrive on time, including borrowing solutions from other agencies or performing the Agency Inspection at</u> <u>another agency at the same time as their inspection in order to use the same solutions. An expired lot of mouth alcohol (2018-B) was found to be in use.</u>		

Action:	
<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC  <input type="checkbox"/> Remain in Evidentiary Use <input type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remove from Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use  <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	Comments: <u>Instructed Agency Inspector to discontinue using the expired mouth alcohol and discard</u> <u>immediately. Provided a new, non-expired lot of mouth alcohol (and acetone) for use at next</u> <u>inspection. Examination of affected Agency Inspections shows instruments demonstrating</u> <u>proper response when introduced to a breath sample containing the expired mouth alcohol.</u>

Taylor Gutschow

Digitally signed by Taylor Gutschow  
Date: 2021.07.21 08:36:59 -04'00'

7/21/2021

Signature of Alcohol Testing Program Staff Member

Date

Original – FDLE

Copy – Agency