



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: NAS Jacksonville Instrument Serial Number: 80-000778 & 2097

Department Inspection: Agency Inspection/Agency Contact: Records Audit: Registration, Agency Inspection Notification, The following records were audited: Agency Inspection Reports, Breath Test Instrument Repair/Maintenance Records, Instrument Registration, Dry Gas Standard Certificate(s) of Analysis

Comments: Instruments needed uploading after AI completed on 9/26. Registration records maintained but not located on visit. Requested registration from Department.

Instrument/Area: Equipment: Supplies: Clean/Dry, Secure, Limited Access, Other: Proper Number of Simulators, Checked Simulators for Air Leak Resistant Seal and Proper Temperature, Class A Glassware, Distilled/Deionized Water, Mouth Alcohol Solution, Acetone Stock Solution, Alcohol Reference Solution, Dry Gas Standard, Mouth Pieces

Comments:

Action: Instrument Complies with Chapter 11D-8, FAC, Instrument Does Not Comply with Chapter 11D-8, FAC, Remain in Evidentiary Use, Return to/Place into Evidentiary Use, Remove from Evidentiary Use, Remain Out of Evidentiary Use, Conduct an Agency Inspection Before Evidentiary Use

Richard A Williams

Digitally signed by Richard A Williams Date: 2021.10.04 12:17:22 -04'00'

Signature of Alcohol Testing Program Staff Member

Date

Original - FDLE

Copy - Agency