



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Pinellas County Sheriff's Office Instrument Serial Number: n/a

Department Inspection: Agency Inspection/Agency Contact: Records Audit: Registration, Agency Inspection Notification, The following records were audited: Agency Inspection Reports, Breath Test Instrument Repair/Maintenance Records, Instrument Registration, Dry Gas Standard Certificate(s) of Analysis

Comments: Compliant with Chapter 11D-8, FAC. Agency Inspector Deane is meticulously organized.

Instrument/Area: Equipment: Supplies: Clean/Dry, Secure, Limited Access, Proper Number of Simulators, Checked Simulators for Air Leak Resistant Seal and Proper Temperature, Class A Glassware, Distilled/Deionized Water, Mouth Alcohol Solution, Acetone Stock Solution, Alcohol Reference Solution, Dry Gas Standard, Mouth Pieces

Comments: Compliant with Chapter 11D-8, FAC. No expired alcohol solutions, stock solutions, or dry gas.

Action: Instrument Complies with Chapter 11D-8, FAC, Instrument Does Not Comply with Chapter 11D-8, FAC, Remain in Evidentiary Use, Return to/Place into Evidentiary Use, Remove from Evidentiary Use, Remain Out of Evidentiary Use, Conduct an Agency Inspection Before Evidentiary Use

Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2021.06.16 09:53:12 -04'00' 6/15/2021 Signature of Alcohol Testing Program Staff Member Date

Original - FDLE Copy - Agency