



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: St. Johns County Sheriff's Office Instrument Serial Number: Multiple, see comments

Department Inspection: Agency Inspection/Agency Contact: Records Audit: Comments: Instruments: 80-007366, 80-001133, 80-007363, 80-001134, 80-001655 Took instrument 80-001655 back to ATP for Department Inspection.

Instrument/Area: Equipment: Supplies: Comments: Checked simulators for cracks.

Action: Instrument Complies with Chapter 11D-8, FAC Instrument Does Not Comply with Chapter 11D-8, FAC Comments: Site inspection, instrument compliance not determined. Site inspected 08-26-2021

Israel Soto

Digitally signed by Israel Soto Date: 2021.08.27 09:14:47 -04'00'

08-27-2021

Signature of Alcohol Testing Program Staff Member

Date

Original - FDLE

Copy - Agency



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: St. Johns Co. Sheriff's Office Instrument Serial Number: 80-001655,1133,1134

Department Inspection: Agency Inspection/Agency Contact: Records Audit: Registration, Agency Inspection Notification, The following records were audited: Agency Inspection Reports, Breath Test Instrument Repair/Maintenance Records, Instrument Registration, Dry Gas Standard Certificate(s) of Analysis, Other: Records kept at SA's, office

Comments: Security for Breath test room requires key check per operator for access. Instrument had a power issue initially before changing outlet and the instrument turned on.

Instrument/Area: Equipment: Supplies: Clean/Dry, Secure, Limited Access, Other: Proper Number of Simulators, Checked Simulators for Air Leak Resistant Seal and Proper Temperature, Class A Glassware, Distilled/Deionized Water, Mouth Alcohol Solution, Acetone Stock Solution, Alcohol Reference Solution, Dry Gas Standard, Mouth Pieces

Comments:

Action: Instrument Complies with Chapter 11D-8, FAC, Instrument Does Not Comply with Chapter 11D-8, FAC, Remain in Evidentiary Use, Return to/Place into Evidentiary Use, Remove from Evidentiary Use, Remain Out of Evidentiary Use, Conduct an Agency Inspection Before Evidentiary Use, Comments:

Richard A Williams

Digitally signed by Richard A Williams Date: 2021.10.04 12:17:22 -04'00'

Signature of Alcohol Testing Program Staff Member

Date

Original - FDLE

Copy - Agency